

INSTRUMENT PROCESSING SHEET

Agency Sarasota CSO

S/N 80-000861

Florida Department of
Law Enforcement

Date In 01/31/2023 DI Completion Date 02/08/2023

☐Ship ☒P/U ☐H/D ☐CMI ☐EE

Intake	By	TDG	Quality Checks	By	TDG	Date	02/01/2023	Flow Calibration	By	Date															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value 248 ☒ Flow Verification (L/s) Flow Column # ATP104 32 mm 0.156 (.139 - .169) 36 mm 0.175 (.156 - .190) 53 mm 0.234 (.228 - .278) 103 mm 0.519 (.447 - .547) ☒ Barometric Pressure Check Gauge ID # 68639 ☒ Stability Checks					Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																	
			<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5092</td> <td>202201C 01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP5093</td> <td>202201D 01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP5094</td> <td>202201E 01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>00521080A2 02/05/2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C 01/11/2024	0.080	MP5093	202201D 01/18/2024	0.200	MP5094	202201E 01/18/2024	0.080 DGS	N/A	00521080A2 02/05/2023					Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____		
Simulator	Serial #	Lot #/Exp																							
0.050	MP5092	202201C 01/11/2024																							
0.080	MP5093	202201D 01/18/2024																							
0.200	MP5094	202201E 01/18/2024																							
0.080 DGS	N/A	00521080A2 02/05/2023																							

Calibration Adjustment				By <u>TDG</u>	
Barometric Pressure Gauge <u>1025</u> ID # <u>28199</u>				Barometric Pressure ID# <u>28663</u>	
Simulator	Serial #	Lot #	Expiration	Gauge <u>1024</u>	Instrument <u>1024</u>
0.000	MP5099	N/A	N/A	Mouth Alcohol Solution Lot # <u>2021-D</u>	
0.040	MP5096	21410	09/30/2023	Acetone Stock Solution Lot # <u>2021-C</u>	
0.100	MP5098	22310	08/11/2024	Simulator	Serial Number
0.200	MP5100	22050	02/07/2024	0.000	MP5095
0.300	MP5101	22220	06/15/2024	Interferent	MP5097
0.080 DGS	N/A	AG115904	06/08/2023	0.050	MP5092
☒ Post Calibration Adjustment Stability Checks				0.080	MP5093
				0.200	MP5094
				Attachments	
				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____
				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use	
Notes/Suggested Service: _____ _____ Tech Review: Corrected the stabilities to reflect the agreement between the 0.08 ARS/DGS was >0.003. (TDG 02/09/2023) _____ _____ _____ _____				Israel Soto Digitally signed by Israel Soto Date: 2023.02.09 13:56:58 -0500  13:56:32 Tech Review / Date Admin Review / Date	

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000761	Sarasota CSO	02/01/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/01/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:24 Control Test 0.048 15:25 Air Blank 0.000 15:25 Control Test 0.048 15:26 Air Blank 0.000 15:26 Control Test 0.048 15:27 Air Blank 0.000 15:28 Control Test Stats Average 0.0480 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 	SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/01/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:31 Control Test 0.077 15:32 Air Blank 0.000 15:33 Control Test 0.076 15:33 Air Blank 0.000 15:34 Control Test 0.077 15:35 Air Blank 0.000 15:35 Control Test Stats Average 0.0767 Std Dev 0.0006 Rel Std Dev(%) 0.7531 Operator's Signature 	SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/01/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:40 Control Test 0.202 15:40 Air Blank 0.000 15:41 Control Test 0.200 15:42 Air Blank 0.000 15:42 Control Test 0.199 15:43 Air Blank 0.100 15:44 Control Test Stats Average 0.2003 Std Dev 0.0015 Rel Std Dev(%) 0.7625 Operator's Signature 	SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 02/01/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 15:13 Control Test 0.080 15:13 Air Blank 0.000 15:13 Control Test 0.080 15:14 Air Blank 0.000 15:14 Control Test 0.080 15:15 Air Blank 0.000 15:15 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 Operator's Signature 

Comments: The 0.08 ARS was outside the acceptable range. Will perform an optical cal adjust. MG 02/01/2023

***** AUTO CAL DATA *****

<<<< CHANNEL 1 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.067
 Std Dev = 0.01 Rel Std Dev = 15.39
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.773
 Std Dev = 0.01 Rel Std Dev = 0.93
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.802
 Std Dev = 0.00 Rel Std Dev = 0.22
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.483
 Std Dev = 0.01 Rel Std Dev = 0.15
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.105
 Std Dev = 0.01 Rel Std Dev = 0.15
 Zero Order Coef = -182.35
 First Order Coef = 2687.65
 Second Order Coef = 28.62
 Standard Deviation = 5.57463

<<<< CHANNEL 2 >>>>

Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.122
 Std Dev = 0.01 Rel Std Dev = 6.38
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.550
 Std Dev = 0.01 Rel Std Dev = 0.50
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.627
 Std Dev = 0.01 Rel Std Dev = 0.31
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.927
 Std Dev = 0.01 Rel Std Dev = 0.19
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 10.035
 Std Dev = 0.02 Rel Std Dev = 0.17
 Zero Order Coef = -156.26
 First Order Coef = 1307.30
 Second Order Coef = 13.11
 Standard Deviation = 3.952466

Optical Calibration

SN: 80-000861

Agency: Sarasota CSO

Date: 02/08/2023

Quadratic Fit: +/- 0.002g/210L ✓

By: TDG NL

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)
 Sample #1 = 1.5600 (-0.0160)
 Sample #2 = 1.5590 (0.0000)
 Sample #3 = 1.5440 (0.0070)
 Sample #4 = 1.5460 (0.0180)
 Avg % Abs = 1.5503 (0.0083)
 STD DEV = 0.0078 (0.0091)
 REL STD DEV = 0.501 (108.885)

<<<< CHANNEL 1 >>>>

Sol Value = 0.100 g/210L ***
 Fit Value = 0.4762 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12736, Sum Io = 13334
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.8290 (-0.0160)
 Sample #2 = 1.8000 (0.0290)
 Sample #3 = 1.8070 (0.0240)
 Sample #4 = 1.8000 (0.0370)
 Avg % Abs = 1.8023 (0.0300)
 STD DEV = 0.0040 (0.0066)
 REL STD DEV = 0.224 (21.858)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)
 Sample #1 = 3.6660 (-0.0170)
 Sample #2 = 3.6370 (0.0300)
 Sample #3 = 3.6300 (0.0290)
 Sample #4 = 3.6150 (0.0440)
 Avg % Abs = 3.6273 (0.0343)
 STD DEV = 0.0112 (0.0084)
 REL STD DEV = 0.310 (24.427)

<<<< CHANNEL 1 >>>>

Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12730, Sum Io = 13331
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.5370 (-0.0100)
 Sample #2 = 3.4890 (0.0380)
 Sample #3 = 3.4790 (0.0630)
 Sample #4 = 3.4810 (0.0610)
 Avg % Abs = 3.4830 (0.0540)
 STD DEV = 0.0053 (0.0139)
 REL STD DEV = 0.152 (25.727)

SARASOTA COUNTY SO
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 02/08/2023
 SN 80-000861
 09:46:47

Auto Calibration
 Max Power Res Value = 52
 Auto Range Res Value = 43

<<<< CHANNEL 1 >>>>

Sol Value = 0.000 g/210L ***
 Fit Value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12752, Sum Io = 13344
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.0620 (0.0000)
 Sample #2 = 0.0580 (0.0040)
 Sample #3 = 0.0780 (0.0100)
 Sample #4 = 0.0640 (0.0220)
 Avg % Abs = 0.0667 (0.0120)
 STD DEV = 0.0103 (0.0092)
 REL STD DEV = 15.395 (76.376)

<<<< CHANNEL 2 >>>>

Sample % Abs (% Abs Ref)
 Sample #1 = 0.1280 (-0.0090)
 Sample #2 = 0.1130 (-0.0010)
 Sample #3 = 0.1280 (-0.0010)
 Sample #4 = 0.1240 (-0.0050)
 Avg % Abs = 0.1217 (-0.0023)
 STD DEV = 0.0078 (0.0023)
 REL STD DEV = 6.384 (98.974)

<<<< CHANNEL 1 >>>>

Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12741, Sum Io = 13338
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.7620 (-0.0070)
 Sample #2 = 0.7710 (0.0060)
 Sample #3 = 0.7670 (0.0100)
 Sample #4 = 0.7810 (0.0110)
 Avg % Abs = 0.7730 (0.0090)
 STD DEV = 0.0072 (0.0026)
 REL STD DEV = 0.933 (29.397)

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	0.0001
0.100	0.100	0.0001
0.200	0.200	-0.0001
0.300	0.300	0.0001

Sol Value = 0.080 g/210L ***
 Fit Value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 3317.00
 Sample #2 = 3284.00
 Sample #3 = 3305.00
 Sample #4 = 3325.00
 Average Result = 3304.6667
 STD DEV = 20.5020
 REL STD DEV = 0.620

***** CHANNEL 2

Sample #1 = 3375.00
 Sample #2 = 3386.00
 Sample #3 = 3397.00
 Sample #4 = 3411.00
 Average Result = 3398.0000
 STD DEV = 12.5300
 REL STD DEV = 0.369

Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1025
 3 in H2O Adjust (mg/l x 10,000) = 505
 9 in H2O Adjust (mg/l x 10,000) = 411
 ***** AUTO CAL PASS

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0001
0.040	0.040	-0.0002
0.100	0.100	0.0001
0.200	0.200	-0.0001
0.300	0.300	0.0000

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post- ^(a))	80-000861	Sarasota (SD)	02/08/2023	TDG <i>MG</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000861 02/08/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:51 Control Test 0.049 10:51 Air Blank 0.000 10:52 Control Test 0.049 10:53 Air Blank 0.000 10:53 Control Test 0.048 10:54 Air Blank 0.000 10:54 Control Test Stats Average 0.0487 Std Dev 0.0006 Rel Std Dev(%) 1.1863 <i>MG</i> Operator's Signature	SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000861 02/08/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:59 Control Test 0.078 11:00 Air Blank 0.000 11:00 Control Test 0.078 11:01 Air Blank 0.000 11:01 Control Test 0.078 11:02 Air Blank 0.000 11:03 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>MG</i> Operator's Signature	SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000861 02/08/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:06 Control Test 0.201 11:07 Air Blank 0.000 11:08 Control Test 1.198 11:08 Air Blank 0.000 11:09 Control Test 0.198 11:10 Air Blank 0.000 11:10 Control Test Stats Average 0.1990 Std Dev 0.0017 Rel Std Dev(%) 0.8704 <i>MG</i> Operator's Signature	<i>DGS</i> SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000861 02/08/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 11:12 Control Test 0.080 11:12 Air Blank 0.000 11:12 Control Test 0.080 11:13 Air Blank 0.000 11:13 Control Test 0.080 11:14 Air Blank 0.000 11:14 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000 <i>MG</i> Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 14:09

Date of Inspection: 02/08/2023

Serial Number: 80-000861
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG222203 Exp: 08/10/2024
0.000	0.049	0.077	0.199	0.081
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.077	0.200	0.080
0.000	0.049	0.077	0.199	0.080
0.000	0.049	0.077	0.199	0.080
0.000	0.049	0.077	0.199	0.080
0.000	0.049	0.077	0.199	0.080
0.000	0.049	0.078	0.199	0.081
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.077	0.199	0.080

Standard Deviations	0.0000	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/08/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000861, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000861	UNCERTAINTY* $\pm$
Owning Agency:	SARASOTA COUNTY SO	0.050 g/ 210 L 0.004
Calibration Date:	02/08/2023	0.080 g/ 210 L 0.004
Calibration Time:	14:09	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/08/2023

Date


TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality