INSTRUMENT PROCESSING SHEET

FDLE Florida Department of Law Enforcement

Agency Pinecrest PD S/N 80-000909

Date In 08/25/2023 DI Completion Date 08/25/2023 ☐Ship ■P/U ☐H/D ☐CMI ☐EE

Intake By TDG

- ■ Annual
- ☐ Registration
- ☐ Return from CMI / EE

Visual Inspection:

- ■ Case ■ Handle
- ■ Keyboard ■ Dry Gas Shelf
- ■ Feet ■ Breath Tube
- ■ Ports ■ Screws Tight

Other Equipment/ Accessories:

- ☐ Power cord ☐ Printer Cable
- ☐ Static Bag ☐ 12V DC Cable

Notes: Sticky keys on keyboard

Quality Checks By TDG Date 08/25/2023

- ■ Breath Tube Screen
- ■ Replace External O-Rings
- ■ Instrument Set Up Verified
- ■ R-Value 129
- ■ Flow Verification (L/s)

Flow Column # ATP104

32 mm 0.144 (.139 - .169)
36 mm 0.164 (.156 - .190)
53 mm 0.234 (.228 - .278)
103 mm 0.500 (.447 - .547)

- ■ Barometric Pressure Check

Gauge ID # 26932

- ■ Stability Checks

Simulator	Serial #	Lot #/Exp
0.050	MP5094	202201C
		01/11/2024
0.080	MP5095	202201D
		01/18/2024
0.200	MP5096	202201E
		01/18/2024
0.080 DGS	N/A	AG223802
		08/26/2024

Flow Calibration By ____ Date ____

Flow Column # ____

- ☐ 5L/min – 17mm
- ☐ 15L/min – 53mm
- ☐ 30L/min – 103mm
- ☐ R-Value ____
- ☐ Post Calibration Verification (L/s)

Flow Column # ____

32 mm ____ (.139 - .169)
36 mm ____ (.156 - .190)
53 mm ____ (.228 - .278)
103 mm ____ (.447 - .547)

Maintenance By ____

- ☐ Battery Replacement
- ☐ Dry Gas Regulator Replacement
- ☐ Breath Tube Replacement
- ☐ Other ____

Calibration Adjustment By ____

Barometric Pressure Gauge ____ ID # ____

Simulator	Serial #	Lot #	Expiration
0.000		N/A	N/A
0.040			
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Department Inspection By TDG

Barometric Pressure ID# 26932

Gauge 1014 Instrument 1011

Mouth Alcohol Solution Lot # 2021-D

Acetone Stock Solution Lot # 2022-B

Simulator	Serial Number
0.000	MP6284
Interferent	MP5093
0.050	MP5094
0.080	MP5095
0.200	MP5096

Attachments

- ■ Form 41
- ■ Stability Checks
- ■ Calibration Certificate
- ☐ Calibration Adjustment
- ☐ Post-Stability Checks
- ☐ Flow Calibration
- ■ Form 40
- ☐ Other ____

Notes/Suggested Service: ____

- ■ **Instrument Complies with Chapter 11D-8, FAC**
- ☐ **Instrument Does Not Comply with Chapter 11D-8, FAC**
- ■ **Return to/Place into Evidentiary Use**
- ☐ **Remain Out of Evidentiary Use**
- ■ **Conduct an Agency Inspection Before Evidentiary Use**

Israel Soto Digitally signed by Israel Soto Date: 2023.08.25 12:23:23 -04'00' — Tech Review / Date

Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.08.25 13:38:04 -04'00' — Admin Review / Date

Florida Department of Law Enforcement
Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINECREST PD
Serial Number: 80-000909
Time of Inspection: 09:23
Date of Inspection: 08/25/2023
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:
AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined MG 8/25/2023

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor Gutschow
TAYLOR D GUTSCHOW
Signature and Printed Name

08/25/2023
Date

FDLE/ATP Form 40 – March 2004

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053 ✓	*0.077 to 0.083* ✓	*0.194 to 0.206* ✓	*0.077 to 0.083* ✓ *≤0.003 of Wet* ✓

PINECREST PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003919
08/25/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:35
Control Test	0.050	09:35
Air Blank	0.000	09:36
Control Test	0.051	09:37
Air Blank	0.000	09:37
Control Test	0.049	09:38
Air Blank	0.000	09:38

Control Test Stats

Average	0.0497
Std Dev	0.0006
Rel Std Dev(%)	1.1625

Operator's Signature

PINECREST PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003919
08/25/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:43
Control Test	0.079	09:44
Air Blank	0.000	09:44
Control Test	0.079	09:45
Air Blank	0.000	09:46
Control Test	0.078	09:46
Air Blank	0.000	09:47

Control Test Stats

Average	0.0787
Std Dev	0.0006
Rel Std Dev(%)	0.7339

Operator's Signature

PINECREST PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-003919
08/25/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:49
Control Test	0.199	09:50
Air Blank	0.000	09:51
Control Test	0.199	09:51
Air Blank	0.000	09:52
Control Test	0.198	09:52
Air Blank	0.000	09:53

Control Test Stats

Average	0.1987
Std Dev	0.0006
Rel Std Dev(%)	0.2936

Operator's Signature

Inadvertently printed to external printer. Results within range. Will attach.

MG
8/25/2023

PINECREST PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000909
08/25/2023
Software: 8100.27

DBS

Test	g/210L	Time
Air Blank	0.000	09:26
Control Test	0.081	09:26
Air Blank	0.000	09:26
Control Test	0.081	09:27
Air Blank	0.000	09:27
Control Test	0.081	09:28
Air Blank	0.000	09:28
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MG

Operator's Signature

Florida Department of Law Enforcement
Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINECREST PD
Time of Inspection: 11:50
Date of Inspection: 08/25/2023
Serial Number: 80-000909
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.050	0.078	0.200	0.080
0.000	0.050	0.078	0.199	0.080
0.000	0.050	0.079	0.200	0.080
0.000	0.050	0.079	0.200	0.079
0.000	0.050	0.079	0.201	0.079
0.000	0.050	0.079	0.201	0.080
0.000	0.050	0.079	0.200	0.080

Standard Deviations	0.0003	0.0004	0.0005	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor Gutschow

TAYLOR D GUTSCHOW
Signature and Printed Name

08/25/2023
Date

FDLE/ATP Form 41 –Revised August 2005

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000909 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	80-000909
Owning Agency:	PINECREST PD
Calibration Date:	08/25/2023
Calibration Time:	11:50

	UNCERTAINTY* ±
0.050 g/ 210 L	0.004
0.080 g/ 210 L	0.004
0.200 g/ 210 L	0.007
0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION
This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

08/25/2023
Date

TAYLOR D GUTSCHOW,
Department Inspector