







INSTRUMENT PROCESSING SHEET

Agency Collier CSOS/N 80-000943Florida Department of
Law EnforcementDate In 11/02/2023 DI Completion Date 11/06/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>11/03/2023</u>	Flow Calibration	By	Date																																									
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Instrument returned from HQ where stripped screws were removed from dry gas regulator plate.</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>144</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																											
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP5095</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5096</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5094	202303K 03/29/2025	0.080	MP5095	202303L 03/29/2025	0.200	MP5096	202304C 04/05/2025	0.080 DGS	N/A	AG223802 08/26/2024			Maintenance By <u>TDG</u> <input type="checkbox"/> Battery Replacement <input checked="" type="checkbox"/> Dry Gas Regulator Replacement on 11/3 <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																												
Simulator	Serial #	Lot #/Exp																																														
0.050	MP5094	202303K 03/29/2025																																														
0.080	MP5095	202303L 03/29/2025																																														
0.200	MP5096	202304C 04/05/2025																																														
0.080 DGS	N/A	AG223802 08/26/2024																																														
Calibration Adjustment By _____		Department Inspection By <u>TDG</u>																																														
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096	
Simulator	Serial #	Lot #	Expiration																																													
0.000		N/A	N/A																																													
0.040																																																
0.100																																																
0.200																																																
0.300																																																
0.080 DGS	N/A																																															
Simulator	Serial Number																																															
0.000	MP5092																																															
Interferent	MP5093																																															
0.050	MP5094																																															
0.080	MP5095																																															
0.200	MP5096																																															
<input type="checkbox"/> Post Calibration Adjustment Stability Checks		Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Notes/Suggested Service: <u>Replaced dry gas regulator and tared tank sensor prior to Quality Checks. Monitored for several days to ensure no leak. (TDG)</u>																										
Simulator	Serial #	Lot #	Expiration																																													
0.050																																																
0.080																																																
0.200																																																
0.080 DGS	N/A																																															
		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																														
Benjamin Siddoway		Digitally signed by Benjamin Siddoway Date: 2023.11.06 16:52:59 -05'00'		Phil Nicodemo		Digitally signed by Phil Nicodemo Date: 2023.11.15 11:23:03 -05'00'																																										
		Tech Review / Date				Admin Review / Date																																										

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000943 11/03/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:26</p> <p>Control Test 0.050 14:27</p> <p>Air Blank 0.000 14:28</p> <p>Control Test 0.050 14:28</p> <p>Air Blank 0.000 14:29</p> <p>Control Test 0.050 14:30</p> <p>Air Blank 0.000 14:30</p> <p>Control Test Status</p> <p>Average 0.0500</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000943 11/03/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:34</p> <p>Control Test 0.080 14:35</p> <p>Air Blank 0.000 14:36</p> <p>Control Test 0.080 14:36</p> <p>Air Blank 0.000 14:37</p> <p>Control Test 0.080 14:38</p> <p>Air Blank 0.000 14:38</p> <p>Control Test Status</p> <p>Average 0.0800</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000943 11/03/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:41</p> <p>Control Test 0.197 14:41</p> <p>Air Blank 0.000 14:42</p> <p>Control Test 0.197 14:43</p> <p>Air Blank 0.000 14:43</p> <p>Control Test 0.198 14:44</p> <p>Air Blank 0.000 14:44</p> <p>Control Test Status</p> <p>Average 0.1973</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2926</p> <p>Operator's Signature </p>	<p>DGS</p> <p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000943 11/03/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 14:46</p> <p>Control Test 0.080 14:46</p> <p>Air Blank 0.000 14:47</p> <p>Control Test 0.079 14:47</p> <p>Air Blank 0.000 14:47</p> <p>Control Test 0.079 14:48</p> <p>Air Blank 0.000 14:48</p> <p>Control Test Status</p> <p>Average 0.0793</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7277</p> <p>Operator's Signature </p>

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 11:29

Date of Inspection: 11/06/2023

Serial Number: 80-000943
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test* (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.081	0.198	0.079
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.081	0.198	0.079
0.000	0.051	0.080	0.198	0.079
0.000	0.051	0.081	0.198	0.079
0.000	0.051	0.081	0.198	0.079
0.000	0.051	0.080	0.197	0.078
0.000	0.051	0.081	0.198	0.079


Standard Deviations	0.0000	0.0005	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/06/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000943, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000943</u>	UNCERTAINTY* \pm
Owning Agency:	<u>COLLIER COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/06/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:29</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/06/2023

Date

Taylor D Gutschow

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality







INSTRUMENT PROCESSING SHEET

Agency Collier CSOS/N 80-000943Florida Department of
Law EnforcementDate In 10/12/2023 DI Completion Date 10/12/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>10/12/2023</u>	Flow Calibration	By	Date																																								
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Hand-delivered.</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>150</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # 32 mm (.139 - .169) 36 mm (.156 - .190) 53 mm (.228 - .278) 103 mm (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024																														
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5094	202201C 01/11/2024																																													
0.080	MP5095	202201D 01/18/2024																																													
0.200	MP5096	202201E 01/18/2024																																													
0.080 DGS	N/A	AG223802 08/26/2024																																													
Calibration Adjustment		By		Department Inspection			By TDG																																								
Barometric Pressure Gauge ID #				Barometric Pressure ID# <u>26932</u> Gauge <u>1007</u> Instrument <u>1008</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP4863</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP4863	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096		
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	MP4863																																														
Interferent	MP5093																																														
0.050	MP5094																																														
0.080	MP5095																																														
0.200	MP5096																																														
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				Attachments																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment		<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other																					
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Notes/Suggested Service: <u>Instrument needs DGS regulator replaced. Sending to FDLE-HQ to remove two stripped screws. Agency will provide shipping box. (TDG)</u>				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																											
Benjamin Siddoway		Digitally signed by Benjamin Siddoway Date: 2023.10.13 16:30:12 -0400		Phil Nicodemo		Digitally signed by Phil Nicodemo Date: 2023.10.18 13:38:51 -0400																																									
Tech Review / Date				Admin Review / Date																																											

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000943 10/12/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:13</p> <p>Control Test 0.050 13:13</p> <p>Air Blank 0.000 13:14</p> <p>Control Test 0.050 13:15</p> <p>Air Blank 0.000 13:15</p> <p>Control Test 0.050 13:16</p> <p>Air Blank 0.000 13:16</p> <p>Control Test Stats</p> <p>Average 0.0500</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000943 10/12/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:20</p> <p>Control Test 0.079 13:21</p> <p>Air Blank 0.000 13:21</p> <p>Control Test 0.078 13:22</p> <p>Air Blank 0.000 13:23</p> <p>Control Test 0.079 13:23</p> <p>Air Blank 0.000 13:24</p> <p>Control Test Stats</p> <p>Average 0.0787</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7339</p> <p>Operator's Signature </p>	<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000943 10/12/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:35</p> <p>Control Test 0.198 13:36</p> <p>Air Blank 0.000 13:36</p> <p>Control Test 0.199 13:37</p> <p>Air Blank 0.000 13:38</p> <p>Control Test 0.198 13:38</p> <p>Air Blank 0.000 13:39</p> <p>Control Test Stats</p> <p>Average 0.1983</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2911</p> <p>Operator's Signature </p>	<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-000943 10/12/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 12:56</p> <p>Control Test 0.079 12:58</p> <p>Air Blank 0.000 12:59</p> <p>Control Test 0.079 12:59</p> <p>Air Blank 0.000 13:00</p> <p>Control Test 0.079 13:00</p> <p>Air Blank 0.000 13:01</p> <p>Control Test Stats</p> <p>Average 0.0790</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>

DGS

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 15:38

Date of Inspection: 10/12/2023

Serial Number: 80-000943
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test* (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.078	0.199	0.079
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.079	0.198	0.078
0.000	0.049	0.079	0.199	0.078
0.000	0.049	0.079	0.199	0.077

Standard Deviations	0.0004	0.0005	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

10/12/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000943, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000943</u>	UNCERTAINTY* \pm
Owning Agency:	<u>COLLIER COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/12/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:38</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



10/12/2023

Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

INSTRUMENT PROCESSING SHEET

Agency Collier CSO

S/N 80-000943

Florida Department of
Law Enforcement

Date In 08/07/2023 DI Completion Date 08/10/2023

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 08/10/2023	Flow Calibration	By	Date																																								
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>148</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5094</td> <td>202201C</td> </tr> <tr> <td>01/11/2024</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5095</td> <td>202201D</td> </tr> <tr> <td>01/18/2024</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5096</td> <td>202201E</td> </tr> <tr> <td>01/18/2024</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG223802</td> </tr> <tr> <td>08/26/2024</td> </tr> </tbody> </table>			Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C	01/11/2024	0.080	MP5095	202201D	01/18/2024	0.200	MP5096	202201E	01/18/2024	0.080 DGS	N/A	AG223802	08/26/2024	Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____																							
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5094	202201C																																													
		01/11/2024																																													
0.080	MP5095	202201D																																													
		01/18/2024																																													
0.200	MP5096	202201E																																													
		01/18/2024																																													
0.080 DGS	N/A	AG223802																																													
		08/26/2024																																													
Calibration Adjustment By _____				Department Inspection By TDG _____																																											
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Barometric Pressure ID# <u>26932</u> Gauge <u>1015</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5092</td> </tr> <tr> <td>Interferent</td> <td>MP5093</td> </tr> <tr> <td>0.050</td> <td>MP5094</td> </tr> <tr> <td>0.080</td> <td>MP5095</td> </tr> <tr> <td>0.200</td> <td>MP5096</td> </tr> </tbody> </table>				Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	MP5092																																														
Interferent	MP5093																																														
0.050	MP5094																																														
0.080	MP5095																																														
0.200	MP5096																																														
<input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <div style="display: flex; justify-content: space-between;"> <div> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment </div> <div> <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Extra Stabilities</u> </div> </div>																							
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																											
Benjamin Siddoway				Phil Nicodemo																																											
Tech Review / Date				Admin Review / Date																																											

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-000943	Collier CSO	08/10/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/10/2023 Software: 8100.27</p> <p>SN 80-000943</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:36</p> <p>Control Test 0.050 10:37</p> <p>Air Blank 0.000 10:37</p> <p>Control Test 0.051 10:38</p> <p>Air Blank 0.000 10:38</p> <p>Control Test 0.049 10:39</p> <p>Air Blank 0.000 10:40</p> <p>Control Test Stats</p> <p>Average 0.0500</p> <p>Std Dev 0.0010</p> <p>Rel Std Dev(%) 2.0000</p> <p>MG</p> <p>Operator's Signature</p>	<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/10/2023 Software: 8100.27</p> <p>SN 80-000943</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:42</p> <p>Control Test 0.079 10:43</p> <p>Air Blank 0.000 10:44</p> <p>Control Test 0.079 10:44</p> <p>Air Blank 0.000 10:45</p> <p>Control Test 0.079 10:45</p> <p>Air Blank 0.000 10:46</p> <p>Control Test Stats</p> <p>Average 0.0790</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>MG</p> <p>Operator's Signature</p>	<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/10/2023 Software: 8100.27</p> <p>SN 80-000943</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:49</p> <p>Control Test 0.200 10:50</p> <p>Air Blank 0.000 10:50</p> <p>Control Test 0.198 10:51</p> <p>Air Blank 0.000 10:52</p> <p>Control Test 0.198 10:52</p> <p>Air Blank 0.000 10:53</p> <p>Control Test Stats</p> <p>Average 0.1987</p> <p>Std Dev 0.0012</p> <p>Rel Std Dev(%) 0.5812</p> <p>MG</p> <p>Operator's Signature</p>	<p>COLLIER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/10/2023 Software: 8100.27</p> <p>SN 80-000943</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:27</p> <p>Control Test 0.078 10:28</p> <p>Air Blank 0.000 10:28</p> <p>Control Test 0.078 10:29</p> <p>Air Blank 0.000 10:29</p> <p>Control Test 0.078 10:29</p> <p>Air Blank 0.000 10:30</p> <p>Control Test Stats</p> <p>Average 0.0780</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>MG</p> <p>Operator's Signature</p>

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 13:19

Date of Inspection: 08/10/2023

Serial Number: 80-000943
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.078	0.199	0.078 / 0.079
0.000	0.049	0.079	0.199	0.078 / 0.078
0.000	0.049	0.078	0.199	0.078 / 0.078
0.000	0.049	0.079	0.199	0.077 / 0.078
0.000	0.049	0.079	0.199	0.077 / 0.077
0.000	0.049	0.078	0.199	0.077 / 0.078
0.000	0.049	0.078	0.199	0.077 / 0.078
0.000	0.049	0.078	0.199	0.077 / 0.077
0.000	0.049	0.079	0.199	0.076 / 0.077
0.000	0.049	0.079	0.199	INT / 0.077

Standard Deviations	0.0000	0.0005	0.0000	0.0006 / 0.0006
---------------------	--------	--------	--------	-----------------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

08: Interferent Detect. UNKNOWN. AIRED OUT ROOM & REPEATED..

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

08/10/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 09:41

Date of Inspection: 08/11/2023

Serial Number: 80-000943
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined *We*
8/11/2023

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

08/11/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000943, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000943</u>	UNCERTAINTY* \pm
Owning Agency:	<u>COLLIER COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/10/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:19</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/10/2023

Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Extra Stabilities

COLLIER COUNTY SO

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000943

08/11/2023

Software: 8100.27

DGS

Test	g/210L	Time
Air Blank	0.000	09:44
Control Test	0.079	09:44
Air Blank	0.000	09:45
Control Test	0.078	09:45
Air Blank	0.000	09:45
Control Test	0.078	09:46
Air Blank	0.000	09:46
Control Test	0.078	09:47
Air Blank	0.000	09:47
Control Test	0.078	09:47
Air Blank	0.000	09:48
Control Test	0.079	09:48
Air Blank	0.000	09:49
Control Test	0.079	09:49
Air Blank	0.000	09:50
Control Test	0.079	09:50
Air Blank	0.000	09:50
Control Test	0.079	09:51
Air Blank	0.000	09:51
Control Test	0.078	09:52
Air Blank	0.000	09:52
Control Test	0.079	09:52
Air Blank	0.000	09:53
Control Test	0.079	09:53
Air Blank	0.000	09:54
Control Test	0.078	09:54
Air Blank	0.000	09:54
Control Test	0.079	09:55
Air Blank	0.000	09:55
Control Test	0.079	09:56
Air Blank	0.000	09:56
Control Test	0.079	09:56
Air Blank	0.000	09:57
Control Test	0.079	09:57
Air Blank	0.000	09:58
Control Test	0.079	09:58
Air Blank	0.000	09:59
Control Test	0.079	09:59
Air Blank	0.000	09:59
Control Test	0.078	10:00
Air Blank	0.000	10:00
Control Test Stats		
Average	0.0786	
Std Dev	0.0005	
Rel Std Dev(%)	0.6222	

Extra DGS stabilities
were run using the
same lot of dry gas
used in the Stability
Checks and Department
Inspection.

MG
08/11/2023

TDG