



INSTRUMENT PROCESSING SHEET

Agency Glades CSOS/N 80-000947Florida Department of
Law EnforcementDate In 08/23/2023 DI Completion Date 08/23/2023☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>08/23/2023</u>	Flow Calibration	By	Date																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>AI advised the analog jack is damaged. Approved to ship to CMI for repair/replacement.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>236</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____</td><td></td></tr></tbody></table>			Maintenance		By	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																					
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5094	202201C 01/11/2024																																													
0.080	MP5095	202201D 01/18/2024																																													
0.200	MP5096	202201E 01/18/2024																																													
0.080 DGS	N/A	AG223802 08/26/2024																																													
Maintenance		By																																													
☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____																																															
Calibration Adjustment		By _____		Department Inspection			By <u>TDG</u>																																								
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>26932</u> Gauge <u>1013</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6284	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096	
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	MP6284																																														
Interferent	MP5093																																														
0.050	MP5094																																														
0.080	MP5095																																														
0.200	MP5096																																														
☐ Post Calibration Adjustment Stability Checks				Attachments																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment			☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☒ Other <u>Form 51</u>																				
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Notes/Suggested Service: <u>Uploaded via direct connect due to the damaged analog jack. (TDG)</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																											
Tech Review: <u>Checked box on Form 51. (TDG 8/23/2023)</u>				Benjamin Siddoway Digitally signed by Benjamin Siddoway Date: 2023.08.23 13:46:33 -0400			Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.08.25 08:12:19 -0400																																								
				Tech Review / Date			Admin Review / Date																																								

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
≤0.003 of Wet			≤0.003 of Wet																																																																																																																																				
✓			✓																																																																																																																																				
GLADES COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/23/2023 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test</td><td>0.047</td><td>10:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0480</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>2.0833</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	10:19	Control Test	0.048	10:20	Air Blank	0.000	10:21	Control Test	0.047	10:22	Air Blank	0.000	10:23	Control Test	0.049	10:23	Air Blank	0.000	10:23	Control Test Stats			Average	0.0480		Std Dev	0.0010		Rel Std Dev(%)	2.0833		GLADES COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/23/2023 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:37</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:38</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:39</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:40</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:41</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	10:37	Control Test	0.078	10:38	Air Blank	0.000	10:38	Control Test	0.078	10:39	Air Blank	0.000	10:40	Control Test	0.078	10:41	Air Blank	0.000	10:41	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		GLADES COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/23/2023 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:50</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:53</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:54</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1993</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2896</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	10:50	Control Test	0.199	10:51	Air Blank	0.000	10:51	Control Test	0.199	10:52	Air Blank	0.000	10:53	Control Test	0.200	10:53	Air Blank	0.000	10:54	Control Test Stats			Average	0.1993		Std Dev	0.0006		Rel Std Dev(%)	0.2896		GLADES COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 08/23/2023 Software: 8100.27 Test g/210L Time <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:11</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr> <tr><td>Control Test</td><td>0.079</td><td>10:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0797</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7247</td><td></td></tr> </table> Operator's Signature 	Air Blank	0.000	10:11	Control Test	0.080	10:12	Air Blank	0.000	10:12	Control Test	0.080	10:13	Air Blank	0.000	10:13	Control Test	0.079	10:13	Air Blank	0.000	10:14	Control Test Stats			Average	0.0797		Std Dev	0.0006		Rel Std Dev(%)	0.7247	
Air Blank	0.000	10:19																																																																																																																																					
Control Test	0.048	10:20																																																																																																																																					
Air Blank	0.000	10:21																																																																																																																																					
Control Test	0.047	10:22																																																																																																																																					
Air Blank	0.000	10:23																																																																																																																																					
Control Test	0.049	10:23																																																																																																																																					
Air Blank	0.000	10:23																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0480																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel Std Dev(%)	2.0833																																																																																																																																						
Air Blank	0.000	10:37																																																																																																																																					
Control Test	0.078	10:38																																																																																																																																					
Air Blank	0.000	10:38																																																																																																																																					
Control Test	0.078	10:39																																																																																																																																					
Air Blank	0.000	10:40																																																																																																																																					
Control Test	0.078	10:41																																																																																																																																					
Air Blank	0.000	10:41																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0780																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	10:50																																																																																																																																					
Control Test	0.199	10:51																																																																																																																																					
Air Blank	0.000	10:51																																																																																																																																					
Control Test	0.199	10:52																																																																																																																																					
Air Blank	0.000	10:53																																																																																																																																					
Control Test	0.200	10:53																																																																																																																																					
Air Blank	0.000	10:54																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1993																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2896																																																																																																																																						
Air Blank	0.000	10:11																																																																																																																																					
Control Test	0.080	10:12																																																																																																																																					
Air Blank	0.000	10:12																																																																																																																																					
Control Test	0.080	10:13																																																																																																																																					
Air Blank	0.000	10:13																																																																																																																																					
Control Test	0.079	10:13																																																																																																																																					
Air Blank	0.000	10:14																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0797																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7247																																																																																																																																						

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GLADES COUNTY SO
Time of Inspection: 12:47

Date of Inspection: 08/23/2023

Serial Number: 80-000947
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.077	0.200	0.080
0.000	0.048	0.077	0.201	0.080
0.000	0.048	0.077	0.201	0.080
0.000	0.049	0.077	0.201	0.080
0.000	0.048	0.077	0.200	0.079
0.000	0.049	0.077	0.201	0.079
0.000	0.048	0.078	0.200	0.080
0.000	0.049	0.078	0.201	0.079
0.000	0.048	0.077	0.200	0.079
0.000	0.049	0.077	0.200	0.080


Standard Deviations	0.0005	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

08/23/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000947, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000947</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>GLADES COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>08/23/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:47</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

08/23/2023

Taylor D Gutschow

TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Travis Arnold on 08/23/2023

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000947

Bill To Address:

Glades County Sheriff's Office

Attn: Travis Arnold

Ship to Address:

Florida Dept. of Law Enforcement

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Analog jack at the back of the instrument is damaged.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$ _____.

TDG 8/23/2023

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Travis Arnold

Phone #: 863-508-8856

Email: TArnold@gladessheriff.org

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us