

INSTRUMENT PROCESSING SHEET

Agency Indialantic PD

S/N 80-001006





Florida Department of
Law Enforcement

Date In 10/04/2023 DI Completion Date 10/09/2023

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	Date 10/09/2023	Flow Calibration By	Date																																																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value 185 <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # ATP104 32 mm 0.164 (.139 - .169) 36 mm 0.175 (.156 - .190) 53 mm 0.257 (.228 - .278) 103 mm 0.523 (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # 26932 <input checked="" type="checkbox"/> Stability Checks		Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																													
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 25%;">Serial #</th> <th style="width: 60%;">Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP5094</td> <td>202201C</td> </tr> <tr> <td>01/11/2024</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP5095</td> <td>202201D</td> </tr> <tr> <td>01/18/2024</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP5096</td> <td>202201E</td> </tr> <tr> <td>01/18/2024</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG223802</td> </tr> <tr> <td>08/26/2024</td> </tr> </tbody> </table>		Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C	01/11/2024	0.080	MP5095	202201D	01/18/2024	0.200	MP5096	202201E	01/18/2024	0.080 DGS	N/A	AG223802	08/26/2024	Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____																																										
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP5094	202201C																																																															
		01/11/2024																																																															
0.080	MP5095	202201D																																																															
		01/18/2024																																																															
0.200	MP5096	202201E																																																															
		01/18/2024																																																															
0.080 DGS	N/A	AG223802																																																															
		08/26/2024																																																															
Calibration Adjustment			Department Inspection																																																														
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 25%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 35%;">Expiration</th> </tr> </thead> <tbody> <tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr> <tr><td>0.040</td><td></td><td></td><td></td></tr> <tr><td>0.100</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.300</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 15%;">Simulator</th> <th style="width: 25%;">Serial #</th> <th style="width: 25%;">Lot #</th> <th style="width: 35%;">Expiration</th> </tr> </thead> <tbody> <tr><td>0.050</td><td></td><td></td><td></td></tr> <tr><td>0.080</td><td></td><td></td><td></td></tr> <tr><td>0.200</td><td></td><td></td><td></td></tr> <tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr> </tbody> </table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# 26932 Gauge 1012 Instrument 1011 Mouth Alcohol Solution Lot # 2023-A Acetone Stock Solution Lot # 2022-B <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 50%;">Simulator</th> <th style="width: 50%;">Serial Number</th> </tr> </thead> <tbody> <tr><td>0.000</td><td>MP4863</td></tr> <tr><td>Interferent</td><td>MP5093</td></tr> <tr><td>0.050</td><td>MP5094</td></tr> <tr><td>0.080</td><td>MP5095</td></tr> <tr><td>0.200</td><td>MP5096</td></tr> </tbody> </table>			Simulator	Serial Number	0.000	MP4863	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	MP4863																																																																
Interferent	MP5093																																																																
0.050	MP5094																																																																
0.080	MP5095																																																																
0.200	MP5096																																																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____			Attachments <div style="display: flex; justify-content: space-between;"> <div> <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment </div> <div> <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____ </div> </div>																																																														
			<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																																														
Benjamin Siddoway <small>Digitally signed by Benjamin Siddoway Date: 2023.10.10 13:23:39 -04'00'</small>			Phil Nicodemmo <small>Digitally signed by Phil Nicodemmo Date: 2023.10.10 14:33:30 -04'00'</small>																																																														
Tech Review / Date			Admin Review / Date																																																														

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 <input checked="" type="checkbox"/> ≤0.003 of Wet <input checked="" type="checkbox"/>																																																																																																																																																
<p>INDIALANTIC PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001006 10/09/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:51</td></tr> <tr><td>Control Test</td><td>0.048</td><td>12:51</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:52</td></tr> <tr><td>Control Test</td><td>0.049</td><td>12:53</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1853</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	12:49	Control Test	0.049	12:50	Air Blank	0.000	12:51	Control Test	0.048	12:51	Air Blank	0.000	12:52	Control Test	0.049	12:53	Air Blank	0.000	12:53	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1853		<p>INDIALANTIC PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001006 10/09/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:57</td></tr> <tr><td>Control Test</td><td>0.079</td><td>12:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:58</td></tr> <tr><td>Control Test</td><td>0.078</td><td>12:59</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:59</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:00</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0783</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	12:57	Control Test	0.079	12:57	Air Blank	0.000	12:58	Control Test	0.078	12:59	Air Blank	0.000	12:59	Control Test	0.078	13:00	Air Blank	0.000	13:00	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		<p>INDIALANTIC PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001006 10/09/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>13:03</td></tr> <tr><td>Control Test</td><td>0.201</td><td>13:04</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:05</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:06</td></tr> <tr><td>Control Test</td><td>0.199</td><td>13:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:07</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0012</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.5783</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	13:03	Control Test	0.201	13:04	Air Blank	0.000	13:05	Control Test	0.199	13:05	Air Blank	0.000	13:06	Control Test	0.199	13:07	Air Blank	0.000	13:07	Control Test Stats			Average	0.1997		Std Dev	0.0012		Rel Std Dev(%)	0.5783		<p>INDIALANTIC PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001006 10/09/2023 Software: 8100.27</p> <p>DGS</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>12:43</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:44</td></tr> <tr><td>Control Test</td><td>0.080</td><td>12:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test</td><td>0.081</td><td>12:45</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	12:43	Control Test	0.080	12:43	Air Blank	0.000	12:44	Control Test	0.080	12:44	Air Blank	0.000	12:45	Control Test	0.081	12:45	Air Blank	0.000	12:45	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test	0.049	12:50																																																																																																																																																	
Air Blank	0.000	12:51																																																																																																																																																	
Control Test	0.048	12:51																																																																																																																																																	
Air Blank	0.000	12:52																																																																																																																																																	
Control Test	0.049	12:53																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1853																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:57																																																																																																																																																	
Control Test	0.079	12:57																																																																																																																																																	
Air Blank	0.000	12:58																																																																																																																																																	
Control Test	0.078	12:59																																																																																																																																																	
Air Blank	0.000	12:59																																																																																																																																																	
Control Test	0.078	13:00																																																																																																																																																	
Air Blank	0.000	13:00																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:03																																																																																																																																																	
Control Test	0.201	13:04																																																																																																																																																	
Air Blank	0.000	13:05																																																																																																																																																	
Control Test	0.199	13:05																																																																																																																																																	
Air Blank	0.000	13:06																																																																																																																																																	
Control Test	0.199	13:07																																																																																																																																																	
Air Blank	0.000	13:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1997																																																																																																																																																		
Std Dev	0.0012																																																																																																																																																		
Rel Std Dev(%)	0.5783																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:43																																																																																																																																																	
Control Test	0.080	12:43																																																																																																																																																	
Air Blank	0.000	12:44																																																																																																																																																	
Control Test	0.080	12:44																																																																																																																																																	
Air Blank	0.000	12:45																																																																																																																																																	
Control Test	0.081	12:45																																																																																																																																																	
Air Blank	0.000	12:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIALANTIC PD
Time of Inspection: 15:02

Date of Inspection: 10/09/2023

Serial Number: 80-001006
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test* (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.050	0.078	0.200	0.079
0.000	0.050	0.078	0.201	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.078	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.078	0.201	0.079


Standard Deviations	0.0004	0.0005	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

10/09/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001006, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001006</u>	UNCERTAINTY* \pm
Owning Agency:	<u>INDIALANTIC PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/09/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:02</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



10/09/2023

Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality