







## INSTRUMENT PROCESSING SHEET

Agency West Melbourne PDS/N 80-001008Florida Department of  
Law EnforcementDate In 10/26/2023 DI Completion Date 11/03/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 11/03/2023	Flow Calibration	By	Date																																																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>152</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP5095</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5096</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>					Simulator	Serial #	Lot #/Exp	0.050	MP5094	202303K 03/29/2025	0.080	MP5095	202303L 03/29/2025	0.200	MP5096	202304C 04/05/2025	0.080 DGS	N/A	AG223802 08/26/2024	<b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____ _____																																															
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP5094	202303K 03/29/2025																																																																	
0.080	MP5095	202303L 03/29/2025																																																																	
0.200	MP5096	202304C 04/05/2025																																																																	
0.080 DGS	N/A	AG223802 08/26/2024																																																																	
<b>Calibration Adjustment</b> By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>					Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<b>Department Inspection</b> By TDG _____ Barometric Pressure ID# <u>26932</u> Gauge <u>1021</u> Instrument <u>1021</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table> <b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____  <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use  Benjamin Siddoway <small>Digitally signed by Benjamin Siddoway Date: 2023.11.03 15:01:55 -0400</small> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.11.03 15:09:46 -0400</small> Tech Review / Date Admin Review / Date			Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	MP5092																																																																		
Interferent	MP5093																																																																		
0.050	MP5094																																																																		
0.080	MP5095																																																																		
0.200	MP5096																																																																		
Notes/Suggested Service: <u>Added plastic caps for return to agency. (TDG)</u> _____ _____ _____ _____ _____ _____ _____ _____ _____																																																																			

# Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083																																																																																																																																				
<p>WEST MELBOURNE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001008 11/03/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:09</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:11</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:12</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0510</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:09	Control Test	0.051	10:10	Air Blank	0.000	10:11	Control Test	0.051	10:11	Air Blank	0.000	10:12	Control Test	0.051	10:13	Air Blank	0.000	10:13	Control Test Stats			Average	0.0510		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>WEST MELBOURNE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001008 11/03/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:17</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:19</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:17	Control Test	0.080	10:18	Air Blank	0.000	10:18	Control Test	0.080	10:19	Air Blank	0.000	10:19	Control Test	0.080	10:20	Air Blank	0.000	10:20	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>WEST MELBOURNE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001008 11/03/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1990</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:24	Control Test	0.199	10:24	Air Blank	0.000	10:25	Control Test	0.199	10:26	Air Blank	0.000	10:26	Control Test	0.199	10:27	Air Blank	0.000	10:27	Control Test Stats			Average	0.1990		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>WEST MELBOURNE PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001008 11/03/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:01</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:02</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0803</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	10:01	Control Test	0.080	10:02	Air Blank	0.000	10:02	Control Test	0.081	10:02	Air Blank	0.000	10:03	Control Test	0.080	10:03	Air Blank	0.000	10:04	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Air Blank	0.000	10:09																																																																																																																																					
Control Test	0.051	10:10																																																																																																																																					
Air Blank	0.000	10:11																																																																																																																																					
Control Test	0.051	10:11																																																																																																																																					
Air Blank	0.000	10:12																																																																																																																																					
Control Test	0.051	10:13																																																																																																																																					
Air Blank	0.000	10:13																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0510																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	10:17																																																																																																																																					
Control Test	0.080	10:18																																																																																																																																					
Air Blank	0.000	10:18																																																																																																																																					
Control Test	0.080	10:19																																																																																																																																					
Air Blank	0.000	10:19																																																																																																																																					
Control Test	0.080	10:20																																																																																																																																					
Air Blank	0.000	10:20																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	10:24																																																																																																																																					
Control Test	0.199	10:24																																																																																																																																					
Air Blank	0.000	10:25																																																																																																																																					
Control Test	0.199	10:26																																																																																																																																					
Air Blank	0.000	10:26																																																																																																																																					
Control Test	0.199	10:27																																																																																																																																					
Air Blank	0.000	10:27																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1990																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	10:01																																																																																																																																					
Control Test	0.080	10:02																																																																																																																																					
Air Blank	0.000	10:02																																																																																																																																					
Control Test	0.081	10:02																																																																																																																																					
Air Blank	0.000	10:03																																																																																																																																					
Control Test	0.080	10:03																																																																																																																																					
Air Blank	0.000	10:04																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0803																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7187																																																																																																																																						



# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: WEST MELBOURNE PD  
Time of Inspection: 12:59

Date of Inspection: 11/03/2023

Serial Number: 80-001008  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test* (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.051	0.080	0.199	0.081
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.200	0.080
0.000	0.051	0.080	0.200	0.080
0.000	0.051	0.080	0.200	0.081
0.000	0.051	0.081	0.199	0.081
0.000	0.051	0.080	0.199	0.080
0.000	0.052	0.081	0.200	0.081
0.000	0.052	0.080	0.200	0.080
0.000	0.051	0.080	0.200	0.080

Standard Deviations	0.0004	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/03/2023  
Date



# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001008, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001008</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>WEST MELBOURNE PD</u>	0.050 g/210 L 0.004
Calibration Date:	<u>11/03/2023</u>	0.080 g/210 L 0.004
Calibration Time:	<u>12:59</u>	0.200 g/210 L 0.007
		0.080 g/210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

11/03/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

Service • Integrity • Respect • Quality