







INSTRUMENT PROCESSING SHEET

Agency Coral Springs PDS/N 80-001050Florida Department of
Law EnforcementDate In 10/02/2023DI Completion Date 10/10/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date	10/04/2023	Flow Calibration	By	Date																																							
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Middle right foot detached from chassis/shelf.		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>177</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks				Flow Column # <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # 32 mm (.139 - .169) 36 mm (.156 - .190) 53 mm (.228 - .278) 103 mm (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024			<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td><input type="checkbox"/> Battery Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Breath Tube Replacement</td><td></td><td></td></tr><tr><td><input type="checkbox"/> Other</td><td></td><td></td></tr></tbody></table>		Maintenance		By	<input type="checkbox"/> Battery Replacement			<input type="checkbox"/> Dry Gas Regulator Replacement			<input type="checkbox"/> Breath Tube Replacement			<input type="checkbox"/> Other												
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5094	202201C 01/11/2024																																													
0.080	MP5095	202201D 01/18/2024																																													
0.200	MP5096	202201E 01/18/2024																																													
0.080 DGS	N/A	AG223802 08/26/2024																																													
Maintenance		By																																													
<input type="checkbox"/> Battery Replacement																																															
<input type="checkbox"/> Dry Gas Regulator Replacement																																															
<input type="checkbox"/> Breath Tube Replacement																																															
<input type="checkbox"/> Other																																															
Calibration Adjustment		By TDG		Department Inspection		By TDG																																									
Barometric Pressure Gauge <u>1014</u> ID # <u>28199</u>				Barometric Pressure ID# <u>26932</u> Gauge <u>1013</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>22460</td><td>12/28/2024</td></tr><tr><td>0.100</td><td>MP5099</td><td>22310</td><td>08/11/2024</td></tr><tr><td>0.200</td><td>MP5100</td><td>22050</td><td>02/07/2024</td></tr><tr><td>0.300</td><td>MP5101</td><td>22220</td><td>06/15/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203</td><td>08/10/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	22460	12/28/2024	0.100	MP5099	22310	08/11/2024	0.200	MP5100	22050	02/07/2024	0.300	MP5101	22220	06/15/2024	0.080 DGS	N/A	AG222203	08/10/2024			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP4863</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP4863	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096		
Simulator	Serial #	Lot #	Expiration																																												
0.000	MP5097	N/A	N/A																																												
0.040	MP5098	22460	12/28/2024																																												
0.100	MP5099	22310	08/11/2024																																												
0.200	MP5100	22050	02/07/2024																																												
0.300	MP5101	22220	06/15/2024																																												
0.080 DGS	N/A	AG222203	08/10/2024																																												
Simulator	Serial Number																																														
0.000	MP4863																																														
Interferent	MP5093																																														
0.050	MP5094																																														
0.080	MP5095																																														
0.200	MP5096																																														
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks				Attachments																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C</td><td>01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D</td><td>01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E</td><td>01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802</td><td>08/26/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5094	202201C	01/11/2024	0.080	MP5095	202201D	01/18/2024	0.200	MP5096	202201E	01/18/2024	0.080 DGS	N/A	AG223802	08/26/2024			<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment		<input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other																					
Simulator	Serial #	Lot #	Expiration																																												
0.050	MP5094	202201C	01/11/2024																																												
0.080	MP5095	202201D	01/18/2024																																												
0.200	MP5096	202201E	01/18/2024																																												
0.080 DGS	N/A	AG223802	08/26/2024																																												
Notes/Suggested Service: <u>Added plastic caps for return to agency. (TDG)</u>				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																											
				Israel Soto Digitally signed by Israel Soto Date: 2023.10.12 08:28:46 -0400		Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.10.12 09:36:21 -0400																																									
				Tech Review / Date		Admin Review / Date																																									

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083
<p>CORAL SPRINGS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001050 10/04/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:21</p> <p>Control Test 0.050 10:22</p> <p>Air Blank 0.000 10:23</p> <p>Control Test 0.050 10:23</p> <p>Air Blank 0.000 10:24</p> <p>Control Test 0.050 10:24</p> <p>Air Blank 0.000 10:25</p> <p>Control Test Stats</p> <p>Average 0.0500</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>CORAL SPRINGS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001050 10/04/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:28</p> <p>Control Test 0.077 10:29</p> <p>Air Blank 0.000 10:29</p> <p>Control Test 0.078 10:30</p> <p>Air Blank 0.000 10:30</p> <p>Control Test 0.077 10:31</p> <p>Air Blank 0.000 10:32</p> <p>Control Test Stats</p> <p>Average 0.0773</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7466</p> <p>Operator's Signature </p>	<p>CORAL SPRINGS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001050 10/04/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:45</p> <p>Control Test 0.200 10:46</p> <p>Air Blank 0.000 10:46</p> <p>Control Test 0.201 10:47</p> <p>Air Blank 0.000 10:48</p> <p>Control Test 0.200 10:48</p> <p>Air Blank 0.000 10:49</p> <p>Control Test Stats</p> <p>Average 0.2003</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2882</p> <p>Operator's Signature </p>	<p>CORAL SPRINGS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001050 10/04/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 10:50</p> <p>Control Test 0.082 10:50</p> <p>Air Blank 0.000 10:51</p> <p>Control Test 0.081 10:51</p> <p>Air Blank 0.000 10:51</p> <p>Control Test 0.081 10:52</p> <p>Air Blank 0.000 10:52</p> <p>Control Test Stats</p> <p>Average 0.0813</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7099</p> <p>Operator's Signature </p>

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 1.5330 (-0.0140)
 Sample #2 = 1.5260 (0.0010)
 Sample #3 = 1.5250 (0.0020)
 Sample #4 = 1.5160 (0.0130)
 Avg % Abs = 1.5223 (0.0053)
 STD DEV = 0.0055 (0.0067)
 REL STD DEV = 0.362 (124.844)

CORAL SPRINGS PD
 Intoxilyzer - Alcohol Analyzer
 Model 8000
 10/10/2023 10:16:19
 SN 80-001050

Auto Calibration
 Max Power Res Value = 86
 Auto Range Res Value = 55

Sol Value = 0.000 g/210L ***
 Fit Value = 0.0000 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12474, Sum Io = 13014
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1780 (-0.0240)
 Sample #2 = 0.1340 (0.0710)
 Sample #3 = 0.1130 (0.1350)
 Sample #4 = 0.0860 (0.2100)
 Avg % Abs = 0.1110 (0.1387)
 STD DEV = 0.0241 (0.0656)
 REL STD DEV = 0.2100

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.1830 (-0.0100)
 Sample #2 = 0.1560 (0.0280)
 Sample #3 = 0.1660 (0.0490)
 Sample #4 = 0.1350 (0.0880)
 Avg % Abs = 0.1523 (0.0550)
 STD DEV = 0.0158 (0.0304)
 REL STD DEV = 10.386 (55.358)

Sol Value = 0.040 g/210L ***
 Fit Value = 0.1905 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12396, Sum Io = 12961
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 0.8620 (-0.0240)
 Sample #2 = 0.8550 (0.0000)
 Sample #3 = 0.8290 (0.0420)
 Sample #4 = 0.8070 (0.0640)
 Avg % Abs = 0.8303 (0.0353)
 STD DEV = 0.0240 (0.0325)
 REL STD DEV = 2.894 (92.028)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 6.7650 (-0.0060)
 Sample #2 = 6.8360 (0.0130)
 Sample #3 = 6.8810 (-0.0060)
 Sample #4 = 6.8600 (0.0220)
 Avg % Abs = 6.8590 (0.0097)
 STD DEV = 0.0225 (0.0143)
 REL STD DEV = 0.328 (147.874)

<<<< CHANNEL 2 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 9.9550 (-0.0170)
 Sample #2 = 9.9370 (-0.0070)
 Sample #3 = 9.9400 (0.0000)
 Sample #4 = 9.9500 (0.0120)
 Avg % Abs = 9.9423 (0.0017)
 STD DEV = 0.0068 (0.0096)
 REL STD DEV = 0.068 (576.541)

Sol Value = 0.200 g/210L ***
 Fit Value = 0.9524 mg/l %%%
 Samples Taken = 4, Discarded = 1
 Sum Io = 12364, Sum Io = 12945
 <<<< CHANNEL 1 >>>>
 Sample % Abs (% Abs Ref)
 Sample #1 = 3.6000 (-0.0090)
 Sample #2 = 3.6570 (0.0080)
 Sample #3 = 3.6610 (0.0040)
 Sample #4 = 3.6580 (0.0460)
 Avg % Abs = 3.6587 (0.0193)
 STD DEV = 0.0021 (0.0232)
 REL STD DEV = 0.057 (119.899)

Optical Calibration Adjustment

By: TDG

***** AUTO CAL DATA *****
 <<<< CHANNEL 1 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.111
 Std Dev = 0.02 Rel Std Dev = 21.68
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.830
 Std Dev = 0.02 Rel Std Dev = 2.89
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.879
 Std Dev = 0.00 Rel Std Dev = 0.19
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.659
 Std Dev = 0.00 Rel Std Dev = 0.06
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.359
 Std Dev = 0.00 Rel Std Dev = 0.04
 Zero Order Coef = -285.55
 First Order Coef = 2636.46
 Second Order Coef = 14.94
 Standard Deviation = 28.691090

<<<< CHANNEL 2 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.152
 Std Dev = 0.02 Rel Std Dev = 10.39
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 1.522
 Std Dev = 0.01 Rel Std Dev = 0.36
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 3.536
 Std Dev = 0.01 Rel Std Dev = 0.38
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 6.859
 Std Dev = 0.02 Rel Std Dev = 0.33
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 9.942
 Std Dev = 0.01 Rel Std Dev = 0.07
 Zero Order Coef = -187.21
 First Order Coef = 1352.41
 Second Order Coef = 10.21
 Standard Deviation = 32.940388

<<<< CHANNEL 1 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.000
 Std Dev = 0.000
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.830
 Std Dev = 0.009
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.879
 Std Dev = 0.001
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.659
 Std Dev = 0.001
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.359
 Std Dev = 0.001





***** AUTO CAL DATA *****
 <<<< CHANNEL 1 >>>>
 Sol Val = 0.0000 mg/l or 0.000 g/210L
 % Abs = 0.111
 Std Dev = 0.02 Rel Std Dev = 21.68
 Sol Val = 0.1905 mg/l or 0.040 g/210L
 % Abs = 0.830
 Std Dev = 0.02 Rel Std Dev = 2.89
 Sol Val = 0.4762 mg/l or 0.100 g/210L
 % Abs = 1.879
 Std Dev = 0.00 Rel Std Dev = 0.19
 Sol Val = 0.9524 mg/l or 0.200 g/210L
 % Abs = 3.659
 Std Dev = 0.00 Rel Std Dev = 0.06
 Sol Val = 1.4286 mg/l or 0.300 g/210L
 % Abs = 5.359
 Std Dev = 0.00 Rel Std Dev = 0.04
 Zero Order Coef = -285.55
 First Order Coef = 2636.46
 Second Order Coef = 14.94
 Standard Deviation = 28.691090

***** CHANNEL 1 *****
 Sample #1 = 3240.00
 Sample #2 = 3230.00
 Sample #3 = 3153.00
 Sample #4 = 3235.00
 Average Result = 3206.0000
 STD DEV = 45.9674
 REL STD DEV = 1.434

 ***** CHANNEL 2 *****
 Sample #1 = 3405.00
 Sample #2 = 3394.00
 Sample #3 = 3361.00
 Sample #4 = 3387.00
 Average Result = 3380.6667
 STD DEV = 17.3877
 REL STD DEV = 0.514

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1014
 3 um H2O Adjust (mg/l*10,000) = 603
 9 um H2O Adjust (mg/l*10,000) = 429
 ***** AUTO CAL PASS *****

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<p>CORAL SPRINGS PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-001050</p>	<p>CORAL SPRINGS PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-001050</p>	<p>CORAL SPRINGS PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-001050</p>	<p>CORAL SPRINGS PD Intoxilyzer - Alcohol Analyzer Model 8000 10/10/2023 Software: 8100.27</p> <p>SN 80-001050</p>																																																																																																																																																
<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	11:28	Control Test	0.049	11:29	Air Blank	0.000	11:29	Control Test	0.048	11:30	Air Blank	0.000	11:30	Control Test	0.049	11:31	Air Blank	0.000	11:32	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:36</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:38</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:38</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	11:35	Control Test	0.078	11:36	Air Blank	0.000	11:36	Control Test	0.078	11:37	Air Blank	0.000	11:38	Control Test	0.078	11:38	Air Blank	0.000	11:39	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:42</td></tr> <tr><td>Control Test</td><td>0.199</td><td>11:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:44</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:45</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:46</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2911</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	11:42	Control Test	0.199	11:43	Air Blank	0.000	11:44	Control Test	0.198	11:44	Air Blank	0.000	11:45	Control Test	0.198	11:46	Air Blank	0.000	11:46	Control Test Stats			Average	0.1983		Std Dev	0.0006		Rel Std Dev(%)	0.2911		<p>DGS</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:24</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> <p>Operator's Signature </p>	Test	g/210L	Time	Air Blank	0.000	11:22	Control Test	0.078	11:23	Air Blank	0.000	11:23	Control Test	0.079	11:23	Air Blank	0.000	11:24	Control Test	0.079	11:24	Air Blank	0.000	11:25	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:28																																																																																																																																																	
Control Test	0.049	11:29																																																																																																																																																	
Air Blank	0.000	11:29																																																																																																																																																	
Control Test	0.048	11:30																																																																																																																																																	
Air Blank	0.000	11:30																																																																																																																																																	
Control Test	0.049	11:31																																																																																																																																																	
Air Blank	0.000	11:32																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:35																																																																																																																																																	
Control Test	0.078	11:36																																																																																																																																																	
Air Blank	0.000	11:36																																																																																																																																																	
Control Test	0.078	11:37																																																																																																																																																	
Air Blank	0.000	11:38																																																																																																																																																	
Control Test	0.078	11:38																																																																																																																																																	
Air Blank	0.000	11:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:42																																																																																																																																																	
Control Test	0.199	11:43																																																																																																																																																	
Air Blank	0.000	11:44																																																																																																																																																	
Control Test	0.198	11:44																																																																																																																																																	
Air Blank	0.000	11:45																																																																																																																																																	
Control Test	0.198	11:46																																																																																																																																																	
Air Blank	0.000	11:46																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1983																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2911																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:22																																																																																																																																																	
Control Test	0.078	11:23																																																																																																																																																	
Air Blank	0.000	11:23																																																																																																																																																	
Control Test	0.079	11:23																																																																																																																																																	
Air Blank	0.000	11:24																																																																																																																																																	
Control Test	0.079	11:24																																																																																																																																																	
Air Blank	0.000	11:25																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0787																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7339																																																																																																																																																		

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: CORAL SPRINGS PD
Time of Inspection: 14:29

Date of Inspection: 10/10/2023

Serial Number: 80-001050
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/18/2024 <i>11 TDC 10/10/23</i>	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.077	0.195	0.080
0.000	0.049	0.077	0.196	0.080
0.000	0.049	0.078	0.196	0.080
0.000	0.049	0.077	0.197	0.080
0.000	0.049	0.077	0.197	0.080
0.000	0.048	0.077	0.199	0.080
0.000	0.049	0.077	0.197	0.079
0.000	0.049	0.077	0.196	0.080
0.000	0.049	0.078	0.197	0.080
0.000	0.049	0.077	0.196	0.080

Standard Deviations	0.0003	0.0004	0.0010	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0005 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

10/10/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001050, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001050</u>	UNCERTAINTY* \pm
Owning Agency:	<u>CORAL SPRINGS PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/10/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:29</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



10/10/2023

Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality