



INSTRUMENT PROCESSING SHEET

Agency New Port Richey Police DepartmentS/N 80-001064Florida Department of
Law EnforcementDate In 06-07-2023DI Completion Date 06-29-2023☐ Ship ☐ P/U ☐ H/D ☐ CMI ☒ EE

Intake	By IS	Quality Checks	By IS	Date <u>06-29-2023</u>	Flow Calibration	By	Date																												
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Instrument would not power on upon arrival.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>196</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																														
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202303K <u>003-29-2025</u></td></tr><tr><td>0.080</td><td>MP6292</td><td>^{IS 6-29-2023} 202303L <u>03-29-2025</u></td></tr><tr><td>0.200</td><td>MP6293</td><td>202304C <u>04-05-2025</u></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 <u>10-25-2024</u></td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202303K <u>003-29-2025</u>	0.080	MP6292	^{IS 6-29-2023} 202303L <u>03-29-2025</u>	0.200	MP6293	202304C <u>04-05-2025</u>	0.080 DGS	N/A	AG229803 <u>10-25-2024</u>	Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____															
Simulator	Serial #	Lot #/Exp																																	
0.050	MP6291	202303K <u>003-29-2025</u>																																	
0.080	MP6292	^{IS 6-29-2023} 202303L <u>03-29-2025</u>																																	
0.200	MP6293	202304C <u>04-05-2025</u>																																	
0.080 DGS	N/A	AG229803 <u>10-25-2024</u>																																	
Calibration Adjustment By _____				Department Inspection By <u>IS</u>																															
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28421</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1013</u> Instrument <u>1008</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
☐ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293																
Simulator	Serial Number																																		
0.000	MP6289																																		
Interferent	MP6290																																		
0.050	MP6291																																		
0.080	MP6292																																		
0.200	MP6293																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☒ Other <u>Form 51</u>											
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: <u>Instrument powered on 6-29-2023. Department Inspection was completed. Instrument will be sent to Enforcement Electronics. IS</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																															
Tech Review: <u>Removed extra "0" from expiration date under Quality Checks IS 06-29-2023</u>				Tech Review / Date _____ Admin Review / Date _____																															
Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2023.06.29 14:54:06 -04'00'				Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.06.30 15:03:18 -04'00'																															

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: NEW PORT RICHEY PD
Time of Inspection: 14:17

Date of Inspection: 06/29/2023

Serial Number: 80-001064
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.050	0.079	0.196	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.048	0.079	0.196	0.077
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.078	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.078

Standard Deviations	0.0004	0.0003	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

06/29/2023
Date

Stability checks

NEW PORT RICHEY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001064
06/29/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:24
Control Test	0.048	12:24
Air Blank	0.000	12:25
Control Test	0.050	12:26
Air Blank	0.000	12:26
Control Test	0.050	12:27
Air Blank	0.000	12:27
Control Test Stats		
Average	0.0493	
Std Dev	0.0012	
Rel Std Dev(%)	2.3406	


Operator's Signature

NEW PORT RICHEY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001064
06/29/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:28
Control Test	0.079	12:29
Air Blank	0.000	12:29
Control Test	0.079	12:30
Air Blank	0.000	12:30
Control Test	0.079	12:31
Air Blank	0.000	12:32
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet


Operator's Signature

NEW PORT RICHEY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001064
06/29/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:33
Control Test	0.197	12:34
Air Blank	0.000	12:35
Control Test	0.195	12:35
Air Blank	0.000	12:36
Control Test	0.196	12:37
Air Blank	0.000	12:37
Control Test Stats		
Average	0.1960	
Std Dev	0.0010	
Rel Std Dev(%)	0.5102	


Operator's Signature

NEW PORT RICHEY PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001064
06/29/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:38
Control Test	0.078	12:38
Air Blank	0.000	12:39
Control Test	0.079	12:39
Air Blank	0.000	12:40
Control Test	0.078	12:40
Air Blank	0.000	12:41
Control Test Stats		
Average	0.0783	
Std Dev	0.0006	
Rel Std Dev(%)	0.7370	

Dry


Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001064 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001064</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>NEW PORT RICHEY PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/29/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:17</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2023.06.29 14:30:10 -04'00'

06/29/2023

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to: ☐ CMI, Inc.

☒ Enforcement Electronics

Shipment to repair facility authorized by: Jason Engel on 06-29-2023

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001064

Bill To Address:

City of New Port Richey Police Dept

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Dept. of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Instrument is having trouble powering on.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Jason Engel

Phone #: 727-858-9535

Email: engelj@cityofnewportrichey.org

ATP Contact Name: Israel Soto

ATP Email: israelsoto@fdle.state.fl.us