

Flow Cal Adjust

80-001077

MB

HARDEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001077
10/12/2023
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff) = 7.070

2: Rate (Liters/min) = 15
SQRT(Diff) = 12.246

3: Rate (Liters/min) = 30
SQRT(Diff) = 21.141

Dependent Data Scale Factor = 100000 L/min





Independent Data Scale Factor = 256

Rounded Slope = 690

Rounded Intercept = -716191

Correlation = 0.99934

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<input checked="" type="checkbox"/> ≤0.003 of Wet			
<p>HARDEE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001077 10/12/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:04</p> <p>Control Test 0.050 13:04</p> <p>Air Blank 0.000 13:05</p> <p>Control Test 0.050 13:05</p> <p>Air Blank 0.000 13:06</p> <p>Control Test 0.049 13:07</p> <p>Air Blank 0.000 13:07</p> <p>Control Test Stats</p> <p>Average 0.0497</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1625</p> <p>Operator's Signature </p>	<p>HARDEE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001077 10/12/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:15</p> <p>Control Test 0.078 13:15</p> <p>Air Blank 0.000 13:17</p> <p>Control Test 0.078 13:17</p> <p>Air Blank 0.000 13:18</p> <p>Control Test 0.078 13:18</p> <p>Air Blank 0.000 13:19</p> <p>Control Test Stats</p> <p>Average 0.0780</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>HARDEE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001077 10/12/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:30</p> <p>Control Test 0.199 13:31</p> <p>Air Blank 0.000 13:31</p> <p>Control Test 0.199 13:32</p> <p>Air Blank 0.000 13:33</p> <p>Control Test 0.198 13:33</p> <p>Air Blank 0.000 13:34</p> <p>Control Test Stats</p> <p>Average 0.1987</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2906</p> <p>Operator's Signature </p>	<p>HARDEE COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-001077 10/12/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:35</p> <p>Control Test 0.077 13:35</p> <p>Air Blank 0.000 13:36</p> <p>Control Test 0.079 13:36</p> <p>Air Blank 0.000 13:37</p> <p>Control Test 0.079 13:37</p> <p>Air Blank 0.000 13:37</p> <p>Control Test Stats</p> <p>Average 0.0783</p> <p>Std Dev 0.0012</p> <p>Rel Std Dev(%) 1.4741</p> <p>Operator's Signature </p>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 15:11

Date of Inspection: 10/12/2023

Serial Number: 80-001077
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:AG223802 Exp: 08/26/2024	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test* (g/210L) Lot#:_____ Exp:_____
0.000	0.050	0.000 / 0.000		
0.000	0.050	0.000 / 0.000		
0.000	0.050	0.000 / 0.000		
0.000	0.049	0.000 / 0.000		
0.000	0.050	0.000 / 0.000		
0.000	0.049	0.000 / 0.000		
0.000	0.050	0.000 / 0.000		
0.000	0.050	0.000 / 0.000		
0.000	0.049	0.000 / 0.000		
0.000	0.050	0.000 / 0.000		

Standard Deviations	0.0004	0.0000 / 0.0000		
---------------------	--------	-----------------	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

After Test #1, I checked the D65 psi and connections. Prior to repeating, I replaced the cylinder with another of the same lot with a higher psi (~1,000). After Test #2, I realized I had inadvertently run a 0.08 D65 Test instead of a 0.08 ARS Test. Recorded 0.08 lot info is for D65 used. Will repeat.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

10/12/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: HARDEE COUNTY SO
Time of Inspection: 15:00

Date of Inspection: 10/13/2023

Serial Number: 80-001077
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.080
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.079

Standard Deviations	0.0003	0.0000	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

10/13/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001077, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001077</u>	UNCERTAINTY* \pm
Owning Agency:	<u>HARDEE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/13/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>15:00</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/13/2023

Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality