

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BOYNTON BEACH PD
Time of Inspection: 09:37

Date of Inspection: 07/17/2023

Serial Number: 80-001190
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

Not determined *7/17/2023*

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow TAYLOR D GUTSCHOW
Signature and Printed Name

07/17/2023
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001190	Boynton Beach PD	07/17/2023	TDG MK

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																												
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																												
<p>BOYNTON BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 07/17/2023 SN 80-001190 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:18</td></tr> <tr><td>Control Test</td><td>0.049</td><td>10:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test</td><td>0.048</td><td>10:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0483</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>1.1945</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>MK</i></p>	Test	g/210L	Time	Air Blank	0.000	10:18	Control Test	0.049	10:19	Air Blank	0.000	10:20	Control Test	0.048	10:20	Air Blank	0.000	10:21	Control Test	0.048	10:22	Air Blank	0.000	10:22	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel. Std Dev(%)	1.1945		<p>BOYNTON BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 07/17/2023 SN 80-001190 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:25</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:26</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:27</td></tr> <tr><td>Control Test</td><td>0.078</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>MK</i></p>	Test	g/210L	Time	Air Blank	0.000	10:25	Control Test	0.078	10:26	Air Blank	0.000	10:26	Control Test	0.078	10:27	Air Blank	0.000	10:27	Control Test	0.078	10:28	Air Blank	0.000	10:29	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel. Std Dev(%)	0.0000		<p>BOYNTON BEACH PD Intoxilyzer - Alcohol Analyzer Model: 8000 07/17/2023 SN 80-001190 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:32</td></tr> <tr><td>Control Test</td><td>0.197</td><td>10:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1983</td><td></td></tr> <tr><td>Std Dev</td><td>0.0015</td><td></td></tr> <tr><td>Rel. Std Dev(%)</td><td>0.7702</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>MK</i></p>	Test	g/210L	Time	Air Blank	0.000	10:31	Control Test	0.200	10:32	Air Blank	0.000	10:32	Control Test	0.197	10:33	Air Blank	0.000	10:33	Control Test	0.198	10:34	Air Blank	0.000	10:35	Control Test Stats			Average	0.1983		Std Dev	0.0015		Rel. Std Dev(%)	0.7702		<p>DGS 0.08g/210L</p> <p>0.077 to 0.083</p> <p>≤0.003 of Wet</p> <p>Inadvertently printed to external printer. Will attach Results in normal range.</p> <p>MK 7/17/2023</p>
Test	g/210L	Time																																																																																																													
Air Blank	0.000	10:18																																																																																																													
Control Test	0.049	10:19																																																																																																													
Air Blank	0.000	10:20																																																																																																													
Control Test	0.048	10:20																																																																																																													
Air Blank	0.000	10:21																																																																																																													
Control Test	0.048	10:22																																																																																																													
Air Blank	0.000	10:22																																																																																																													
Control Test Stats																																																																																																															
Average	0.0483																																																																																																														
Std Dev	0.0006																																																																																																														
Rel. Std Dev(%)	1.1945																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	10:25																																																																																																													
Control Test	0.078	10:26																																																																																																													
Air Blank	0.000	10:26																																																																																																													
Control Test	0.078	10:27																																																																																																													
Air Blank	0.000	10:27																																																																																																													
Control Test	0.078	10:28																																																																																																													
Air Blank	0.000	10:29																																																																																																													
Control Test Stats																																																																																																															
Average	0.0780																																																																																																														
Std Dev	0.0000																																																																																																														
Rel. Std Dev(%)	0.0000																																																																																																														
Test	g/210L	Time																																																																																																													
Air Blank	0.000	10:31																																																																																																													
Control Test	0.200	10:32																																																																																																													
Air Blank	0.000	10:32																																																																																																													
Control Test	0.197	10:33																																																																																																													
Air Blank	0.000	10:33																																																																																																													
Control Test	0.198	10:34																																																																																																													
Air Blank	0.000	10:35																																																																																																													
Control Test Stats																																																																																																															
Average	0.1983																																																																																																														
Std Dev	0.0015																																																																																																														
Rel. Std Dev(%)	0.7702																																																																																																														

Comments:

BOYNTON BEACH PD

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-001190

07/17/2023

Software: 8100.27

DGS

Test	g/210L	Time
Air Blank	0.000	10:09
Control Test	0.080	10:09
Air Blank	0.000	10:10
Control Test	0.079	10:10
Air Blank	0.000	10:10
Control Test	0.079	10:11
Air Blank	0.000	10:11
Control Test Stats		
Average	0.0793	
Std Dev	0.0006	
Rel Std Dev(%)	0.7277	

ML

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BOYNTON BEACH PD
Time of Inspection: 12:20

Date of Inspection: 07/17/2023

Serial Number: 80-001190
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.079	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.078
0.000	0.048	0.078	0.199	0.078
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.078
0.000	0.048	0.079	0.199	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.079	0.200	0.078
Standard Deviations	0.0005	0.0005	0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/17/2023
Date



Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

Calibration Certificate

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001190, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001190</u>	UNCERTAINTY* ±	
Owning Agency:	<u>BOYNTON BEACH PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/17/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:20</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/17/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality