



INSTRUMENT PROCESSING SHEET

Agency Melbourne PDS/N 80-001262Florida Department of
Law EnforcementDate In 12/27/2023 DI Completion Date 12/27/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By BS	Quality Checks	By BS	Date <u>12/27/2023</u>	Flow Calibration	By	Date																																										
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>188</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.523</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K 3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 4/5/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG325603 9/13/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 3/29/2025	0.080	MP5089	202303L 3/29/2025	0.200	MP5090	202304C 4/5/2025	0.080 DGS	N/A	AG325603 9/13/2025	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td><td></td></tr><tr><td>☐ Other _____</td><td></td><td></td></tr><tr><td>_____</td><td></td><td></td></tr><tr><td>_____</td><td></td><td></td></tr><tr><td>_____</td><td></td><td></td></tr><tr><td>_____</td><td></td><td></td></tr></tbody></table>			Maintenance		By	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other _____			_____			_____			_____			_____		
Simulator	Serial #	Lot #/Exp																																															
0.050	MP5088	202303K 3/29/2025																																															
0.080	MP5089	202303L 3/29/2025																																															
0.200	MP5090	202304C 4/5/2025																																															
0.080 DGS	N/A	AG325603 9/13/2025																																															
Maintenance		By																																															
☐ Battery Replacement																																																	
☐ Dry Gas Regulator Replacement																																																	
☐ Breath Tube Replacement																																																	
☐ Other _____																																																	

Calibration Adjustment		By _____		Department Inspection			By BS _____																																										
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28427</u> Gauge <u>1012</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2023-B</u>																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090			
Simulator	Serial #	Lot #	Expiration																																														
0.000		N/A	N/A																																														
0.040																																																	
0.100																																																	
0.200																																																	
0.300																																																	
0.080 DGS	N/A																																																
Simulator	Serial Number																																																
0.000	MP5086																																																
Interferent	MP5087																																																
0.050	MP5088																																																
0.080	MP5089																																																
0.200	MP5090																																																
☐ Post Calibration Adjustment Stability Checks				Attachments																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table>			☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☐ Other _____															
Simulator	Serial #	Lot #	Expiration																																														
0.050																																																	
0.080																																																	
0.200																																																	
0.080 DGS	N/A																																																
☒ Form 41	☐ Post-Stability Checks																																																
☒ Stability Checks	☐ Flow Calibration																																																
☒ Calibration Certificate	☐ Form 40																																																
☐ Calibration Adjustment	☐ Other _____																																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																													
		Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2024.01.03 12:18:52 -05'00'		Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.01.04 15:36:59 -05'00'																																													
		Tech Review / Date		Admin Review / Date																																													

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MELBOURNE PD

Time of Inspection: 12:34

Date of Inspection: 12/27/2023

Serial Number: 80-001262

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG325603 Exp: 09/13/2025
0.000	0.050	0.079	0.196	0.080
0.000	0.050	0.079	0.196	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.196	0.078
0.000	0.049	0.079	0.197	0.079
0.000	0.050	0.079	0.196	0.078
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.079	0.196	0.078
0.000	0.050	0.079	0.196	0.078


Standard Deviations	0.0003	0.0000	0.0005	0.0006
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BENJAMIN W SIDDOWAY
Signature and Printed Name

12/27/2023
Date

Stability Checks

MELBOURNE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001262
12/27/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.050	09:58
Air Blank	0.000	09:59
Control Test	0.050	09:59
Air Blank	0.000	10:00
Control Test	0.050	10:00
Air Blank	0.000	10:01
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Siddle
Operator's Signature

MELBOURNE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001262
12/27/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:05
Control Test	0.079	10:06
Air Blank	0.000	10:06
Control Test	0.078	10:07
Air Blank	0.000	10:08
Control Test	0.079	10:08
Air Blank	0.000	10:09
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Benjamin Siddle
Operator's Signature

MELBOURNE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001262
12/27/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:51
Control Test	0.197	09:52
Air Blank	0.000	09:52
Control Test	0.196	09:53
Air Blank	0.000	09:53
Control Test	0.197	09:54
Air Blank	0.000	09:55
Control Test Stats		
Average	0.1967	
Std Dev	0.0006	
Rel Std Dev(%)	0.2936	

Benjamin Siddle
Operator's Signature

MELBOURNE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001262
12/27/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.080	10:21
Air Blank	0.000	10:21
Control Test	0.080	10:22
Air Blank	0.000	10:22
Control Test	0.080	10:22
Air Blank	0.000	10:23
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Benjamin Siddle
Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001262, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001262</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MELBOURNE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/27/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:34</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Benjamin
Siddoway

Digitally signed by Benjamin Siddoway
Date: 2024.01.03 10:01:20 -05'00'

12/27/2023

Date

BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality