



INSTRUMENT PROCESSING SHEET

Agency Okeechobee CSOS/N 80-001321Florida Department of
Law EnforcementDate In 03/06/2023DI Completion Date 04/28/2023☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>04/24/2023</u>	Flow Calibration	By	Date																											
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Static apron (not bag)</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>214</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP4864</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP6288</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6286	202201C 01/11/2024	0.080	MP4864	202201D 01/18/2024	0.200	MP6288	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024			Maintenance By TDG <input checked="" type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ <u>Changed battery on 4/14. Noticed the internal speaker was detached from interior wall. Will send to CMI to reattach. (TDG)</u>														
Simulator	Serial #	Lot #/Exp																																
0.050	MP6286	202201C 01/11/2024																																
0.080	MP4864	202201D 01/18/2024																																
0.200	MP6288	202201E 01/18/2024																																
0.080 DGS	N/A	AG223802 08/26/2024																																
Calibration Adjustment By TDG		Department Inspection By TDG																																
Barometric Pressure Gauge <u>1015/1014</u> ID # <u>28199</u>		Barometric Pressure ID# <u>26932</u>																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP4864</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>21410</td><td>09/30/2023</td></tr><tr><td>0.100</td><td>MP5099</td><td>22310</td><td>08/11/2024</td></tr><tr><td>0.200</td><td>MP5100</td><td>22050</td><td>02/07/2024</td></tr><tr><td>0.300</td><td>MP5101</td><td>22220</td><td>06/15/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG115904</td><td>06/08/2023</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP4864	N/A	N/A	0.040	MP5098	21410	09/30/2023	0.100	MP5099	22310	08/11/2024	0.200	MP5100	22050	02/07/2024	0.300	MP5101	22220	06/15/2024	0.080 DGS	N/A	AG115904	06/08/2023	Gauge <u>1012</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>				
Simulator	Serial #	Lot #	Expiration																															
0.000	MP4864	N/A	N/A																															
0.040	MP5098	21410	09/30/2023																															
0.100	MP5099	22310	08/11/2024																															
0.200	MP5100	22050	02/07/2024																															
0.300	MP5101	22220	06/15/2024																															
0.080 DGS	N/A	AG115904	06/08/2023																															
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096	Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks (x1) <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment (x2) <input checked="" type="checkbox"/> Other <u>Form 51</u>																		
Simulator	Serial Number																																	
0.000	MP5092																																	
Interferent	MP5093																																	
0.050	MP5094																																	
0.080	MP5095																																	
0.200	MP5096																																	
Notes/Suggested Service: <u>Checked breath tube screen and replaced o-rings on 3/7. Finished Quality Checks on 4/24. (TDG)</u>		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																
During the optical calibration, the top portion of data did not print on the internal printer slip. Repeated the optical using the same gauge, solutions, and dry gas. Did not conduct post-cal stabilities on Optical #1. (TDG)		Israel Soto <small>Digitally signed by Israel Soto Date: 2023.05.01 14:30:08 +04'00'</small>		Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.05.10 12:41:10 -04'00'</small>																														
		Tech Review / Date		Admin Review / Date																														

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE COUNTY SO
Time of Inspection: 10:15

Date of Inspection: 04/24/2023

Serial Number: 80-001321
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT.

Not determined

76
04/24/2023

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.







TAYLOR D GUTSCHOW

Signature and Printed Name

04/24/2023
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001321	Okeechobee CSD	04/24/2023	TDG ML

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L																																																																																																																																																	
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083																																																																																																																																																	
✓		✓			✓		✗																																																																																																																																																
<p>OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 04/24/2023 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:07</td></tr><tr><td>Control Test</td><td>0.052</td><td>12:08</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:08</td></tr><tr><td>Control Test</td><td>0.052</td><td>12:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:10</td></tr><tr><td>Control Test</td><td>0.052</td><td>12:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:11</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0520</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <p>Operator's Signature </p>		Test	g/210L	Time	Air Blank	0.000	12:07	Control Test	0.052	12:08	Air Blank	0.000	12:08	Control Test	0.052	12:09	Air Blank	0.000	12:10	Control Test	0.052	12:10	Air Blank	0.000	12:11	Control Test Stats			Average	0.0520		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 04/24/2023 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:12</td></tr><tr><td>Control Test</td><td>0.081</td><td>12:13</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:13</td></tr><tr><td>Control Test</td><td>0.081</td><td>12:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:14</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:16</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0807</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr></tbody></table> <p>Operator's Signature </p>		Test	g/210L	Time	Air Blank	0.000	12:12	Control Test	0.081	12:13	Air Blank	0.000	12:13	Control Test	0.081	12:14	Air Blank	0.000	12:14	Control Test	0.080	12:15	Air Blank	0.000	12:16	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157		<p>OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 04/24/2023 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:20</td></tr><tr><td>Control Test</td><td>0.201</td><td>12:20</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:21</td></tr><tr><td>Control Test</td><td>0.202</td><td>12:22</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:22</td></tr><tr><td>Control Test</td><td>0.202</td><td>12:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:23</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.2017</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2863</td><td></td></tr></tbody></table> <p>Operator's Signature </p>		Test	g/210L	Time	Air Blank	0.000	12:20	Control Test	0.201	12:20	Air Blank	0.000	12:21	Control Test	0.202	12:22	Air Blank	0.000	12:22	Control Test	0.202	12:23	Air Blank	0.000	12:23	Control Test Stats			Average	0.2017		Std Dev	0.0006		Rel Std Dev(%)	0.2863		<p>OKEECHOBEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model: 8000 04/24/2023 Software: 8100.27</p> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:25</td></tr><tr><td>Control Test</td><td>0.076</td><td>12:25</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr><tr><td>Control Test</td><td>0.077</td><td>12:26</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:26</td></tr><tr><td>Control Test</td><td>0.077</td><td>12:27</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:27</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0767</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7531</td><td></td></tr></tbody></table> <p>Operator's Signature </p>		Test	g/210L	Time	Air Blank	0.000	12:25	Control Test	0.076	12:25	Air Blank	0.000	12:26	Control Test	0.077	12:26	Air Blank	0.000	12:26	Control Test	0.077	12:27	Air Blank	0.000	12:27	Control Test Stats			Average	0.0767		Std Dev	0.0006		Rel Std Dev(%)	0.7531	
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	12:07																																																																																																																																																					
Control Test	0.052	12:08																																																																																																																																																					
Air Blank	0.000	12:08																																																																																																																																																					
Control Test	0.052	12:09																																																																																																																																																					
Air Blank	0.000	12:10																																																																																																																																																					
Control Test	0.052	12:10																																																																																																																																																					
Air Blank	0.000	12:11																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0520																																																																																																																																																						
Std Dev	0.0000																																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	12:12																																																																																																																																																					
Control Test	0.081	12:13																																																																																																																																																					
Air Blank	0.000	12:13																																																																																																																																																					
Control Test	0.081	12:14																																																																																																																																																					
Air Blank	0.000	12:14																																																																																																																																																					
Control Test	0.080	12:15																																																																																																																																																					
Air Blank	0.000	12:16																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0807																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.7157																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	12:20																																																																																																																																																					
Control Test	0.201	12:20																																																																																																																																																					
Air Blank	0.000	12:21																																																																																																																																																					
Control Test	0.202	12:22																																																																																																																																																					
Air Blank	0.000	12:22																																																																																																																																																					
Control Test	0.202	12:23																																																																																																																																																					
Air Blank	0.000	12:23																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.2017																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.2863																																																																																																																																																						
Test	g/210L	Time																																																																																																																																																					
Air Blank	0.000	12:25																																																																																																																																																					
Control Test	0.076	12:25																																																																																																																																																					
Air Blank	0.000	12:26																																																																																																																																																					
Control Test	0.077	12:26																																																																																																																																																					
Air Blank	0.000	12:26																																																																																																																																																					
Control Test	0.077	12:27																																																																																																																																																					
Air Blank	0.000	12:27																																																																																																																																																					
Control Test Stats																																																																																																																																																							
Average	0.0767																																																																																																																																																						
Std Dev	0.0006																																																																																																																																																						
Rel Std Dev(%)	0.7531																																																																																																																																																						

Comments: Will perform an optical cal adjust. ML 04/24/2023

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %
Samples Taken = 4, Discarded = 1
3um Io = 12821, 9um Io = 12757
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.1180 (-0.0250)
Sample #2 = 0.0970 (-0.0040)
Sample #3 = 0.0880 (-0.0150)
Sample #4 = 0.0900 (-0.0340)
Avg % Abs = 0.0917 (-0.0150)
STD DEV = 0.0047 (-0.0190)
REL STD DEV = 5.155 (126.667)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.2110 (-0.0210)
Sample #2 = 0.1970 (-0.0020)
Sample #3 = 0.2000 (-0.0050)
Sample #4 = 0.1940 (-0.0170)
Avg % Abs = 0.1970 (-0.0067)
STD DEV = 0.0030 (-0.0096)
REL STD DEV = 1.523 (144.135)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %
Samples Taken = 4, Discarded = 1
3um Io = 12810, 9um Io = 12752
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.8540 (-0.0270)
Sample #2 = 0.8400 (-0.0100)
Sample #3 = 0.8410 (-0.0170)
Sample #4 = 0.8260 (-0.0320)
Avg % Abs = 0.8357 (-0.0197)
STD DEV = 0.0084 (-0.0112)
REL STD DEV = 1.004 (57.152)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.6220 (-0.0030)
Sample #2 = 1.6290 (-0.0190)
Sample #3 = 1.6150 (-0.0230)
Sample #4 = 1.6140 (-0.0340)
Avg % Abs = 1.6193 (-0.0253)
STD DEV = 0.0084 (-0.0078)
REL STD DEV = 0.518 (30.661)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 mg/l %
Samples Taken = 4, Discarded = 1
3um Io = 12788, 9um Io = 12739
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 5.3890 (-0.0020)
Sample #2 = 5.3810 (-0.0380)
Sample #3 = 5.3660 (-0.0250)
Sample #4 = 5.3700 (-0.0430)
Avg % Abs = 5.3790 (-0.0353)
STD DEV = 0.0082 (-0.0093)
REL STD DEV = 0.152 (26.297)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 10.1140 (-0.0040)
Sample #2 = 10.0870 (-0.0410)
Sample #3 = 10.0750 (-0.0290)
Sample #4 = 10.0710 (-0.0470)
Avg % Abs = 10.0777 (-0.0390)
STD DEV = 0.0083 (-0.0092)
REL STD DEV = 0.083 (23.500)

Sol Value = 0.000 mg/l or 0.000 g/210L
Fit Value = 0.0000 mg/l or 0.000 g/210L
Samples Taken = 4, Discarded = 1
3um Io = 12805, 9um Io = 12749
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.9410 (-0.0280)
Sample #2 = 1.8850 (-0.0200)
Sample #3 = 1.9190 (-0.0040)
Sample #4 = 1.8950 (-0.0300)
Avg % Abs = 1.8937 (-0.0180)
STD DEV = 0.0175 (-0.0131)
REL STD DEV = 0.920 (72.860)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 7.0230 (-0.0010)
Sample #2 = 6.9910 (-0.0100)
Sample #3 = 6.9930 (-0.0160)
Sample #4 = 6.9790 (-0.0160)
Avg % Abs = 6.9877 (-0.0140)
STD DEV = 0.0076 (-0.0035)
REL STD DEV = 0.108 (24.744)

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.000
0.040	0.040	-0.0004
0.100	0.099	0.0008
0.200	0.201	-0.0006
0.300	0.300	0.0002

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0001
0.040	0.040	-0.0003
0.100	0.099	0.0010
0.200	0.201	-0.0009
0.300	0.300	0.0003

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 mg/l %
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****

Sample #1 = 3240.00
Sample #2 = 3213.00
Sample #3 = 3179.00
Sample #4 = 3307.00
Average Result = 3233.0000
STD DEV = 66.3023
REL STD DEV = 2.051

***** CHANNEL 2 *****
Sample #1 = 3192.00
Sample #2 = 3201.00
Sample #3 = 3179.00
Sample #4 = 3208.00
Average Result = 3196.0000
STD DEV = 15.1327
REL STD DEV = 0.473

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l*10,000) = 576
9 um H2O Adjust (mg/l*10,000) = 613
***** AUTO CAL PASS

SN:	80-001321
Agency:	Okeechobee CSD
Date:	04/27/2023
Quadratic Fit:	+/- 0.002g/210L ✓
By:	TDG <i>ML</i>

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post - Cal)	80-001321	Okeechobee CSO	04/27/2023	TDG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
<p>OKEECHOBEE COUNTY SO Intoxilyzer - Alconol Analyzer Model 8000 04/27/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:29</p> <p>Control Test 0.049 13:30</p> <p>Air Blank 0.000 13:31</p> <p>Control Test 0.049 13:31</p> <p>Air Blank 0.000 13:32</p> <p>Control Test 0.049 13:32</p> <p>Air Blank 0.000 13:33</p> <p>Control Test Stats</p> <p>Average 0.0490</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>OKEECHOBEE COUNTY SO Intoxilyzer - Alconol Analyzer Model 8000 04/27/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:36</p> <p>Control Test 0.079 13:37</p> <p>Air Blank 0.000 13:37</p> <p>Control Test 0.079 13:38</p> <p>Air Blank 0.000 13:39</p> <p>Control Test 0.079 13:39</p> <p>Air Blank 0.000 13:40</p> <p>Control Test Stats</p> <p>Average 0.0790</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>OKEECHOBEE COUNTY SO Intoxilyzer - Alconol Analyzer Model 8000 04/27/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:43</p> <p>Control Test 0.199 13:44</p> <p>Air Blank 0.000 13:44</p> <p>Control Test 0.199 13:45</p> <p>Air Blank 0.000 13:45</p> <p>Control Test 0.199 13:46</p> <p>Air Blank 0.000 13:47</p> <p>Control Test Stats</p> <p>Average 0.1990</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>OKEECHOBEE COUNTY SO Intoxilyzer - Alconol Analyzer Model 8000 04/27/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:22</p> <p>Control Test 0.080 13:23</p> <p>Air Blank 0.000 13:23</p> <p>Control Test 0.080 13:24</p> <p>Air Blank 0.000 13:24</p> <p>Control Test 0.079 13:24</p> <p>Air Blank 0.000 13:25</p> <p>Control Test Stats</p> <p>Average 0.0797</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7247</p>
<p>Operator's Signature</p>	<p>Operator's Signature</p>	<p>Operator's Signature</p>	<p>Operator's Signature</p>

Comments:

Return Material Authorization

Ship to:



CMI, Inc.



Enforcement Electronics

Shipment to repair facility authorized by: Mark Garland on 4/26/2023

Items Returned: Instrument ☐ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001321

Bill To Address:

Okeechobee CSO

Attn: Mark Garland

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Internal speaker is detached from the inside casing.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Mark Garland

Phone #: 863-763-3117

Email: mgarland@okeesheriff.com

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OKEECHOBEE COUNTY SO
Time of Inspection: 13:04

Date of Inspection: 04/28/2023

Serial Number: 80-001321
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.080
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.079
0.000	0.050	0.078	0.198	0.078
0.000	0.049	0.078	0.199	0.079
Standard Deviations	0.0003	0.0000	0.0004	0.0005

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

04/28/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001321, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001321</u>	UNCERTAINTY* \pm
Owning Agency:	<u>OKEECHOBEE COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>04/28/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:04</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).


The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



04/28/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality