



INSTRUMENT PROCESSING SHEET

Agency Sarasota CSOS/N 80-001347Florida Department of
Law EnforcementDate In 09/14/2023 DI Completion Date 09/19/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date	Flow Calibration	By	Date																																								
☒ Annual ☐ Registration ☒ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>244</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		<u>09/19/2023</u>	Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024			Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____																											
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5094	202201C 01/11/2024																																													
0.080	MP5095	202201D 01/18/2024																																													
0.200	MP5096	202201E 01/18/2024																																													
0.080 DGS	N/A	AG223802 08/26/2024																																													
Calibration Adjustment By _____		Department Inspection By TDG _____																																													
Barometric Pressure Gauge ID # _____		Barometric Pressure ID# <u>26932</u> Gauge <u>1015</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP4863</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP4863	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	MP4863																																														
Interferent	MP5093																																														
0.050	MP5094																																														
0.080	MP5095																																														
0.200	MP5096																																														
☐ Post Calibration Adjustment Stability Checks		Attachments ☒ Form 41 ☐ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☐ Calibration Adjustment ☐ Other _____																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																									
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____		Benjamin Siddoway Digitally signed by Benjamin Siddoway Date: 2023.09.19 16:54:31 +0400 Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.09.21 09:01:08 -0400 Tech Review / Date _____ Admin Review / Date _____																																													

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
0.003 of Wet	✓	✓	✓
≤0.003 of Wet	✓	✓	✓
SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001347 09/19/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:05 Control Test 0.051 10:06 Air Blank 0.000 10:07 Control Test 0.048 10:07 Air Blank 0.000 10:08 Control Test 0.049 10:08 Air Blank 0.000 10:09 Control Test Stats Average 0.0493 Std Dev 0.0015 Rel Std Dev(%) 3.0953	SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001347 09/19/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:12 Control Test 0.078 10:12 Air Blank 0.000 10:13 Control Test 0.078 10:14 Air Blank 0.000 10:14 Control Test 0.078 10:15 Air Blank 0.000 10:15 Control Test Stats Average 0.0780 Std Dev 0.0000 Rel Std Dev(%) 0.0000	SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001347 09/19/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 10:16 Control Test 0.200 10:17 Air Blank 0.000 10:18 Control Test 0.200 10:18 Air Blank 0.000 10:19 Control Test 0.199 10:20 Air Blank 0.000 10:20 Control Test Stats Average 0.1997 Std Dev 0.0006 Rel Std Dev(%) 0.2892	SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-001347 09/19/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 09:55 Control Test 0.080 09:55 Air Blank 0.000 09:55 Control Test 0.080 09:56 Air Blank 0.000 09:56 Control Test 0.080 09:57 Air Blank 0.000 09:57 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000
Operator's Signature <i>WV</i>	Operator's Signature <i>WV</i>	Operator's Signature <i>WV</i>	Operator's Signature <i>WV</i>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 12:54

Date of Inspection: 09/19/2023

Serial Number: 80-001347
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.077	0.199	0.080
0.000	0.048	0.076	0.198	0.080
0.000	0.048	0.077	0.199	0.080
0.000	0.047	0.077	0.199	0.080
0.000	0.048	0.076	0.199	0.080
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.076	0.199	0.080

Standard Deviations	0.0003	0.0006	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

09/19/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001347, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001347</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>09/19/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:54</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/19/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Sarasota CSOS/N 80-001347Florida Department of
Law EnforcementDate In 04/14/2023DI Completion Date 05/18/2023☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>05/03/2023</u>	Flow Calibration	By TDG	Date <u>05/03/2023</u>																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>87*</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.113**</u> (.139 - .169) 36 mm <u>0.128**</u> (.156 - .190) 53 mm <u>0.199**</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks			Flow Column # <u>ATP106 (x2)</u> ☒ 5L/min - 17mm ☒ 15L/min - 53mm ☒ 30L/min - 103mm ☒ R-Value <u>89* / 89*</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP104 (x2)</u> 32 mm <u>0.140 / 0.144</u> (.139 - .169) 36 mm <u>0.156 / 0.156</u> (.156 - .190) 53 mm <u>0.222** / 0.230</u> (.228 - .278) 103 mm <u>0.500 / 0.515</u> (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP5094	202201C 01/11/2024																																												
0.080	MP5095	202201D 01/18/2024																																												
0.200	MP5096	202201E 01/18/2024																																												
0.080 DGS	N/A	AG223802 08/26/2024																																												
Calibration Adjustment		By _____		Department Inspection			By TDG _____																																							
Barometric Pressure Gauge ID # _____				Barometric Pressure ID# <u>28663</u> Gauge <u>1016</u> Instrument <u>1012</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP5092																																													
Interferent	MP5093																																													
0.050	MP5094																																													
0.080	MP5095																																													
0.200	MP5096																																													
☐ Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☒ Flow Calibrations</td></tr><tr><td>☐ Calibration Certificate</td><td>☐ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☒ Other <u>Form 51</u></td></tr></tbody></table>			☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☒ Flow Calibrations	☐ Calibration Certificate	☐ Form 40	☐ Calibration Adjustment	☒ Other <u>Form 51</u>												
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
☒ Form 41	☐ Post-Stability Checks																																													
☒ Stability Checks	☒ Flow Calibrations																																													
☐ Calibration Certificate	☐ Form 40																																													
☐ Calibration Adjustment	☒ Other <u>Form 51</u>																																													
Notes/Suggested Service: <u>*R-value is below 100. (TDG)</u> <u>**Flow values outside nominal range. Conducted flow calibrations. (TDG)</u> <u>Optical calibration adjustment not performed due to the instrument being sent to repair for R-value. Al Feltovic sent a box and static bag for shipment to CMI. (TDG)</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																										
				2023.05.1 Israel Soto Digitally signed by Israel Soto Date: 2023.05.19 07:30:51 -04'00' 9 07:50:45 04'00' Tech Review / Date Admin Review / Date																																										

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-001347	Sarasota CSO	05/03/2023	TDG

0.05g/210L	0.08g/210L	0.20g/210L	0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
SARASOTA COUNTY SO Intoxilyzer - Alcolon Analyzer Model 8000 05/03/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:14 Control Test 0.050 12:15 Air Blank 0.000 12:15 Control Test 0.050 12:16 Air Blank 0.000 12:16 Control Test 0.050 12:17 Air Blank 0.000 12:18 Control Test Stats Average 0.0500 Std Dev 0.0000 Rel Std Dev(%) 0.0000	SARASOTA COUNTY SO Intoxilyzer - Alcolon Analyzer Model 8000 05/03/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:20 Control Test 0.079 12:21 Air Blank 0.000 12:22 Control Test 0.079 12:23 Air Blank 0.000 12:23 Control Test 0.080 12:24 Control Test Stats Average 0.0793 Std Dev 0.0006 Rel Std Dev(%) 0.7277	SARASOTA COUNTY SO Intoxilyzer - Alcolon Analyzer Model 8000 05/03/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:30 Control Test 0.200 12:31 Air Blank 0.000 12:31 Control Test 0.200 12:32 Air Blank 0.000 12:33 Control Test 0.200 12:33 Air Blank 0.000 12:34 Control Test Stats Average 0.2000 Std Dev 0.0000 Rel Std Dev(%) 0.0000	SARASOTA COUNTY SO Intoxilyzer - Alcolon Analyzer Model 8000 05/03/2023 Software: 8100.27 Test g/210L Time Air Blank 0.000 12:05 Control Test 0.083 12:05 Air Blank 0.000 12:06 Control Test 0.083 12:06 Air Blank 0.000 12:07 Control Test 0.082 12:07 Air Blank 0.000 12:07 Control Test Stats Average 0.0827 Std Dev 0.0006 Rel Std Dev(%) 0.6984
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Comments:

Flow Calibrations

80-001347

5/3/2023

ML

#1

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001347
05/03/2023
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 4.242

2: Rate (Liters/min) = 15

SQRT(Diff) = 8.773

3: Rate (Liters/min) = 30

SQRT(Diff) = 19.207

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 637

Rounded Intercept = -84083

Correlation = 0.99407

#2

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001347
05/03/2023
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 4.355

2: Rate (Liters/min) = 15

SQRT(Diff) = 8.715

3: Rate (Liters/min) = 30

SQRT(Diff) = 18.785

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 660

Rounded Intercept = -127870

Correlation = 0.99399

Return Material Authorization

Ship to:



CMI, Inc.



Enforcement Electronics

Shipment to repair facility authorized by: Michael Feltovic on 05/04/2023

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001347

Bill To Address:

Sarasota County Sheriff's Office

Attn: Michael Feltovic

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Alcohol Testing Program

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

R-value is less than 100. Agreement between 0.08 ARS and 0.08 DGS is >0.003 g/210 L.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$ _____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Michael Feltovic

Phone #: 941-706-6943

Email: Michael.Feltovic@sarasotasheriff.org

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 14:14

Date of Inspection: 05/18/2023

Serial Number: 80-001347
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.050	0.080	0.200	0.081
0.000	0.050	0.079	0.199	0.081
0.000	0.050	0.080	0.199	0.081
0.000	0.050	0.079	0.199	0.081
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.049	0.080	0.199	0.080
0.000	0.050	0.079	0.199	0.080

Standard Deviations	0.0004	0.0005	0.0003	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

05/18/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001347, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001347</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/18/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:14</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

05/18/2023

Date

TAYLOR D GUTSCHOW,
Department Inspector

Service • Integrity • Respect • Quality