



Agency Pinellas County Sheriff's Office

S/N 80-001415

Date In 06/16/2023 DI Completion Date 06-19-2023

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By_

Barometric Pressure Gauge _____ ID # _____

☐ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

Department Inspection

By ISBarometric Pressure ID# 28421Gauge 1013 Instrument 1012Mouth Alcohol Solution Lot # 2022-A

Acetone Stock Solution Lot # 2022-B

Attachments

- Instrument Complies with Chapter 11D-8, FAC**

- ☐
- Instrument Does Not Comply with Chapter 11D-8, FAC

- #### Return to/Place into Evidentiary Use

- ☐
- Remain Out of Evidentiary Use**

- ### ■ Conduct an Agency Inspection Before Evidentiary Use

Taylor
Gutschow

Digitally signed by Taylor
Gutschow
Date: 2023.06.19 11:45:10
-04'00'

Phil Nicodemo

Digitally signed by Phil Nicodemo
Date: 2023.06.22 08:19:25 -04'00'

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 11:09

Date of Inspection: 06/19/2023

Serial Number: 80-001415
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.049	0.079	0.200	0.077
0.000	0.048	0.080	0.201	0.077
0.000	0.049	0.079	0.201	0.077
0.000	0.048	0.080	0.201	0.077
0.000	0.048	0.080	0.201	0.078
0.000	0.047	0.080	0.199	0.077
0.000	0.047	0.081	0.199	0.077
0.000	0.046	0.080	0.199	0.077
0.000	0.046	0.080	0.198	0.077
0.000	0.047	0.081	0.199	0.077

Standard Deviations	0.0010	0.0006	0.0011	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

06/19/2023
Date

Stability checks

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:06
Control Test	0.048	09:07
Air Blank	0.000	09:07
Control Test	0.049	09:08
Air Blank	0.000	09:08
Control Test	0.048	09:09
Air Blank	0.000	09:10
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:11
Control Test	0.079	09:12
Air Blank	0.000	09:13
Control Test	0.079	09:13
Air Blank	0.000	09:14
Control Test	0.079	09:15
Air Blank	0.000	09:15
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:16
Control Test	0.201	09:17
Air Blank	0.000	09:18
Control Test	0.200	09:18
Air Blank	0.000	09:19
Control Test	0.200	09:20
Air Blank	0.000	09:20
Control Test Stats		
Average	0.2003	
Std Dev	0.0006	
Rel Std Dev(%)	0.2882	

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
06/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.077	09:32
Air Blank	0.000	09:33
Control Test	0.077	09:33
Air Blank	0.000	09:34
Control Test	0.077	09:34
Air Blank	0.000	09:35
Control Test Stats		
Average	0.0770	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

Dry



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001415, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001415</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>PINELLAS COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/19/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:09</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Date: 2023.06.19 11:18:08
-04'00'

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Pinellas County Sheriff's OfficeS/N 80-001415Florida Department of
Law EnforcementDate In 02-09-2023 DI Completion Date N/A☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By IS	Quality Checks	By IS	Date <u>02-09-2023</u>	Flow Calibration	By	Date																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>214</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D 01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 10-25-2024</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG229803 10-25-2024	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By IS</th></tr></thead><tbody><tr><td colspan="2">☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____</td><td>_____ _____ _____ _____ _____</td></tr></tbody></table>			Maintenance		By IS	☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____		_____ _____ _____ _____ _____																			
Simulator	Serial #	Lot #/Exp																																													
0.050	MP6291	202201C 01-11-2024																																													
0.080	MP6292	202201D 01-18-2024																																													
0.200	MP6293	202201E 01-18-2024																																													
0.080 DGS	N/A	AG229803 10-25-2024																																													
Maintenance		By IS																																													
☒ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____		_____ _____ _____ _____ _____																																													
Calibration Adjustment		By PN		Department Inspection			By IS																																								
Barometric Pressure Gauge <u>1014</u> ID # <u>28427</u>				Barometric Pressure ID# <u>30793</u> Gauge <u>1015</u> Instrument <u>1014</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>21410</td><td>9/30/2023</td></tr><tr><td>0.100</td><td>MP6296</td><td>22310</td><td>8/11/2024</td></tr><tr><td>0.200</td><td>MP6297</td><td>22050</td><td>2/7/2024</td></tr><tr><td>0.300</td><td>MP6298</td><td>22220</td><td>6/15/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>08121080A1</td><td>5/05/2023</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	21410	9/30/2023	0.100	MP6296	22310	8/11/2024	0.200	MP6297	22050	2/7/2024	0.300	MP6298	22220	6/15/2024	0.080 DGS	N/A	08121080A1	5/05/2023			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090	
Simulator	Serial #	Lot #	Expiration																																												
0.000	MP6294	N/A	N/A																																												
0.040	MP6295	21410	9/30/2023																																												
0.100	MP6296	22310	8/11/2024																																												
0.200	MP6297	22050	2/7/2024																																												
0.300	MP6298	22220	6/15/2024																																												
0.080 DGS	N/A	08121080A1	5/05/2023																																												
Simulator	Serial Number																																														
0.000	MP5086																																														
Interferent	MP5087																																														
0.050	MP5088																																														
0.080	MP5089																																														
0.200	MP5090																																														
☒ Post Calibration Adjustment Stability Checks				Attachments																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C</td><td>01/11/2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D</td><td>01/18/2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E</td><td>01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803</td><td>10/25/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202201C	01/11/2024	0.080	MP6292	202201D	01/18/2024	0.200	MP6293	202201E	01/18/2024	0.080 DGS	N/A	AG229803	10/25/2024			<table border="1"><tbody><tr><td>☒ Form 41 ☒ Stability Checks ☐ Calibration Certificate ☒ Calibration Adjustment</td><td>☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Form 51</u></td></tr></tbody></table>			☒ Form 41 ☒ Stability Checks ☐ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Form 51</u>																			
Simulator	Serial #	Lot #	Expiration																																												
0.050	MP6291	202201C	01/11/2024																																												
0.080	MP6292	202201D	01/18/2024																																												
0.200	MP6293	202201E	01/18/2024																																												
0.080 DGS	N/A	AG229803	10/25/2024																																												
☒ Form 41 ☒ Stability Checks ☐ Calibration Certificate ☒ Calibration Adjustment	☒ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☒ Other <u>Form 51</u>																																														
Notes/Suggested Service: <u>Battery replaced prior to Quality Checks and Department Inspection.</u> <u>Instrument had Ambient Fail and Interferent Detect during Stability Checks. Optical Bench Calibration Adjustment was performed. Instrument continued to have Ambient Fail and Interferent Detect during Department Inspection. IS</u> _____ _____ _____				☐ Instrument Complies with Chapter 11D-8, FAC ☒ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																											
				Tech Review / Date _____			Admin Review / Date _____																																								

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 12:56

Date of Inspection: 02/10/2023

Serial Number: 80-001415
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:
0.000	0.048			
0.000	0.048			
0.000	0.051			
0.000	0.052			
0.000	0.052			
0.000	0.051			
0.000	0.052			
0.000	0.053			
0.000	0.053			
0.000	INT			

Standard Deviations	0.0028			
---------------------	--------	--	--	--


Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

Int Det: Ambient Fail CLEARED AREA. 05: Interferent Detect. Non-compliance:WILL SEND TO REPAIR.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



ISRAEL SOTO

Signature and Printed Name

02/10/2023
Date

stability checks

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
02/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:10
Control Test	0.046	16:10
Air Blank	0.000	16:11
Control Test	0.050	16:11
Air Blank	0.000	16:12
Control Test	INT*	16:13
Air Blank	PUR**	16:13
Air Blank	PUR**	16:14

*Interferent Detect
**Purge Fail

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
02/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:19
Control Test	0.078	16:20
Air Blank	AMB*	16:21
Air Blank	PUR**	16:21
*Ambient Fail		
**Purge Fail		

wet

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
02/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:23
Control Test	0.201	16:24
Air Blank	PUR*	16:25
Air Blank	PUR*	16:25

*Purge Fail

Operator's Signature

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
02/09/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	16:29
Control Test	0.079	16:29
Air Blank	0.000	16:30
Control Test	0.080	16:30
Air Blank	0.000	16:31
Control Test	0.078	16:31
Air Blank	0.000	16:32
Control Test Stats		
Average	0.0790	
Std Dev	0.0010	
Rel Std Dev(%)	1.2658	

Dry

Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
02/10/2023 08:52:48

optical bench calibration
Adjustment


Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 0.0880 (0.0650) 0.0860 (0.0430)
Sample #2 0.0670 (0.1160) 0.1010 (0.0480)
Sample #3 0.0680 (0.1230) 0.0930 (0.0490)
Sample #4 0.0720 (0.1290) 0.0780 (0.0460)
Avg % Abs 0.0690 (0.1227) 0.0907 (0.0477)
STD DEV 0.0026 (0.0065) 0.0117 (0.0015)
REL STD DEV 3.834 (5.304) 12.878 (3.205)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 0.7950 (-0.0260) 1.5150 (0.0060)
Sample #2 0.8050 (0.1690) 1.5300 (0.0940)
Sample #3 0.8040 (0.1730) 1.5390 (0.0880)
Sample #4 0.7780 (0.2630) 1.5140 (0.1350)
Avg % Abs 0.7957 (0.2017) 1.5277 (0.1057)
STD DEV 0.0153 (0.0532) 0.0127 (0.0256)
REL STD DEV 1.924 (26.357) 0.829 (24.208)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 1.8790 (-0.0310) 3.6250 (-0.0060)
Sample #2 1.8870 (-0.0310) 3.6440 (-0.0170)
Sample #3 1.8730 (-0.0170) 3.6130 (0.0010)
Sample #4 1.8840 (0.0040) 3.6420 (0.0000)
Avg % Abs 1.8813 (-0.0147) 3.6330 (-0.0053)
STD DEV 0.0074 (0.0176) 0.0173 (0.0101)
REL STD DEV 0.392 (120.111) 0.478 (189.675)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 3.6070 (-0.0120) 6.9680 (-0.0080)
Sample #2 3.6370 (0.0270) 6.9580 (0.0240)
Sample #3 3.5800 (0.0630) 6.9590 (0.0290)
Sample #4 3.5910 (0.0730) 6.9680 (0.0290)
Avg % Abs 3.6027 (0.0543) 6.9617 (0.0273)
STD DEV 0.0302 (0.0242) 0.0055 (0.0029)
REL STD DEV 0.839 (44.528) 0.079 (10.561)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 5.2220 (-0.0240) 10.0600 (-0.0040)
Sample #2 5.2660 (-0.0840) 10.0860 (-0.0230)
Sample #3 5.2450 (-0.0610) 10.0720 (-0.0120)
Sample #4 5.2340 (-0.0600) 10.0880 (-0.0250)
Avg % Abs 5.2483 (-0.0683) 10.0820 (-0.0200)
STD DEV 0.0163 (0.0136) 0.0087 (0.0070)
REL STD DEV 0.310 (19.869) 0.086 (35.000)

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
02/10/2023 08:52:48

Optical Bench Calibration
Adjustment



Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>

Zero Order Coef -167.04
First Order Coef 2551.75
Second Order Coef 38.44

<<<<< 9um >>>>>

Zero Order Coef -109.00
First Order Coef 1290.57
Second Order Coef 13.56

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0002
0.040	0.040	0.0004
0.100	0.100	-0.0002
0.200	0.200	-0.0000
0.300	0.300	0.0000

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0002
0.040	0.040	0.0002
0.100	0.100	0.0001
0.200	0.200	-0.0002
0.300	0.300	0.0001

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 2954.00 3297.00
Sample #2 2858.00 3293.00
Sample #3 2914.00 3335.00
Sample #4 2930.00 3323.00
Avg 2900.6667 3317.0000
STD DEV 37.8065 21.6333
REL STD DEV 1.303 0.652
H2O adjust (mg/l*10k) 909 492

Barometric Pressure = 1015

*****CALIBRATION SUCCESSFUL*****

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
02/10/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:14
Control Test	0.049	10:14
Air Blank	0.000	10:15
Control Test	0.049	10:15
Air Blank	0.000	10:16
Control Test	0.049	10:17
Air Blank	0.000	10:17
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
02/10/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.078	10:19
Air Blank	0.000	10:20
Control Test	0.078	10:20
Air Blank	0.000	10:21
Control Test	0.082	10:22
Air Blank	0.000	10:22
Control Test Stats		
Average	0.0793	
Std Dev	0.0023	
Rel Std Dev(%)	2.9110	



Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
02/10/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:24
Control Test	0.198	10:24
Air Blank	0.000	10:25
Control Test	0.199	10:25
Air Blank	0.000	10:26
Control Test	0.199	10:27
Air Blank	0.000	10:27
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	



Operator's Signature

PINELLAS COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001415
02/10/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:28
Control Test	0.080	10:29
Air Blank	0.000	10:29
Control Test	0.079	10:29
Air Blank	0.000	10:30
Control Test	0.080	10:30
Air Blank	0.000	10:31
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

D65



Operator's Signature

POST

STABILITY CHECKS

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PINELLAS COUNTY SO
Time of Inspection: 15:30

Date of Inspection: 02/09/2023

Serial Number: 80-001415
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

N/A compliance not determined

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

02/09/2023
Date

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Malcolm Deane on 02-15-2023

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001415

Bill To Address:

Pinellas County Sheriff's Office

Florida

Ship to Address:

Florida Department of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Instrument is producing Ambient Fail and Purge Fail intermittently, and produces

Interferent Detect on alcohol reference solutions.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Malcolm Dean

Phone #: 727-464-6330

Email: mdeane@pcsonet.com

ATP Contact Name: Israel Soto

ATP Email: israelsoto@fdle.state.fl.us