



INSTRUMENT PROCESSING SHEET

Agency Orange County Sheriff's OfficeS/N 80-001420Florida Department of
Law EnforcementDate In 1/27/2023 DI Completion Date 2/1/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By PN	Quality Checks	By PN	Date <u>2/1/2023</u>	Flow Calibration	By	Date																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>210</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>.164</u> (.139 - .169) 36 mm <u>.183</u> (.156 - .190) 53 mm <u>.246</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28421</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5090</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 10/25/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202201C 01/11/2024	0.080	MP5089	202201D 01/18/2024	0.200	MP5090	202201E 01/18/2024	0.080 DGS	N/A	AG229803 10/25/2024			Maintenance By _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____ _____ _____ _____ _____ _____																											
Simulator	Serial #	Lot #/Exp																																													
0.050	MP5088	202201C 01/11/2024																																													
0.080	MP5089	202201D 01/18/2024																																													
0.200	MP5090	202201E 01/18/2024																																													
0.080 DGS	N/A	AG229803 10/25/2024																																													
Calibration Adjustment By PN _____		Department Inspection By PN _____																																													
Barometric Pressure Gauge <u>1021</u> ID # <u>30793</u>		Barometric Pressure ID# <u>28421</u>		Gauge <u>1019</u> Instrument <u>1018</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>21410</td><td>9/30/2023</td></tr><tr><td>0.100</td><td>MP6296</td><td>22310</td><td>8/11/2024</td></tr><tr><td>0.200</td><td>MP6297</td><td>22050</td><td>2/7/2024</td></tr><tr><td>0.300</td><td>MP6298</td><td>22220</td><td>6/15/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>08121080A1</td><td>5/5/2023</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	21410	9/30/2023	0.100	MP6296	22310	8/11/2024	0.200	MP6297	22050	2/7/2024	0.300	MP6298	22220	6/15/2024	0.080 DGS	N/A	08121080A1	5/5/2023	Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>		<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																												
0.000	MP6294	N/A	N/A																																												
0.040	MP6295	21410	9/30/2023																																												
0.100	MP6296	22310	8/11/2024																																												
0.200	MP6297	22050	2/7/2024																																												
0.300	MP6298	22220	6/15/2024																																												
0.080 DGS	N/A	08121080A1	5/5/2023																																												
Simulator	Serial Number																																														
0.000	MP5086																																														
Interferent	MP5087																																														
0.050	MP5088																																														
0.080	MP5089																																														
0.200	MP5090																																														
☒ Post Calibration Adjustment Stability Checks		Attachments																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202201C</td><td>01/11/2024</td></tr><tr><td>0.080</td><td>MP5089</td><td>202201D</td><td>01/18/2024</td></tr><tr><td>0.200</td><td>MP5090</td><td>202201E</td><td>01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803</td><td>10/25/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202201C	01/11/2024	0.080	MP5089	202201D	01/18/2024	0.200	MP5090	202201E	01/18/2024	0.080 DGS	N/A	AG229803	10/25/2024	☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment		☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____																							
Simulator	Serial #	Lot #	Expiration																																												
0.050	MP5088	202201C	01/11/2024																																												
0.080	MP5089	202201D	01/18/2024																																												
0.200	MP5090	202201E	01/18/2024																																												
0.080 DGS	N/A	AG229803	10/25/2024																																												
Notes/Suggested Service: _____ During 0.08 DGS Stability Check, inadvertently only two measurements were performed; since a calibration adjustment was required based on the measurements obtained during the 0.08 wet bath Stability Check, the 0.08 DGS check was not repeated to obtain three measurements. PN _____ _____ _____ _____		☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																													
		Tech Review / Date _____		Admin Review / Date _____																																											

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: ORANGE COUNTY S.O.
Time of Inspection: 15:37

Date of Inspection: 02/01/2023

Serial Number: 80-001420
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.048	0.078	0.197	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.080
0.000	0.048	0.078	0.197	0.080
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.079	0.198	0.079


Standard Deviations	0.0004	0.0003	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


 PHIL NICODEMO
 Signature and Printed Name

02/01/2023
 Date

STABILITY CHECKS

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
02/01/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:41
Control Test	0.047	09:42
Air Blank	0.000	09:42
Control Test	0.047	09:43
Air Blank	0.000	09:43
Control Test	0.047	09:44
Air Blank	0.000	09:44
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
02/01/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:46
Control Test	0.076	09:46
Air Blank	0.000	09:47
Control Test	0.076	09:47
Air Blank	0.000	09:48
Control Test	0.076	09:49
Air Blank	0.000	09:49
Control Test Stats		
Average	0.0760	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
02/01/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:52
Control Test	0.196	09:53
Air Blank	0.000	09:53
Control Test	0.197	09:54
Air Blank	0.000	09:54
Control Test	0.196	09:55
Air Blank	0.000	09:56
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	



Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
02/01/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:57
Control Test	0.079	09:57
Air Blank	0.000	09:58
Control Test	0.081	09:58
Air Blank	0.000	09:59
Control Test Stats		
Average	0.0800	
Std Dev	0.0014	
Rel Std Dev(%)	1.7678	

DGS* See notes



Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001420, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001420</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>ORANGE COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/01/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:37</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/01/2023

Date


PHIL NICODEMO,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
02/01/2023 10:28:03

Optical Calibration Adjustment



Auto Calibration

pg 1 of 2

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0740	(-0.0050)	0.0950	(-0.0030)
Sample #2	0.0840	(0.0410)	0.0780	(0.0080)
Sample #3	0.0650	(0.0520)	0.1020	(0.0060)
Sample #4	0.0390	(0.0890)	0.0810	(0.0190)
Avg % Abs	0.0627	(0.0607)	0.0870	(0.0110)
STD DEV	0.0226	(0.0251)	0.0131	(0.0070)
REL STD DEV	36.049	(41.450)	15.031	(63.636)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8460	(-0.0200)	1.5500	(-0.0130)
Sample #2	0.8400	(-0.0060)	1.5480	(-0.0030)
Sample #3	0.8200	(0.0300)	1.5460	(-0.0020)
Sample #4	0.8290	(0.0450)	1.5560	(0.0120)
Avg % Abs	0.8297	(0.0230)	1.5500	(0.0023)
STD DEV	0.0100	(0.0262)	0.0053	(0.0084)
REL STD DEV	1.207	(113.959)	0.341	(359.421)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	2.0270	(-0.0290)	3.6700	(0.0000)
Sample #2	2.0150	(-0.0090)	3.6910	(-0.0080)
Sample #3	1.9670	(0.0070)	3.6890	(0.0060)
Sample #4	2.0120	(-0.0040)	3.6980	(-0.0050)
Avg % Abs	1.9980	(-0.0020)	3.6927	(-0.0023)
STD DEV	0.0269	(0.0082)	0.0047	(0.0074)
REL STD DEV	1.346	(409.268)	0.128	(315.905)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	3.8500	(-0.0140)	7.0730	(-0.0150)
Sample #2	3.8350	(-0.0070)	7.0480	(0.0050)
Sample #3	3.8240	(-0.0040)	7.0760	(-0.0150)
Sample #4	3.7980	(0.0270)	7.0380	(0.0110)
Avg % Abs	3.8190	(0.0053)	7.0540	(0.0003)
STD DEV	0.0190	(0.0188)	0.0197	(0.0136)
REL STD DEV	0.498	(352.945)	0.279	(4084.115)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1

Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.6090	(-0.0130)	10.2380	(-0.0260)
Sample #2	5.6230	(-0.0030)	10.2110	(0.0140)
Sample #3	5.6200	(-0.0080)	10.2000	(0.0090)
Sample #4	5.5810	(0.0310)	10.2110	(0.0210)
Avg % Abs	5.6080	(0.0067)	10.2073	(0.0147)
STD DEV	0.0234	(0.0212)	0.0064	(0.0060)
REL STD DEV	0.418	(318.316)	0.062	(41.098)

Optical Calibration Adjustment

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
02/01/2023 10:28:03

(Handwritten signature)

Auto Calibration

pg 2 of 2

```
<<<<<      3um      >>>>>
-----
Zero Order Coef      -146.74
First Order Coef     2420.33
Second Order Coef    27.56
```

```
<<<<<      9um      >>>>>
-----
Zero Order Coef      -103.39
First Order Coef     1266.52
Second Order Coef     14.00
```

```
-----
Act      Fit      Residual
(g/210L) (g/210L) (g/210L)
0.000    0.000    -0.0001
0.040    0.039    0.0005
0.100    0.101    -0.0008
0.200    0.199    0.0005
0.300    0.300    -0.0002
-----
```

```
-----
Act      Fit      Residual
(g/210L) (g/210L) (g/210L)
0.000    0.000    -0.0001
0.040    0.040    0.0002
0.100    0.100    -0.0001
0.200    0.200    -0.0001
0.300    0.300    0.0001
-----
```

```
<<<<<      3um      >>>>>
-----
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1      3332.00      3377.00
Sample #2      3289.00      3395.00
Sample #3      3262.00      3368.00
Sample #4      3356.00      3383.00
Avg            3302.3333      3382.0000
STD DEV        48.3977        13.5277
REL STD DEV    1.466          0.400
H2O adjust (mg/l*10k) 507          427
```

Barometric Pressure = 1021

*****CALIBRATION SUCCESSFUL*****

Post Stability CHECKS

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
02/01/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:07
Control Test	0.048 •	13:07
Air Blank	0.000	13:08
Control Test	0.048 •	13:09
Air Blank	0.000	13:09
Control Test	0.048 •	13:10
Air Blank	0.000	13:10
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
02/01/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:16
Control Test	0.077 •	13:16
Air Blank	0.000	13:17
Control Test	0.077 •	13:17
Air Blank	0.000	13:18
Control Test	0.077 •	13:19
Air Blank	0.000	13:19
Control Test Stats		
Average	0.0770 •	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
02/01/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:20
Control Test	0.197 •	13:21
Air Blank	0.000	13:22
Control Test	0.198 •	13:22
Air Blank	0.000	13:23
Control Test	0.196 •	13:24
Air Blank	0.000	13:24
Control Test Stats		
Average	0.1970	
Std Dev	0.0010	
Rel Std Dev(%)	0.5076	

Operator's Signature

ORANGE COUNTY S.O.
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001420
02/01/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:26
Control Test	0.079 •	13:26
Air Blank	0.000	13:27
Control Test	0.080 •	13:27
Air Blank	0.000	13:27
Control Test	0.080 •	13:28
Air Blank	0.000	13:28
Control Test Stats		
Average	0.0797 •	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Operator's Signature

D65