



# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST LUCIE PD  
Time of Inspection: 10:33

Date of Inspection: 12/20/2023

Serial Number: 80-001962  
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: \_\_\_\_\_

**Remarks:**

AI NOT CONDUCTED. BYPASSED TO BRING OUT OF DISABLED MODE.

*Not determined* <sup>MG</sup> 12/20/2023

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

*Taylor D Gutschow*

TAYLOR D GUTSCHOW

Signature and Printed Name

12/20/2023  
Date



# Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
<p>PORT ST LUCIE PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-001962 12/20/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:41</td></tr> <tr><td>Control Test</td><td>0.052</td><td>10:41</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:42</td></tr> <tr><td>Control Test</td><td>0.052</td><td>10:43</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:43</td></tr> <tr><td>Control Test</td><td>0.052</td><td>10:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:44</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0520</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table>	Air Blank	0.000	10:41	Control Test	0.052	10:41	Air Blank	0.000	10:42	Control Test	0.052	10:43	Air Blank	0.000	10:43	Control Test	0.052	10:44	Air Blank	0.000	10:44	Control Test Stats			Average	0.0520		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>PORT ST LUCIE PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-001962 12/20/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:49</td></tr> <tr><td>Control Test</td><td>0.082</td><td>10:49</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:50</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:50</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:51</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:52</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:52</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0813</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7099</td><td></td></tr> </table>	Air Blank	0.000	10:49	Control Test	0.082	10:49	Air Blank	0.000	10:50	Control Test	0.081	10:50	Air Blank	0.000	10:51	Control Test	0.081	10:52	Air Blank	0.000	10:52	Control Test Stats			Average	0.0813		Std Dev	0.0006		Rel Std Dev(%)	0.7099		<p>PORT ST LUCIE PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-001962 12/20/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:55</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:56</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:56</td></tr> <tr><td>Control Test</td><td>0.200</td><td>10:57</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:58</td></tr> <tr><td>Control Test</td><td>0.199</td><td>10:58</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:59</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1997</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2892</td><td></td></tr> </table>	Air Blank	0.000	10:55	Control Test	0.200	10:56	Air Blank	0.000	10:56	Control Test	0.200	10:57	Air Blank	0.000	10:58	Control Test	0.199	10:58	Air Blank	0.000	10:59	Control Test Stats			Average	0.1997		Std Dev	0.0006		Rel Std Dev(%)	0.2892		<p>PORT ST LUCIE PD Intoxilyzer - Alcolon Analyzer Model 8000 SN 80-001962 12/20/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test</td><td>0.081</td><td>11:00</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:00</td></tr> <tr><td>Control Test</td><td>0.082</td><td>11:01</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:01</td></tr> <tr><td>Control Test</td><td>0.082</td><td>11:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:02</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0817</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7070</td><td></td></tr> </table>	Air Blank	0.000	11:00	Control Test	0.081	11:00	Air Blank	0.000	11:00	Control Test	0.082	11:01	Air Blank	0.000	11:01	Control Test	0.082	11:02	Air Blank	0.000	11:02	Control Test Stats			Average	0.0817		Std Dev	0.0006		Rel Std Dev(%)	0.7070	
Air Blank	0.000	10:41																																																																																																																																					
Control Test	0.052	10:41																																																																																																																																					
Air Blank	0.000	10:42																																																																																																																																					
Control Test	0.052	10:43																																																																																																																																					
Air Blank	0.000	10:43																																																																																																																																					
Control Test	0.052	10:44																																																																																																																																					
Air Blank	0.000	10:44																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0520																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	10:49																																																																																																																																					
Control Test	0.082	10:49																																																																																																																																					
Air Blank	0.000	10:50																																																																																																																																					
Control Test	0.081	10:50																																																																																																																																					
Air Blank	0.000	10:51																																																																																																																																					
Control Test	0.081	10:52																																																																																																																																					
Air Blank	0.000	10:52																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0813																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7099																																																																																																																																						
Air Blank	0.000	10:55																																																																																																																																					
Control Test	0.200	10:56																																																																																																																																					
Air Blank	0.000	10:56																																																																																																																																					
Control Test	0.200	10:57																																																																																																																																					
Air Blank	0.000	10:58																																																																																																																																					
Control Test	0.199	10:58																																																																																																																																					
Air Blank	0.000	10:59																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1997																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2892																																																																																																																																						
Air Blank	0.000	11:00																																																																																																																																					
Control Test	0.081	11:00																																																																																																																																					
Air Blank	0.000	11:00																																																																																																																																					
Control Test	0.082	11:01																																																																																																																																					
Air Blank	0.000	11:01																																																																																																																																					
Control Test	0.082	11:02																																																																																																																																					
Air Blank	0.000	11:02																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0817																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7070																																																																																																																																						
<p>Operator's Signature</p> <p><i>ML</i></p>	<p>Operator's Signature</p> <p><i>ML</i></p>	<p>Operator's Signature</p> <p><i>ML</i></p>	<p>Operator's Signature</p> <p><i>ML</i></p>																																																																																																																																				

# Florida Department of Law Enforcement

## Alcohol Testing Program

### DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ST LUCIE PD  
Time of Inspection: 12:36

Date of Inspection: 12/20/2023

Serial Number: 80-001962  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test* (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.052	0.083	0.199	0.082
0.000	0.053	0.082	0.200	0.082
0.000	0.053	0.083	0.199	0.081
0.000	0.053	0.082	0.200	0.083
0.000	0.053	0.083	0.200	0.082
0.000	0.053	0.083	0.201	0.081
0.000	0.053	0.083	0.201	0.083
0.000	0.054	0.083	0.200	0.082
0.000	0.053	0.083	0.201	0.083
0.000	0.054	0.083	0.201	0.081

Standard Deviations	0.0005	0.0004	0.0007	0.0008
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0006 Number of Simulators Used: 5

Remarks:

Repeated the Minimum Sample Volume Check. A sip of sparkling water from a couple minutes prior caused a faint reading. The repeat had a result of 0.000.  
TNG 12/20/2023

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.




TAYLOR D GUTSCHOW

Signature and Printed Name

12/20/2023  
Date





# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001962, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001962</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PORT ST LUCIE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/20/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:36</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence. ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021  
Issuing Authority: Alcohol Testing Program

Taylor D Gutschow  
Date 12/20/2023  
TAYLOR D GUTSCHOW,  
Department Inspector

Service • Integrity • Respect • Quality