

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 3408	FHP Troop E	02/16/2023	TDG ML

0.05g/210L		0.08g/210L		0.20g/210L		DGS 0.08g/210L					
0.047 to 0.053		0.077 to 0.083		0.194 to 0.206		0.077 to 0.083 ≤0.003 of Wet					
FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003408 02/16/2023 Software: 8100.27		FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003408 02/16/2023 Software: 8100.27		FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003408 02/16/2023 Software: 8100.27		FHP TROOP E MIAMI Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-003408 02/16/2023 Software: 8100.27					
Test	g/210L	Time	Test	g/210L	Time	Test	g/210L	Time			
Air Blank	0.000	12:17	Air Blank	0.000	12:10	Air Blank	0.000	12:24	Air Blank	0.000	12:28
Control Test	0.050	12:18	Control Test	0.079	12:11	Control Test	0.201	12:24	Control Test	0.080	12:28
Air Blank	0.000	12:18	Air Blank	0.000	12:11	Air Blank	0.000	12:25	Air Blank	0.000	12:29
Control Test	0.049	12:19	Control Test	0.079	12:12	Control Test	0.202	12:26	Control Test	0.078	12:29
Air Blank	0.000	12:19	Air Blank	0.000	12:12	Air Blank	0.000	12:26	Air Blank	0.000	12:30
Control Test	0.049	12:20	Control Test	0.079	12:13	Control Test	0.201	12:27	Control Test	0.079	12:30
Air Blank	0.000	12:20	Air Blank	0.000	12:14	Air Blank	0.000	12:27	Air Blank	0.000	12:31
Control Test Stats			Control Test Stats			Control Test Stats			Control Test Stats		
Average	0.0493		Average	0.0790		Average	0.2013		Average	0.0790	
Std Dev	0.0006		Std Dev	0.0000		Std Dev	0.0006		Std Dev	0.0010	
Rel Std Dev(%)	1.1703		Rel Std Dev(%)	0.0000		Rel Std Dev(%)	0.2868		Rel Std Dev(%)	1.2658	
Operator's Signature			Operator's Signature			Operator's Signature			Operator's Signature		

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP TROOP E MIAMI
Time of Inspection: 16:57

Date of Inspection: 02/17/2023

Serial Number: 80-003408
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.079	0.199	0.078
0.000	0.050	0.078	0.199	0.079
0.000	0.050	0.078	0.199	0.079
0.000	0.050	0.079	0.198	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.199	0.079
Standard Deviations	0.0004	0.0004	0.0003	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/17/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003408, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003408</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP TROOP E MIAMI</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/17/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>16:57</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/17/2023

Date


TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1