



INSTRUMENT PROCESSING SHEET

Agency Fort Walton Beach PDS/N 80-004770Florida Department of
Law EnforcementDate In 12/22/2023 DI Completion Date 12/22/2023☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By BS	Quality Checks	By BS	Date <u>12/22/2023</u>	Flow Calibration	By	Date																															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Drop-off</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>186</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.167</u> (.139 - .169) 36 mm <u>0.187</u> (.156 - .190) 53 mm <u>0.261</u> (.228 - .278) 103 mm <u>0.535</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																	
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K 3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 4/5/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG325603 9/13/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 3/29/2025	0.080	MP5089	202303L 3/29/2025	0.200	MP5090	202304C 4/5/2025	0.080 DGS	N/A	AG325603 9/13/2025	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td><td></td></tr><tr><td>☐ Other _____</td><td></td><td></td></tr></tbody></table>			Maintenance		By	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other _____			
Simulator	Serial #	Lot #/Exp																																				
0.050	MP5088	202303K 3/29/2025																																				
0.080	MP5089	202303L 3/29/2025																																				
0.200	MP5090	202304C 4/5/2025																																				
0.080 DGS	N/A	AG325603 9/13/2025																																				
Maintenance		By																																				
☐ Battery Replacement																																						
☐ Dry Gas Regulator Replacement																																						
☐ Breath Tube Replacement																																						
☐ Other _____																																						
Calibration Adjustment		By BS			Department Inspection			By BS																														
Barometric Pressure Gauge <u>1025</u> ID # <u>30793</u>					Barometric Pressure ID# <u>28427</u>																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>22460</td><td>12/28/2024</td></tr><tr><td>0.100</td><td>MP6296</td><td>23390</td><td>10/17/2025</td></tr><tr><td>0.200</td><td>MP6297</td><td>23340</td><td>9/18/2025</td></tr><tr><td>0.300</td><td>MP6298</td><td>23070</td><td>3/6/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>4/5/2025</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	22460	12/28/2024	0.100	MP6296	23390	10/17/2025	0.200	MP6297	23340	9/18/2025	0.300	MP6298	23070	3/6/2025	0.080 DGS	N/A	06723080A5	4/5/2025				Gauge <u>1023</u> Instrument <u>1024</u>					
Simulator	Serial #	Lot #	Expiration																																			
0.000	MP6294	N/A	N/A																																			
0.040	MP6295	22460	12/28/2024																																			
0.100	MP6296	23390	10/17/2025																																			
0.200	MP6297	23340	9/18/2025																																			
0.300	MP6298	23070	3/6/2025																																			
0.080 DGS	N/A	06723080A5	4/5/2025																																			
					Mouth Alcohol Solution Lot # <u>2022-A</u>																																	
					Acetone Stock Solution Lot # <u>2023-B</u>																																	
					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090																			
Simulator	Serial Number																																					
0.000	MP5086																																					
Interferent	MP5087																																					
0.050	MP5088																																					
0.080	MP5089																																					
0.200	MP5090																																					
☒ Post Calibration Adjustment Stability Checks					Attachments																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K</td><td>3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L</td><td>3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C</td><td>4/5/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG310901</td><td>4/19/2025</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202303K	3/29/2025	0.080	MP5089	202303L	3/29/2025	0.200	MP5090	202304C	4/5/2025	0.080 DGS	N/A	AG310901	4/19/2025				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment			☒ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____										
Simulator	Serial #	Lot #	Expiration																																			
0.050	MP5088	202303K	3/29/2025																																			
0.080	MP5089	202303L	3/29/2025																																			
0.200	MP5090	202304C	4/5/2025																																			
0.080 DGS	N/A	AG310901	4/19/2025																																			
Notes/Suggested Service: <u>Quality check barometric pressure read 986, gauge read 1023; 3.6% difference. 0.080 ARS and 0.080 DGS were not within 0.003. Calibration adjustment required. (BS 12/22/23)</u>					☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																	
					Taylor Gutschow Gutschow Date: 2023.12.22 15:08:06 -0500			Shayla Platt Platt Date: 2023.12.22 15:19:20 -0500																														
					Tech Review / Date			Admin Review / Date																														

Stability Checks

FORT WALTON BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004770
12/22/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:07
Control Test	0.049	08:08
Air Blank	0.000	08:09
Control Test	0.049	08:09
Air Blank	0.000	08:10
Control Test	0.049	08:11
Air Blank	0.000	08:11
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Siddle
Operator's Signature

FORT WALTON BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004770
12/22/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:14
Control Test	0.079	08:15
Air Blank	0.000	08:15
Control Test	0.079	08:16
Air Blank	0.000	08:17
Control Test	0.079	08:17
Air Blank	0.000	08:18
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Siddle
Operator's Signature

FORT WALTON BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004770
12/22/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:19
Control Test	0.198	08:20
Air Blank	0.000	08:20
Control Test	0.197	08:21
Air Blank	0.000	08:21
Control Test	0.196	08:22
Air Blank	0.000	08:23
Control Test Stats		
Average	0.1970	
Std Dev	0.0010	
Rel Std Dev(%)	0.5076	

Benjamin Siddle
Operator's Signature

FORT WALTON BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004770
12/22/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:29
Control Test	0.083	08:29
Air Blank	0.000	08:30
Control Test	0.083	08:30
Air Blank	0.000	08:31
Control Test	0.083	08:31
Air Blank	0.000	08:31
Control Test Stats		
Average	0.0830	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Benjamin Siddle
Operator's Signature

Optical Bench Calibration Adjustment

FORT WALTON BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000
12/22/2023
08:48:56

Auto Calibration
Max Power Res Value = 41
Auto Range Res Value = 24

Sol Value = 0.000 g/210L ***
Fit Value = 0.000 ng/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12796, Sum Io = 13015

CHANN 1

Sample	% Abs	(% Abs Ref)
Sample #1	0.0930	(0.0000)
Sample #2	0.0890	(0.0200)
Sample #3	0.1800	(0.0450)
Sample #4	0.0790	(0.0610)
Avg % Abs	0.0827	(0.0420)
STD DEV	0.0055	(0.0207)
REL STD DEV	0.662	(49.200)

CHANN 2

Sample	% Abs	(% Abs Ref)
Sample #1	0.1780	(-0.0150)
Sample #2	0.1490	(0.0040)
Sample #3	0.1550	(0.0000)
Sample #4	0.1670	(0.0240)
Avg % Abs	0.1570	(0.0027)
STD DEV	0.0092	(0.0023)
REL STD DEV	5.838	(86.503)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 ng/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12788, Sum Io = 13012

CHANN 1

Sample	% Abs	(% Abs Ref)
Sample #1	0.8530	(-0.0120)
Sample #2	0.8470	(0.0050)
Sample #3	0.8510	(0.0090)
Sample #4	0.8520	(0.0080)
Avg % Abs	0.8510	(0.0073)
STD DEV	0.0025	(0.0021)
REL STD DEV	0.311	(28.386)

CHANN 2

Sample	% Abs	(% Abs Ref)
Sample #1	1.5600	(-0.0160)
Sample #2	1.5470	(0.0070)
Sample #3	1.5600	(0.0000)
Sample #4	1.5450	(0.0070)
Avg % Abs	1.5507	(0.0047)
STD DEV	0.0081	(0.0040)
REL STD DEV	0.525	(86.603)

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 ng/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12782, Sum Io = 13008

CHANN 1

Sample	% Abs	(% Abs Ref)
Sample #1	1.9580	(-0.0080)
Sample #2	1.9690	(-0.0060)
Sample #3	1.9480	(0.0150)
Sample #4	1.9390	(0.0240)
Avg % Abs	1.9520	(0.0110)
STD DEV	0.0154	(0.0154)
REL STD DEV	0.789	(139.953)

CHANN 2

Sample	% Abs	(% Abs Ref)
Sample #1	3.5720	(-0.0190)
Sample #2	3.5660	(0.0050)
Sample #3	3.5490	(0.0040)
Sample #4	3.5440	(0.0170)
Avg % Abs	3.5530	(0.0087)
STD DEV	0.0115	(0.0072)
REL STD DEV	0.325	(83.471)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 ng/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12780, Sum Io = 13009

CHANN 1

Sample	% Abs	(% Abs Ref)
Sample #1	3.7550	(-0.0070)
Sample #2	3.7440	(0.0270)
Sample #3	3.7320	(0.0310)
Sample #4	3.7250	(0.0370)
Avg % Abs	3.7340	(0.0317)
STD DEV	0.0092	(0.0050)
REL STD DEV	0.245	(15.894)

CHANN 2

Sample	% Abs	(% Abs Ref)
Sample #1	6.7490	(0.0000)
Sample #2	6.6970	(0.0460)
Sample #3	6.6900	(0.0380)
Sample #4	6.6920	(0.0540)
Avg % Abs	6.6930	(0.0460)
STD DEV	0.0036	(0.0080)
REL STD DEV	0.054	(17.391)

Sol Value = 0.300 g/210L ***
Fit Value = 1.4286 ng/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12780, Sum Io = 13006

CHANN 1

Sample	% Abs	(% Abs Ref)
Sample #1	5.5280	(-0.0200)
Sample #2	5.5170	(0.0220)
Sample #3	5.4930	(0.0460)
Sample #4	5.4910	(0.0580)
Avg % Abs	5.5003	(0.0420)
STD DEV	0.0145	(0.0183)
REL STD DEV	0.263	(43.644)

CHANN 2

Sample	% Abs	(% Abs Ref)
Sample #1	9.8120	(-0.0050)
Sample #2	9.7610	(0.0510)
Sample #3	9.7570	(0.0610)
Sample #4	9.7440	(0.0680)
Avg % Abs	9.7540	(0.0600)
STD DEV	0.0089	(0.0085)
REL STD DEV	0.091	(14.240)

AUTO CAL DATA

CHANN 1

Sol Val	% Abs	Rel Std Dev
0.0000 ng/l or 0.000 g/210L	0.083	6.66
0.001	0.041	0.0005
0.002	0.040	-0.0005
0.003	0.100	-0.0003
0.004	0.199	0.0007
0.005	0.300	-0.0002

CHANN 2

Sol Val	% Abs	Rel Std Dev
0.0000 ng/l or 0.000 g/210L	0.157	5.84
0.001	0.1905 ng/l or 0.040 g/210L	1.551
0.002	0.4762 ng/l or 0.100 g/210L	3.553
0.003	0.001	0.32
0.004	0.9524 ng/l or 0.200 g/210L	6.693
0.005	1.4286 ng/l or 0.300 g/210L	9.754
0.006	0.001	0.09
0.007	-0.234	0.08
0.008	1370.08	12.29
0.009	26.866159	

CHANN 2

Sol Val	% Abs	Rel Std Dev
0.0000 ng/l or 0.000 g/210L	0.157	5.84
0.001	0.1905 ng/l or 0.040 g/210L	1.551
0.002	0.4762 ng/l or 0.100 g/210L	3.553
0.003	0.001	0.32
0.004	0.9524 ng/l or 0.200 g/210L	6.693
0.005	1.4286 ng/l or 0.300 g/210L	9.754
0.006	0.001	0.09
0.007	-0.234	0.08
0.008	1370.08	12.29
0.009	26.866159	

Solution Stats Quadratic Fit Chan 1

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0005
0.040	0.041	-0.0005
0.100	0.100	-0.0003
0.200	0.199	0.0007
0.300	0.300	-0.0002

Solution Stats Quadratic Fit Chan 2

Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0004
0.040	0.040	-0.0003
0.100	0.101	-0.0006
0.200	0.199	0.0008
0.300	0.300	-0.0003

Sol Value = 0.080 g/210L ***
Fit Value = 0.3810 ng/l %%%
Samples Taken = 4, Discarded = 1
CHANN 1
Sample #1 = 3290.00
Sample #2 = 3297.00
Sample #3 = 3285.00
Sample #4 = 3290.00
Average Result = 3290.6667
STD DEV = 6.0277
REL STD DEV = 0.183
CHANN 2
Sample #1 = 3317.00
Sample #2 = 3335.00
Sample #3 = 3325.00
Sample #4 = 3360.00
Average Result = 3340.0000
STD DEV = 18.0278
REL STD DEV = 0.540
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1025
3 in H2O Adjust (mg/l*10.000) = 519
9 in H2O Adjust (mg/l*10.000) = 469
***** AUTO CAL PASS

Bryan Siding

Post Calibration Adjustment Stability Checks

FORT WALTON BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004770
12/22/2023
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:54	
Control Test	0.050	09:54	
Air Blank	0.000	09:55	
Control Test	0.049	09:55	
Air Blank	0.000	09:56	
Control Test	0.050	09:57	
Air Blank	0.000	09:57	
Control Test Stats			
Average	0.0497		
Std Dev	0.0006		
Rel Std Dev(%)	1.1625		

Benjamin Soliday
Operator's Signature

FORT WALTON BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004770
12/22/2023
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	09:59	
Control Test	0.080	09:59	
Air Blank	0.000	10:00	
Control Test	0.080	10:01	
Air Blank	0.000	10:01	
Control Test	0.079	10:02	
Air Blank	0.000	10:03	
Control Test Stats			
Average	0.0797		
Std Dev	0.0006		
Rel Std Dev(%)	0.7247		

Benjamin Soliday
Operator's Signature

FORT WALTON BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004770
12/22/2023
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	10:15	
Control Test	0.198	10:16	
Air Blank	0.000	10:16	
Control Test	0.197	10:17	
Air Blank	0.000	10:18	
Control Test	0.196	10:18	
Air Blank	0.000	10:19	
Control Test Stats			
Average	0.1970		
Std Dev	0.0010		
Rel Std Dev(%)	0.5076		

Benjamin Soliday
Operator's Signature

FORT WALTON BEACH PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004770
12/22/2023
Software: 8100.27

Test	g/210L	Time	
Air Blank	0.000	10:24	
Control Test	0.079	10:24	
Air Blank	0.000	10:25	
Control Test	0.079	10:25	
Air Blank	0.000	10:26	
Control Test	0.079	10:26	
Air Blank	0.000	10:27	
Control Test Stats			
Average	0.0790		
Std Dev	0.0000		
Rel Std Dev(%)	0.0000		

DGS

Benjamin Soliday
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT WALTON BEACH PD
Time of Inspection: 12:03

Date of Inspection: 12/22/2023

Serial Number: 80-004770
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.049	0.078	0.196	0.079
0.000	0.050	0.078	0.196	0.080
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.050	0.078	0.197	0.079
0.000	0.050	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.050	0.078	0.197	0.079
0.000	0.049	0.078	0.197	0.079
Standard Deviations	0.0005	0.0000	0.0005	0.0003

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BENJAMIN W SIDDOWAY
Signature and Printed Name

12/22/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-004770, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-004770</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FORT WALTON BEACH PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>12/22/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:03</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Benjamin
Siddoway
Digitally signed by
Benjamin Siddoway
Date: 2023.12.22 13:21:02
-05'00'

12/22/2023 Date
BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality