



INSTRUMENT PROCESSING SHEET

Agency Port Orange PDS/N 80-005402Florida Department of
Law EnforcementDate In 12-11-2023 DI Completion Date 12/14/2023☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

2023 (BS 2/9/2024)

Intake	By ALL	Quality Checks	By BS	Date 12/13/2023	Flow Calibration	By BS	Date 12/13/2023																																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>90</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.089</u> (.139 - .169) 36 mm <u>0.101</u> (.156 - .190) 53 mm <u>0.175</u> (.228 - .278) 103 mm <u>0.457</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28427</u> ☒ Stability Checks 12/14/2023			☒ Flow Column # <u>ATP102</u> ☒ 5L/min – 17mm ☒ 15L/min – 53mm ☒ 30L/min – 103mm ☒ R-Value <u>90</u> ☒ Post Calibration Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.156</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.523</u> (.447 - .547)																																															
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K 3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 4/5/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG325603 9/13/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 3/29/2025	0.080	MP5089	202303L 3/29/2025	0.200	MP5090	202304C 4/5/2025	0.080 DGS	N/A	AG325603 9/13/2025	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By _____</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td colspan="2"></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td colspan="2"></td></tr><tr><td>☐ Breath Tube Replacement</td><td colspan="2"></td></tr><tr><td>☐ Other _____</td><td colspan="2"></td></tr><tr><td>_____</td><td colspan="2"></td></tr><tr><td>_____</td><td colspan="2"></td></tr><tr><td>_____</td><td colspan="2"></td></tr><tr><td>_____</td><td colspan="2"></td></tr><tr><td>_____</td><td colspan="2"></td></tr></tbody></table>			Maintenance		By _____	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other _____			_____			_____			_____			_____			_____		
Simulator	Serial #	Lot #/Exp																																																		
0.050	MP5088	202303K 3/29/2025																																																		
0.080	MP5089	202303L 3/29/2025																																																		
0.200	MP5090	202304C 4/5/2025																																																		
0.080 DGS	N/A	AG325603 9/13/2025																																																		
Maintenance		By _____																																																		
☐ Battery Replacement																																																				
☐ Dry Gas Regulator Replacement																																																				
☐ Breath Tube Replacement																																																				
☐ Other _____																																																				

Calibration Adjustment By BS				Department Inspection By BS																																																
Barometric Pressure Gauge <u>1031</u> ID # <u>28663</u>				Barometric Pressure ID# <u>28427</u>																																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>22460</td><td>12/28/2024</td></tr><tr><td>0.100</td><td>MP6296</td><td>23390</td><td>10/17/2025</td></tr><tr><td>0.200</td><td>MP6297</td><td>22400</td><td>10/12/2024</td></tr><tr><td>0.300</td><td>MP6298</td><td>23070</td><td>3/6/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>4/5/2025</td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	22460	12/28/2024	0.100	MP6296	23390	10/17/2025	0.200	MP6297	22400	10/12/2024	0.300	MP6298	23070	3/6/2025	0.080 DGS	N/A	06723080A5	4/5/2025	Gauge <u>1028</u> Instrument <u>1030</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2023-B</u>																				
Simulator	Serial #	Lot #	Expiration																																																	
0.000	MP6294	N/A	N/A																																																	
0.040	MP6295	22460	12/28/2024																																																	
0.100	MP6296	23390	10/17/2025																																																	
0.200	MP6297	22400	10/12/2024																																																	
0.300	MP6298	23070	3/6/2025																																																	
0.080 DGS	N/A	06723080A5	4/5/2025																																																	
☒ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090																																	
Simulator	Serial Number																																																			
0.000	MP5086																																																			
Interferent	MP5087																																																			
0.050	MP5088																																																			
0.080	MP5089																																																			
0.200	MP5090																																																			
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K</td><td>3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L</td><td>3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C</td><td>4/5/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG325603</td><td>9/13/2025</td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202303K	3/29/2025	0.080	MP5089	202303L	3/29/2025	0.200	MP5090	202304C	4/5/2025	0.080 DGS	N/A	AG325603	9/13/2025	Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☒ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☒ Other <u>Form 51</u>																												
Simulator	Serial #	Lot #	Expiration																																																	
0.050	MP5088	202303K	3/29/2025																																																	
0.080	MP5089	202303L	3/29/2025																																																	
0.200	MP5090	202304C	4/5/2025																																																	
0.080 DGS	N/A	AG325603	9/13/2025																																																	
Notes/Suggested Service: <u>R-Value<100, send to repair.</u> <u>Quality check barometric pressure read 1016, gauge</u> <u>read 1028. 1.17% difference requires calibration</u> <u>adjustment.</u> <u>Administrative Review: amended typo to DI Completion</u> <u>Date. (BS 2/9/2024)</u> _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																																
Israel Soto Digitally signed by Israel Soto Date: 2023.12.19 09:17:12 -05'00'				Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.12.20 15:00:41 -05'00'																																																
Tech Review / Date _____				Admin Review / Date _____																																																

Stability Checks and Flow Calibration

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005402
12/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:20
Control Test	0.050	08:20
Air Blank	0.000	08:21
Control Test	0.051	08:21
Air Blank	0.000	08:22
Control Test	0.050	08:23
Air Blank	0.000	08:23
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

Benjamin Sidbey
Operator's Signature

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005402
12/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:14
Control Test	0.080	08:15
Air Blank	0.000	08:15
Control Test	0.080	08:16
Air Blank	0.000	08:17
Control Test	0.080	08:17
Air Blank	0.000	08:18
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Sidbey
Operator's Signature

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005402
12/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:07
Control Test	0.197	08:08
Air Blank	0.000	08:08
Control Test	0.198	08:09
Air Blank	0.000	08:09
Control Test	0.199	08:10
Air Blank	0.000	08:11
Control Test Stats		
Average	0.1980	
Std Dev	0.0010	
Rel Std Dev(%)	0.5051	

Benjamin Sidbey
Operator's Signature

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005402
12/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.080	08:36
Air Blank	0.000	08:37
Control Test	0.081	08:37
Air Blank	0.000	08:38
Control Test	0.081	08:38
Air Blank	0.000	08:38
Control Test Stats		
Average	0.0807	
Std Dev	0.0006	
Rel Std Dev(%)	0.7157	

DGS

Benjamin Sidbey
Operator's Signature

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005402
12/13/2023
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5
SQRT(Diff) = 4.355
2: Rate (Liters/min) = 15
SQRT(Diff) = 8.941
3: Rate (Liters/min) = 30
SQRT(Diff) = 19.078
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 649
Rounded Intercept = -126902
Correlation = 0.99507

Benjamin Sidbey

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005402
12/14/2023 08:52:59

```

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/L 888%
Samples Taken = 4, discarded = 1
Sum lo = 12639, Sum lo = 12734
<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = -0.0170      (-0.0030)
Sample #2 = -0.0180      (0.0180)
Sample #3 = -0.0190      (0.0420)
Sample #4 = -0.0120      (0.0510)
Avg % Abs = -0.0153      (0.0370)
STD DEV = 0.0038      (0.0171)
REL STD DEV = 23.179      (46.105)

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = -0.0500  (-0.0040)
Sample #2 = -0.0170  (-0.0140)
Sample #3 = -0.120  (-0.0190)
Sample #4 = -0.110  (-0.0090)
Avg. % Abs = -0.0133  (-0.0140)
STD DEV = 0.0032  (0.0050)
REL STD DEV = 24.109  (35.714)

```

```
Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l 422%
Samples Taken = 4, Discarded = 1
Sum Io = 12626, Num Io = 12732
<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 0.7630      (-0.0038)
Sample #2 = 0.7720      (0.0040)
Sample #3 = 0.8210      (-0.0140)
Sample #4 = 0.8140      (0.0010)
Avg % Abs = 0.8023      (-0.0030)
STD DEV = 0.0265      (0.0096)
REL STD DEV = 3.3033      (321.455)
```

```

Sol Value = 0.100 g/20UL ***
Fit value = 0.4762 mg/L  %%%
Samples Taken = 4, Discarded = 1
Sum Lo = 1265, Sum Lo = 1273!
<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 1.8790      (-0.0180)
Sample #2 = 1.8759      (0.0203)
Sample #3 = 1.8660      (0.0190)
Sample #4 = 1.8140      (0.0310)
Avg % Abs = 1.8517      (0.0173)
STD DEV = 0.0329      (0.0146)
REL STD DEV = 1.778      (84.06%)

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs
Sample #1 = 3.6030
Sample #2 = 3.6400
Sample #3 = 3.5950
Sample #4 = 3.5750
Avg % Abs = 3.6033
STD DEV = 0.0333
REL STD DEV = 0.924

```

```
Sol. Value = 0.210L g/210L ***
Fit value = 0.9524 ng/1 %%%
Samples Taken = 4, Discarded = 1
Sum Lo = 12623, Sum Hi = 12731
<<<< CHANNEL 1 >>>>
Sample      Sample      (% Abs Ref)
Sample #1 = 3.5400 (-0.0070)
Sample #2 = 3.5020 (0.0330)
Sample #3 = 3.5780 (0.0100)
Sample #4 = 3.5180 (0.0400)
Avg % Abs = 3.5327 (0.0277)
STD DEV = 0.0401 (0.0157)
REL STD DEV = 1.134 (56.729)
```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 9.9130   (-0.0130)
Sample #2 = 9.9240   (0.0060)
Sample #3 = 9.8640   (0.0240)
Sample #4 = 9.8790   (0.0270)
Avg % Abs = 9.8957   (0.0190)
STD DEV = 0.0247   (0.0114)
REL STD DEV = 0.249   (59.778)

```

```
<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = -0.013
Std Dev = 0.00 Rel Std Dev = 24.11
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.572
Std Dev = 0.02 Rel Std Dev = 1.46
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.613
Std Dev = 0.03 Rel Std Dev = 0.92
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.835
Std Dev = 0.03 Rel Std Dev = 0.47
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.896
Std Dev = 0.02 Rel Std Dev = 0.25
Zero Order Coef = -38.10
First Order Coef = 1262.24
Second Order Coef = 18.94
Standard Deviation = 58.379265
```

% Abs = 9.896
Std Dev = 0.02 Rel Std Dev = 0.25
Zero Order Coef = -38.10
First Order Coef = 1262.24
Second Order Coef = 18.94
Standard Deviation = 58.379265

Solution Stats Quadratic Fit Chan 1			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.300	-0.002	0.0016	
0.040	0.042	-0.0023	
0.100	0.100	-0.0005	
0.200	0.198	0.0019	
0.300	0.301	-0.0008	

Solution Stats Quadratic Fit Chan 2			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.006	-0.001	0.0012	
0.040	0.042	-0.0018	
0.100	0.106	0.0001	
0.200	0.199	0.0010	
0.300	0.300	-0.0005	

Sol value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%

ENCLOSURE

[illegible]

Sample #1 - 3417 00

Sample #1 = 3417.00

Sample #2 - 3404 00

Sample #2 = 3404.00

Complc H2	3434, 00
Complc H3	3441, 00

Sample #3	2471.00
Sample #4	2471.00

Quercus Dougl. - 3425

STP NEW = 18 5203

DEI STD DEI = 0.541

✕ ✕ ✕ ✕ ✕ ✕ ✕

(END)

Sample #1 = 3389.00

Sample #2 = 3362.00

Sample #3 = 3377.00

Sample #4 = 3381.90

Average Result = 3373

 $2910^{\circ}01 = \text{VED DIS}$
$$\text{REL STD DEV} = 0.297$$

✕✕✕✕✕✕✕✕

Dry Gas H₂O Adjust Re

Barometric Pressure

3 μm H₂O Adjust (mg)

9. **UF H2O Adjust (mg)**

SCHEIDT

3

2

Post Calibration Adjustment Stability Checks

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005402
12/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:59
Control Test	0.050	09:59
Air Blank	0.000	10:00
Control Test	0.050	10:01
Air Blank	0.000	10:01
Control Test	0.050	10:02
Air Blank	0.000	10:02
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Siddley
Operator's Signature

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005402
12/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:13
Control Test	0.078	10:13
Air Blank	0.000	10:14
Control Test	0.079	10:15
Air Blank	0.000	10:15
Control Test	0.079	10:16
Air Blank	0.000	10:16
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Benjamin Siddley
Operator's Signature

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005402
12/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:25
Control Test	0.195	10:26
Air Blank	0.000	10:27
Control Test	0.194	10:27
Air Blank	0.000	10:28
Control Test	0.194	10:28
Air Blank	0.000	10:29
Control Test Stats		
Average	0.1943	
Std Dev	0.0006	
Rel Std Dev(%)	0.2971	

Benjamin Siddley
Operator's Signature

PORT ORANGE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005402
12/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:30
Control Test	0.079	10:30
Air Blank	0.000	10:31
Control Test	0.079	10:31
Air Blank	0.000	10:32
Control Test	0.079	10:32
Air Blank	0.000	10:33
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Benjamin Siddley
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: PORT ORANGE PD
Time of Inspection: 13:58

Date of Inspection: 12/14/2023

Serial Number: 80-005402
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG325603 Exp: 09/13/2025
0.000	0.050	0.079	0.194	0.079
0.000	0.051	0.079	0.195	0.079
0.000	0.050	0.080	0.194	0.079
0.000	0.051	0.079	0.194	0.078
0.000	0.051	0.079	0.194	0.078
0.000	0.051	0.079	0.195	0.078
0.000	0.051	0.079	0.194	0.079
0.000	0.051	0.080	0.194	0.078
0.000	0.051	0.079	0.194	0.078
0.000	0.051	0.079	0.194	0.078


Standard Deviations	0.0004	0.0004	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BENJAMIN W SIDDOWAY

Signature and Printed Name

12/14/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005402, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005402</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>PORT ORANGE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/14/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:58</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Benjamin
Siddoway
Digitally signed by Benjamin
Siddoway
Date: 2023.12.14 14:26:13
-05'00'

12/14/2023

Date

BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: William Harrison on 12/14/2023

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-005402

Bill To Address:

William Harrison

Port Orange PD

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Dept of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

R-Value is low, at 90.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: William Harrison

Phone #: 386-506-5846 Email: wharrison@port-orange.org

ATP Contact Name: Benjamin Siddoway ATP Email: BenjaminSiddoway@fdle.state.fl.us