



# INSTRUMENT PROCESSING SHEET

Agency Manatee CSOS/N 80-006236Florida Department of  
Law EnforcementDate In 02/16/2023 DI Completion Date 02/17/2023☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>02/17/2023</u>	Flow Calibration	By	Date																														
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: <u>Missing middle left foot</u> <u>on dry gas shelf.</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>247</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																
				<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5092</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5093</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5094</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C 01/11/2024	0.080	MP5093	202201D 01/18/2024	0.200	MP5094	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024	<table border="1"><thead><tr><th>Maintenance</th><th>By</th></tr></thead><tbody><tr><td><input type="checkbox"/> Battery Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Breath Tube Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Other _____</td><td></td></tr></tbody></table>			Maintenance	By	<input type="checkbox"/> Battery Replacement		<input type="checkbox"/> Dry Gas Regulator Replacement		<input type="checkbox"/> Breath Tube Replacement		<input type="checkbox"/> Other _____						
Simulator	Serial #	Lot #/Exp																																			
0.050	MP5092	202201C 01/11/2024																																			
0.080	MP5093	202201D 01/18/2024																																			
0.200	MP5094	202201E 01/18/2024																																			
0.080 DGS	N/A	AG223802 08/26/2024																																			
Maintenance	By																																				
<input type="checkbox"/> Battery Replacement																																					
<input type="checkbox"/> Dry Gas Regulator Replacement																																					
<input type="checkbox"/> Breath Tube Replacement																																					
<input type="checkbox"/> Other _____																																					
<b>Calibration Adjustment</b> By _____				<b>Department Inspection</b> By <u>TDG</u>																																	
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28663</u>																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1019</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>					
Simulator	Serial #	Lot #	Expiration																																		
0.000		N/A	N/A																																		
0.040																																					
0.100																																					
0.200																																					
0.300																																					
0.080 DGS	N/A																																				
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5095</td></tr><tr><td>Interferent</td><td>MP5097</td></tr><tr><td>0.050</td><td>MP5092</td></tr><tr><td>0.080</td><td>MP5093</td></tr><tr><td>0.200</td><td>MP5094</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP5095	Interferent	MP5097	0.050	MP5092	0.080	MP5093	0.200	MP5094																		
Simulator	Serial Number																																				
0.000	MP5095																																				
Interferent	MP5097																																				
0.050	MP5092																																				
0.080	MP5093																																				
0.200	MP5094																																				
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td><input checked="" type="checkbox"/> Form 41</td><td><input type="checkbox"/> Post-Stability Checks</td></tr><tr><td><input checked="" type="checkbox"/> Stability Checks</td><td><input type="checkbox"/> Flow Calibration</td></tr><tr><td><input checked="" type="checkbox"/> Calibration Certificate</td><td><input type="checkbox"/> Form 40</td></tr><tr><td><input type="checkbox"/> Calibration Adjustment</td><td><input type="checkbox"/> Other _____</td></tr></tbody></table>				Attachments		<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks	<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration	<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40	<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____
Simulator	Serial #	Lot #	Expiration																																		
0.050																																					
0.080																																					
0.200																																					
0.080 DGS	N/A																																				
Attachments																																					
<input checked="" type="checkbox"/> Form 41	<input type="checkbox"/> Post-Stability Checks																																				
<input checked="" type="checkbox"/> Stability Checks	<input type="checkbox"/> Flow Calibration																																				
<input checked="" type="checkbox"/> Calibration Certificate	<input type="checkbox"/> Form 40																																				
<input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Other _____																																				
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																	
Israel Soto <small>Digitally signed by Israel Soto Date: 2023.02.20 10:03:10 -05'00'</small>				Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.02.21 08:44:09 -05'00'</small>																																	
Tech Review / Date				Admin Review / Date																																	

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 6236	Manatee CSO	02/17/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ✓	0.077 to 0.083 ✓	0.194 to 0.206 ✓	0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-016236 02/17/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr><tr><td>Control Test</td><td>0.049</td><td>11:36</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:36</td></tr><tr><td>Control Test</td><td>0.048</td><td>11:37</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:38</td></tr><tr><td>Control Test</td><td>0.048</td><td>11:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:39</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0483</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>1.1945</td><td></td></tr></table> <p>Operator's Signature <i>MG</i></p>	Test	g/210L	Time	Air Blank	0.000	11:35	Control Test	0.049	11:36	Air Blank	0.000	11:36	Control Test	0.048	11:37	Air Blank	0.000	11:38	Control Test	0.048	11:38	Air Blank	0.000	11:39	Control Test Stats			Average	0.0483		Std Dev	0.0006		Rel Std Dev(%)	1.1945		<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-016236 02/17/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:43</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:44</td></tr><tr><td>Control Test</td><td>0.078</td><td>11:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:46</td></tr><tr><td>Control Test</td><td>0.077</td><td>11:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:47</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0777</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7434</td><td></td></tr></table> <p>Operator's Signature <i>MG</i></p>	Test	g/210L	Time	Air Blank	0.000	11:43	Control Test	0.078	11:44	Air Blank	0.000	11:44	Control Test	0.078	11:45	Air Blank	0.000	11:46	Control Test	0.077	11:47	Air Blank	0.000	11:47	Control Test Stats			Average	0.0777		Std Dev	0.0006		Rel Std Dev(%)	0.7434		<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-016236 02/17/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>11:53</td></tr><tr><td>Control Test</td><td>0.197</td><td>11:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:54</td></tr><tr><td>Control Test</td><td>0.194</td><td>11:55</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:56</td></tr><tr><td>Control Test</td><td>0.195</td><td>11:57</td></tr><tr><td>Air Blank</td><td>0.000</td><td>11:57</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1953</td><td></td></tr><tr><td>Std Dev</td><td>0.0015</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7820</td><td></td></tr></table> <p>Operator's Signature <i>MG</i></p>	Test	g/210L	Time	Air Blank	0.000	11:53	Control Test	0.197	11:54	Air Blank	0.000	11:54	Control Test	0.194	11:55	Air Blank	0.000	11:56	Control Test	0.195	11:57	Air Blank	0.000	11:57	Control Test Stats			Average	0.1953		Std Dev	0.0015		Rel Std Dev(%)	0.7820		<p><b>DGS</b></p> <p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 60-016236 02/17/2023 Software: 8100.27</p> <table><tr><th>Test</th><th>g/210L</th><th>Time</th></tr><tr><td>Air Blank</td><td>0.000</td><td>12:00</td></tr><tr><td>Control Test</td><td>0.078</td><td>12:00</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:00</td></tr><tr><td>Control Test</td><td>0.078</td><td>12:01</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:01</td></tr><tr><td>Control Test</td><td>0.078</td><td>12:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:02</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0780</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></table> <p>Operator's Signature <i>MG</i></p>	Test	g/210L	Time	Air Blank	0.000	12:00	Control Test	0.078	12:00	Air Blank	0.000	12:00	Control Test	0.078	12:01	Air Blank	0.000	12:01	Control Test	0.078	12:02	Air Blank	0.000	12:02	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:35																																																																																																																																																	
Control Test	0.049	11:36																																																																																																																																																	
Air Blank	0.000	11:36																																																																																																																																																	
Control Test	0.048	11:37																																																																																																																																																	
Air Blank	0.000	11:38																																																																																																																																																	
Control Test	0.048	11:38																																																																																																																																																	
Air Blank	0.000	11:39																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0483																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1945																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:43																																																																																																																																																	
Control Test	0.078	11:44																																																																																																																																																	
Air Blank	0.000	11:44																																																																																																																																																	
Control Test	0.078	11:45																																																																																																																																																	
Air Blank	0.000	11:46																																																																																																																																																	
Control Test	0.077	11:47																																																																																																																																																	
Air Blank	0.000	11:47																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0777																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7434																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	11:53																																																																																																																																																	
Control Test	0.197	11:54																																																																																																																																																	
Air Blank	0.000	11:54																																																																																																																																																	
Control Test	0.194	11:55																																																																																																																																																	
Air Blank	0.000	11:56																																																																																																																																																	
Control Test	0.195	11:57																																																																																																																																																	
Air Blank	0.000	11:57																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1953																																																																																																																																																		
Std Dev	0.0015																																																																																																																																																		
Rel Std Dev(%)	0.7820																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:00																																																																																																																																																	
Control Test	0.078	12:00																																																																																																																																																	
Air Blank	0.000	12:00																																																																																																																																																	
Control Test	0.078	12:01																																																																																																																																																	
Air Blank	0.000	12:01																																																																																																																																																	
Control Test	0.078	12:02																																																																																																																																																	
Air Blank	0.000	12:02																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO  
Time of Inspection: 15:48

Date of Inspection: 02/17/2023

Serial Number: 80-006236  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#: 202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#: 202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: AG223802 Exp: 08/26/2024
0.000	0.049	0.079	0.196	0.078
0.000	0.048	0.078	0.195	0.077
0.000	0.048	0.078	0.195	0.078
0.000	0.048	0.078	0.196	0.078
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.077	0.195	0.077
0.000	0.048	0.078	0.196	0.077
0.000	0.048	0.078	0.195	0.077
0.000	0.048	0.078	0.195	0.078


Standard Deviations	0.0003	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
 TAYLOR D GUTSCHOW  
 Signature and Printed Name

02/17/2023  
 Date



## Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006236, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006236</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>MANATEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/17/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:48</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/17/2023

Date

TAYLOR D GUTSCHOW,  
Department Inspector