



# INSTRUMENT PROCESSING SHEET

Agency Gainesville Police DepartmentS/N 80-006384Florida Department of  
Law EnforcementDate In 10/24/2023 DI Completion Date 10/24/2023☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By PN	Quality Checks	By PN	Date <u>10/24/2023</u>	Flow Calibration	By	Date																																								
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE  Visual Inspection: <input checked="" type="checkbox"/> Case <input type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input type="checkbox"/> Screws Tight  Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable  Notes: <u>Instrument Hand Delivered</u> <u>by Agency. PN</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>172</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 103</u> 32 mm <u>.164</u> (.139 - .169) 36 mm <u>.171</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202303K 3/29/2025</td></tr><tr><td>0.080</td><td>MP6292</td><td>202303L 3/29/2025</td></tr><tr><td>0.200</td><td>MP6293</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 10/25/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202303K 3/29/2025	0.080	MP6292	202303L 3/29/2025	0.200	MP6293	202304C 04/05/2025	0.080 DGS	N/A	AG229803 10/25/2024			<b>Maintenance</b> By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____																											
Simulator	Serial #	Lot #/Exp																																													
0.050	MP6291	202303K 3/29/2025																																													
0.080	MP6292	202303L 3/29/2025																																													
0.200	MP6293	202304C 04/05/2025																																													
0.080 DGS	N/A	AG229803 10/25/2024																																													
<b>Calibration Adjustment</b> By _____		<b>Department Inspection</b> By <u>PN</u>																																													
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>30793</u> Gauge <u>1021</u> Instrument <u>1022</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																												
0.000		N/A	N/A																																												
0.040																																															
0.100																																															
0.200																																															
0.300																																															
0.080 DGS	N/A																																														
Simulator	Serial Number																																														
0.000	MP6289																																														
Interferent	MP6290																																														
0.050	MP6291																																														
0.080	MP6292																																														
0.200	MP6293																																														
<input type="checkbox"/> Post Calibration Adjustment Stability Checks		<b>Attachments</b> <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																													
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																									
Simulator	Serial #	Lot #	Expiration																																												
0.050																																															
0.080																																															
0.200																																															
0.080 DGS	N/A																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____		Tech Review / Date _____ Admin Review / Date _____																																													

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GAINESVILLE PD  
Time of Inspection: 11:18

Date of Inspection: 10/24/2023

Serial Number: 80-006384  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.052	0.081	0.203	0.081
0.000	0.052	0.081	0.203	0.081
0.000	0.052	0.081	0.203	0.081
0.000	0.052	0.082	0.203	0.081
0.000	0.052	0.081	0.203	0.081
0.000	0.052	0.081	0.203	0.081
0.000	0.052	0.082	0.204	0.081
0.000	0.052	0.082	0.204	0.081
0.000	0.052	0.081	0.204	0.081
0.000	0.053	0.081	0.204	0.081

Standard Deviations	0.0003	0.0004	0.0005	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PHIL NICODEMO

Signature and Printed Name

10/24/2023  
Date

# STABILITY CHECKS

GAINESVILLE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006384  
10/24/2023  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:37
Control Test	0.052	09:38
Air Blank	0.000	09:38
Control Test	0.052	09:39
Air Blank	0.000	09:40
Control Test	0.052	09:40
Air Blank	0.000	09:41
Control Test Stats		
Average	0.0520	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

GAINESVILLE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006384  
10/24/2023  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:42
Control Test	0.081	09:42
Air Blank	0.000	09:43
Control Test	0.082	09:44
Air Blank	0.000	09:44
Control Test	0.081	09:45
Air Blank	0.000	09:46
Control Test Stats		
Average	0.0813	
Std Dev	0.0006	
Rel Std Dev(%)	0.7099	



Operator's Signature

GAINESVILLE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006384  
10/24/2023  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:47
Control Test	0.203	09:47
Air Blank	0.000	09:48
Control Test	0.203	09:49
Air Blank	0.000	09:49
Control Test	0.203	09:50
Air Blank	0.000	09:50
Control Test Stats		
Average	0.2030	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

GAINESVILLE PD  
Intoxilyzer - Alcohol Analyzer  
Model 8000 SN 80-006384  
10/24/2023  
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:34
Control Test	0.080	09:34
Air Blank	0.000	09:34
Control Test	0.080	09:35
Air Blank	0.000	09:35
Control Test	0.080	09:36
Air Blank	0.000	09:36
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	



Operator's Signature

D65



## Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
2331 Phillips Road.  
Suite B1032  
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006384, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006384</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>GAINESVILLE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/24/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:18</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

### TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

10/24/2023

Date

Phil Nicodemo

PHIL NICODEMO,  
Department Inspector

Digitally signed by Phil  
Nicodemo  
Date: 2023.10.24 11:26:05  
-04'00'

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

*Service • Integrity • Respect • Quality*

Page 1 of 1