



INSTRUMENT PROCESSING SHEET

Agency Monroe CSOS/N 80-006471Florida Department of
Law EnforcementDate In 06/09/2023 DI Completion Date 06/16/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date	06/13/2023	Flow Calibration	By	Date																											
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Missing right support peg for keyboard</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>196</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.160</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks				Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024			<table border="1"><thead><tr><th>Maintenance</th><th>By</th></tr></thead><tbody><tr><td><input type="checkbox"/> Battery Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Dry Gas Regulator Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Breath Tube Replacement</td><td></td></tr><tr><td><input type="checkbox"/> Other _____</td><td></td></tr></tbody></table>		Maintenance	By	<input type="checkbox"/> Battery Replacement		<input type="checkbox"/> Dry Gas Regulator Replacement		<input type="checkbox"/> Breath Tube Replacement		<input type="checkbox"/> Other _____				
Simulator	Serial #	Lot #/Exp																																	
0.050	MP5094	202201C 01/11/2024																																	
0.080	MP5095	202201D 01/18/2024																																	
0.200	MP5096	202201E 01/18/2024																																	
0.080 DGS	N/A	AG223802 08/26/2024																																	
Maintenance	By																																		
<input type="checkbox"/> Battery Replacement																																			
<input type="checkbox"/> Dry Gas Regulator Replacement																																			
<input type="checkbox"/> Breath Tube Replacement																																			
<input type="checkbox"/> Other _____																																			
Calibration Adjustment		By TDG		Department Inspection		By TDG																													
Barometric Pressure Gauge <u>1016</u> ID # <u>28199</u>				Barometric Pressure ID# <u>26932</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>21410</td><td>09/30/2023</td></tr><tr><td>0.100</td><td>MP5099</td><td>22310</td><td>08/11/2024</td></tr><tr><td>0.200</td><td>MP5100</td><td>22050</td><td>02/07/2024</td></tr><tr><td>0.300</td><td>MP5101</td><td>22220</td><td>06/15/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203</td><td>08/10/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	21410	09/30/2023	0.100	MP5099	22310	08/11/2024	0.200	MP5100	22050	02/07/2024	0.300	MP5101	22220	06/15/2024	0.080 DGS	N/A	AG222203	08/10/2024			Gauge <u>1015</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000	MP5097	N/A	N/A																																
0.040	MP5098	21410	09/30/2023																																
0.100	MP5099	22310	08/11/2024																																
0.200	MP5100	22050	02/07/2024																																
0.300	MP5101	22220	06/15/2024																																
0.080 DGS	N/A	AG222203	08/10/2024																																
<input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096			Attachments															
Simulator	Serial Number																																		
0.000	MP5092																																		
Interferent	MP5093																																		
0.050	MP5094																																		
0.080	MP5095																																		
0.200	MP5096																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C</td><td>01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D</td><td>01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E</td><td>01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802</td><td>08/26/2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP5094	202201C	01/11/2024	0.080	MP5095	202201D	01/18/2024	0.200	MP5096	202201E	01/18/2024	0.080 DGS	N/A	AG223802	08/26/2024			<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment		<input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____									
Simulator	Serial #	Lot #	Expiration																																
0.050	MP5094	202201C	01/11/2024																																
0.080	MP5095	202201D	01/18/2024																																
0.200	MP5096	202201E	01/18/2024																																
0.080 DGS	N/A	AG223802	08/26/2024																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																															
				Israel Soto <small>Digitally signed by Israel Soto Date: 2023.06.19 08:25:36 +04'00'</small>		Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.06.19 08:40:34 -04'00'</small>																													
				Tech Review / Date		Admin Review / Date																													

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006471	Monroe CSO	06/13/2023	TDG MC

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	0.077 to 0.083	DGS 0.08g/210L ≤0.003 of Wet
<p>MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/13/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:46 Control Test 0.047 15:47 Air Blank 0.000 15:47 Control Test 0.047 15:48 Air Blank 0.000 15:48 Control Test 0.047 15:49 Air Blank 0.000 15:50 Control Test Stats Average 0.0470 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <u>MC</u></p>	<p>MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/13/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:57 Control Test 0.076 15:57 Air Blank 0.000 15:58 Control Test 0.076 15:59 Air Blank 0.000 15:59 Control Test 0.076 16:00 Air Blank 0.000 16:00 Control Test Stats Average 0.0760 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <u>MC</u></p>	<p>MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/13/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 16:02 Control Test 0.197 16:03 Air Blank 0.000 16:03 Control Test 0.197 16:04 Air Blank 0.000 16:05 Control Test 0.196 16:05 Air Blank 0.000 16:06 Control Test Stats Average 0.1967 Std Dev 0.0006 Rel Std Dev(%) 0.2936</p> <p>Operator's Signature <u>MC</u></p>	<p>MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/13/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:34 Control Test 0.077 15:35 Air Blank 0.000 15:35 Control Test 0.077 15:35 Air Blank 0.000 15:36 Control Test 0.077 15:36 Air Blank 0.000 15:37 Control Test Stats Average 0.0770 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature <u>MC</u></p>	

Comments:

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Bot-Gal)	80-006471	Monroe CSD	06/16/2023	TDG <i>ML</i>

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/16/2023 Software: 8100.27 SN 80-006471	MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/16/2023 Software: 8100.27 SN 80-006471	MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/16/2023 Software: 8100.27 SN 80-006471	MONROE COUNTY S.O. Intoxilyzer - Alcohol Analyzer Model 8000 06/16/2023 Software: 8100.27 SN 80-006471
065			
Test g/210L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time
Air Blank 0.000 10:57	Air Blank 0.000 11:04	Air Blank 0.000 11:12	Air Blank 0.000 10:47
Control Test 0.048 10:58	Control Test 0.079 11:05	Control Test 0.200 11:13	Control Test 0.078 10:47
Air Blank 0.000 10:59	Air Blank 0.000 11:06	Air Blank 0.000 11:13	Air Blank 0.000 10:48
Control Test 0.049 10:59	Control Test 0.078 11:06	Control Test 0.199 11:14	Control Test 0.078 10:48
Air Blank 0.000 11:00	Air Blank 0.000 11:07	Air Blank 0.000 11:15	Air Blank 0.000 10:49
Control Test 0.049 11:01	Control Test 0.079 11:08	Control Test 0.200 11:15	Control Test 0.079 10:49
Air Blank 0.000 11:01	Air Blank 0.000 11:08	Air Blank 0.000 11:16	Air Blank 0.000 10:50
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0487	Average 0.0787	Average 0.1997	Average 0.0783
Std Dev 0.0006	Std Dev 0.0006	Std Dev 0.0006	Std Dev 0.0006
Rel Std Dev(%) 1.1863	Rel Std Dev(%) 0.7339	Rel Std Dev(%) 0.2892	Rel Std Dev(%) 0.7370
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MONROE COUNTY S.O.
Time of Inspection: 14:25

Date of Inspection: 06/16/2023

Serial Number: 80-006471
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.199	0.077
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.199	0.078
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.078


Standard Deviations	0.0000	0.0000	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

06/16/2023
Date



ANAB
ANSI National Accreditation Board
ACCREDITED
ISO/IEC 17025
FORENSIC CALIBRATION
LABORATORY

Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006471, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006471</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>MONROE COUNTY S.O.</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/16/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:25</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

06/16/2023

Date



TAYLOR D GUTSCHOW,

Department Inspector

Service • Integrity • Respect • Quality