



INSTRUMENT PROCESSING SHEET

Agency Flagler County Sheriff's OfficeS/N 80-006616Florida Department of
Law EnforcementDate In 2/22/2023DI Completion Date 02-24-2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By PN	Quality Checks	By IS	Date <u>02-23-2023</u>	Flow Calibration	By	Date																																										
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>227</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																												
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D 01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 10-25-2024</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG229803 10-25-2024	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td><td></td></tr><tr><td>☐ Other _____</td><td></td><td></td></tr><tr><td>_____</td><td></td><td></td></tr><tr><td>_____</td><td></td><td></td></tr><tr><td>_____</td><td></td><td></td></tr><tr><td>_____</td><td></td><td></td></tr></tbody></table>			Maintenance		By	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other _____			_____			_____			_____			_____		
Simulator	Serial #	Lot #/Exp																																															
0.050	MP6291	202201C 01-11-2024																																															
0.080	MP6292	202201D 01-18-2024																																															
0.200	MP6293	202201E 01-18-2024																																															
0.080 DGS	N/A	AG229803 10-25-2024																																															
Maintenance		By																																															
☐ Battery Replacement																																																	
☐ Dry Gas Regulator Replacement																																																	
☐ Breath Tube Replacement																																																	
☐ Other _____																																																	

Calibration Adjustment		By IS			Department Inspection																																												
Barometric Pressure Gauge <u>1017</u> ID # <u>28427</u>					Barometric Pressure ID# <u>30793</u>																																												
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>21410</td><td>09-30-2023</td></tr><tr><td>0.100</td><td>MP6296</td><td>22310</td><td>08-11-2024</td></tr><tr><td>0.200</td><td>MP6297</td><td>22050</td><td>02-07-2024</td></tr><tr><td>0.300</td><td>MP6298</td><td>22220</td><td>06-15-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>08121080A1</td><td>05-05-2023</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	21410	09-30-2023	0.100	MP6296	22310	08-11-2024	0.200	MP6297	22050	02-07-2024	0.300	MP6298	22220	06-15-2024	0.080 DGS	N/A	08121080A1	05-05-2023				Gauge <u>1021</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																
Simulator	Serial #	Lot #	Expiration																																														
0.000	MP6294	N/A	N/A																																														
0.040	MP6295	21410	09-30-2023																																														
0.100	MP6296	22310	08-11-2024																																														
0.200	MP6297	22050	02-07-2024																																														
0.300	MP6298	22220	06-15-2024																																														
0.080 DGS	N/A	08121080A1	05-05-2023																																														
☒ Post Calibration Adjustment Stability Checks					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293																														
Simulator	Serial Number																																																
0.000	MP6289																																																
Interferent	MP6290																																																
0.050	MP6291																																																
0.080	MP6292																																																
0.200	MP6293																																																
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C</td><td>01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D</td><td>01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E</td><td>01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803</td><td>10-25-2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202201C	01-11-2024	0.080	MP6292	202201D	01-18-2024	0.200	MP6293	202201E	01-18-2024	0.080 DGS	N/A	AG229803	10-25-2024				Attachments ☒ Form 41 ☒ Post-Stability Checks ☒ Stability Checks ☐ Flow Calibration ☒ Calibration Certificate ☐ Form 40 ☒ Calibration Adjustment ☐ Other _____																								
Simulator	Serial #	Lot #	Expiration																																														
0.050	MP6291	202201C	01-11-2024																																														
0.080	MP6292	202201D	01-18-2024																																														
0.200	MP6293	202201E	01-18-2024																																														
0.080 DGS	N/A	AG229803	10-25-2024																																														
Notes/Suggested Service: <u>Barometric pressure from instrument compared to barometric pressure gauge greater than 1%. Performed Optical Bench Calibration Adjustment. IS</u> _____ _____ _____ _____ _____					☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use 2023.03.																																												
		Taylor Gutschow			06 13:47:28 05/00																																												
		Tech Review / Date			Admin: Review / Date																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FLAGLER COUNTY SO
Time of Inspection: 14:44

Date of Inspection: 02/24/2023

Serial Number: 80-006616
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.077	0.198	0.081
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.076	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080

Standard Deviations	0.0000	0.0003	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

02/24/2023
Date

Stability checks

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
02/23/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:22
Control Test	0.048	13:23
Air Blank	0.000	13:24
Control Test	0.048	13:24
Air Blank	0.000	13:25
Control Test	0.048	13:26
Air Blank	0.000	13:26
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
02/23/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:43
Control Test	0.077	13:44
Air Blank	0.000	13:44
Control Test	0.078	13:45
Air Blank	0.000	13:45
Control Test	0.077	13:46
Air Blank	0.000	13:47
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
02/23/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:48
Control Test	0.198	13:48
Air Blank	0.000	13:49
Control Test	0.198	13:49
Air Blank	0.000	13:50
Control Test	0.198	13:51
Air Blank	0.000	13:51
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

wet

Operator's Signature

Operator's Signature

Operator's Signature

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
02/23/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:52
Control Test	0.081	13:53
Air Blank	0.000	13:53
Control Test	0.080	13:54
Air Blank	0.000	13:54
Control Test	0.080	13:54
Air Blank	0.000	13:55
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Dry

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006616, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006616</u>	Serial Number:	<u>80-006616</u>
Owning Agency:	<u>FLAGLER COUNTY SO</u>	Owning Agency:	<u>FLAGLER COUNTY SO</u>
Calibration Date:	<u>02/24/2023</u>	Calibration Date:	<u>02/24/2023</u>
Calibration Time:	<u>14:44</u>	Calibration Time:	<u>14:44</u>
		UNCERTAINTY* $\pm$	
		0.050 g/ 210 L	0.004
		0.080 g/ 210 L	0.004
		0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2023.02.24 15:23:53
-05'00'

02/24/2023

Date

ISRAEL SOTO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
02/23/2023 14:06:54

Optical Bench Calibration Adjustment

Auto Calibration

pg 1 of 2

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 0.0870 (-0.0150) 0.1390 (-0.0280)
Sample #2 0.0690 (0.0680) 0.1280 (-0.0050)
Sample #3 0.0410 (0.1360) 0.1260 (0.0180)
Sample #4 0.0010 (0.1760) 0.1050 (0.0340)
Avg % Abs 0.0370 (0.1267) 0.1197 (0.0157)
STD DEV 0.0342 (0.0546) 0.0127 (0.0196)
REL STD DEV 92.368 (43.107) 10.647 (125.135)

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 0.7260 (-0.0070) 1.4940 (0.0040)
Sample #2 0.7170 (0.0120) 1.4800 (-0.0020)
Sample #3 0.7550 (0.0170) 1.5140 (0.0020)
Sample #4 0.7220 (0.0450) 1.5050 (0.0110)
Avg % Abs 0.7313 (0.0247) 1.4997 (0.0037)
STD DEV 0.0206 (0.0178) 0.0176 (0.0067)
REL STD DEV 2.823 (72.104) 1.175 (181.591)

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 1.8170 (-0.0180) 3.5390 (0.0000)
Sample #2 1.8140 (0.0050) 3.5480 (-0.0030)
Sample #3 1.7530 (0.0410) 3.5380 (0.0070)
Sample #4 1.8080 (0.0190) 3.5460 (-0.0070)
Avg % Abs 1.7917 (0.0217) 3.5440 (-0.0010)
STD DEV 0.0336 (0.0181) 0.0053 (0.0072)
REL STD DEV 1.876 (83.758) 0.149 (721.110)

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 3.4820 (-0.0060) 6.7780 (-0.0060)
Sample #2 3.4600 (0.0130) 6.7620 (0.0100)
Sample #3 3.4980 (-0.0050) 6.7760 (-0.0040)
Sample #4 3.5070 (-0.0010) 6.7750 (-0.0030)
Avg % Abs 3.4883 (0.0023) 6.7710 (0.0010)
STD DEV 0.0249 (0.0095) 0.0078 (0.0078)
REL STD DEV 0.715 (405.070) 0.115 (781.025)

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1
Sample % Abs (% Abs Ref) % Abs (% Abs Ref)
Sample #1 5.0980 (-0.0050) 9.7750 (-0.0230)
Sample #2 5.0960 (0.0160) 9.7610 (0.0060)
Sample #3 5.1110 (0.0010) 9.7880 (-0.0160)
Sample #4 5.0830 (0.0290) 9.7600 (0.0000)
Avg % Abs 5.0967 (0.0153) 9.7697 (-0.0033)
STD DEV 0.0140 (0.0140) 0.0159 (0.0114)
REL STD DEV 0.275 (91.382) 0.163 (341.174)

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
02/23/2023 14:06:54

Optical Bench Calibration Adjustment

✓

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -72.99
First Order Coef 2631.90
Second Order Coef 36.01

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.001	-0.0005
0.040	0.039	0.0007
0.100	0.100	0.0001
0.200	0.200	-0.0005
0.300	0.300	0.0002

<<<< 9um >>>>

-142.26
1328.01
15.14

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0004
0.040	0.040	0.0004
0.100	0.100	0.0002
0.200	0.200	-0.0004
0.300	0.300	0.0002

<<<< 3um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 3448.00
Sample #2 3502.00
Sample #3 3429.00
Sample #4 3413.00
Avg 3448.0000
STD DEV 47.4447
REL STD DEV 1.376
H2O adjust (mg/l*10k) 361

<<<< 9um >>>>

3305.00
3305.00
3279.00
3299.00
3294.3333
13.6137
0.413
515

Barometric Pressure = 1017

*****CALIBRATION SUCCESSFUL*****

Post Stability Checks

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
02/23/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:25
Control Test	0.049	15:25
Air Blank	0.000	15:26
Control Test	0.048	15:27
Air Blank	0.000	15:27
Control Test	0.048	15:28
Air Blank	0.000	15:28
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Operator's Signature

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
02/23/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:41
Control Test	0.078	15:42
Air Blank	0.000	15:42
Control Test	0.078	15:43
Air Blank	0.000	15:43
Control Test	0.077	15:44
Air Blank	0.000	15:45
Control Test Stats		
Average	0.0777	
Std Dev	0.0006	
Rel Std Dev(%)	0.7434	

Operator's Signature

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
02/23/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:48
Control Test	0.198	15:49
Air Blank	0.000	15:49
Control Test	0.198	15:50
Air Blank	0.000	15:51
Control Test	0.198	15:51
Air Blank	0.000	15:52
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

wet

FLAGLER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006616
02/23/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:58
Control Test	0.080	15:58
Air Blank	0.000	15:59
Control Test	0.079	15:59
Air Blank	0.000	15:59
Control Test	0.080	16:00
Air Blank	0.000	16:00
Control Test Stats		
Average	0.0797	
Std Dev	0.0006	
Rel Std Dev(%)	0.7247	

Operator's Signature

Dry