





Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006621	Kennedy Space Center	05/30/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
<p>KENNEDY SPACE CENTER Intoxilyzer - Alcohol Analyzer Model: 8000 05/30/2023 Software: 8100.27</p> <p>SN 80-006621</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:36</p> <p>Control Test 0.050 11:36</p> <p>Air Blank 0.000 11:37</p> <p>Control Test 0.050 11:38</p> <p>Air Blank 0.000 11:38</p> <p>Control Test 0.050 11:39</p> <p>Air Blank 0.000 11:39</p> <p>Control Test Stats</p> <p>Average 0.0500</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>KENNEDY SPACE CENTER Intoxilyzer - Alcohol Analyzer Model: 8000 05/30/2023 Software: 8100.27</p> <p>SN 80-006621</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:45</p> <p>Control Test 0.079 11:46</p> <p>Air Blank 0.000 11:46</p> <p>Control Test 0.079 11:47</p> <p>Air Blank 0.000 11:47</p> <p>Control Test 0.079 11:48</p> <p>Air Blank 0.000 11:49</p> <p>Control Test Stats</p> <p>Average 0.0790</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>KENNEDY SPACE CENTER Intoxilyzer - Alcohol Analyzer Model: 8000 05/30/2023 Software: 8100.27</p> <p>SN 80-006621</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:52</p> <p>Control Test 0.200 11:53</p> <p>Air Blank 0.000 11:54</p> <p>Control Test 0.199 11:54</p> <p>Air Blank 0.000 11:55</p> <p>Control Test 0.199 11:56</p> <p>Air Blank 0.000 11:56</p> <p>Control Test Stats</p> <p>Average 0.1993</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2896</p>	<p>KENNEDY SPACE CENTER Intoxilyzer - Alcohol Analyzer Model: 8000 05/30/2023 Software: 8100.27</p> <p>SN 80-006621</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:00</p> <p>Control Test 0.081 11:00</p> <p>Air Blank 0.000 11:01</p> <p>Control Test 0.082 11:01</p> <p>Air Blank 0.000 11:02</p> <p>Control Test 0.081 11:02</p> <p>Air Blank 0.000 11:03</p> <p>Control Test Stats</p> <p>Average 0.0813</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7099</p>
<p>Operator's Signature</p> <p>MG</p>	<p>Operator's Signature</p> <p>MG</p>	<p>Operator's Signature</p> <p>MG</p>	<p>Operator's Signature</p> <p>MG</p>

Comments:

## KENNEDY SPACE CENTER

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-006621

06/01/2023

09:01:40

## Auto Calibration

Max Power Res Value = 118

Auto Range Res Value = 91

Sol Value = 0.000 g/210L \*\*\*

Fit value = 0.0000 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12612, Sum Io = 12524

&lt;&lt;&lt;&lt; CHANNEL 1 &gt;&gt;&gt;&gt;

Sample % Abs (% Abs Ref)

Sample #1 = 0.1450 (-0.0090)

Sample #2 = 0.1280 (0.0060)

Sample #3 = 0.0960 (0.0540)

Sample #4 = 0.1460 (0.0600)

Avg % Abs = 0.1233 (0.0400)

STD DEV = 0.0253 (0.0296)

REL STD DEV = 20.533 (73.953)

&lt;&lt;&lt;&lt; CHANNEL 2 &gt;&gt;&gt;&gt;

Sample % Abs (% Abs Ref)

Sample #1 = 0.1800 (-0.0050)

Sample #2 = 0.1710 (0.0040)

Sample #3 = 0.1480 (0.0350)

Sample #4 = 0.1920 (0.0310)

Avg % Abs = 0.1703 (0.0233)

STD DEV = 0.0220 (0.0169)

REL STD DEV = 12.920 (72.267)

Sol Value = 0.040 g/210L \*\*\*

Fit value = 0.1905 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12589, Sum Io = 12512

&lt;&lt;&lt;&lt; CHANNEL 1 &gt;&gt;&gt;&gt;

Sample % Abs (% Abs Ref)

Sample #1 = 0.8280 (-0.0020)

Sample #2 = 0.8740 (-0.0010)

Sample #3 = 0.8320 (0.0250)

Sample #4 = 0.8880 (0.0220)

Avg % Abs = 0.8647 (0.0153)

STD DEV = 0.0291 (0.0142)

REL STD DEV = 3.370 (92.768)

&lt;&lt;&lt;&lt; CHANNEL 2 &gt;&gt;&gt;&gt;

Sample % Abs (% Abs Ref)

Sample #1 = 1.5660 (0.0000)

Sample #2 = 1.5960 (-0.0060)

Sample #3 = 1.5750 (0.0010)

Sample #4 = 1.5890 (0.0090)

Avg % Abs = 1.5867 (0.0013)

STD DEV = 0.0107 (0.0075)

REL STD DEV = 0.674 (562.917)

Sol Value = 0.100 g/210L \*\*\*

Fit value = 0.4762 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12571, Sum Io = 12505

&lt;&lt;&lt;&lt; CHANNEL 1 &gt;&gt;&gt;&gt;

Sample % Abs (% Abs Ref)

Sample #1 = 1.8990 (-0.0060)

Sample #2 = 1.9340 (0.0060)

Sample #3 = 1.9070 (0.0180)

Sample #4 = 1.9120 (0.0280)

Avg % Abs = 1.9177 (0.0173)

STD DEV = 0.0144 (0.0110)

REL STD DEV = 0.749 (63.549)

&lt;&lt;&lt;&lt; CHANNEL 2 &gt;&gt;&gt;&gt;

Sample % Abs (% Abs Ref)

Sample #1 = 3.6150 (-0.0090)

Sample #2 = 3.6220 (0.0080)

Sample #3 = 3.6170 (0.0050)

Sample #4 = 3.6020 (0.0120)

Avg % Abs = 3.6137 (0.0083)

STD DEV = 0.0104 (0.0035)

REL STD DEV = 0.288 (42.143)

Sol Value = 0.200 g/210L \*\*\*

Fit value = 0.9524 mg/l %%%

Samples Taken = 4, Discarded = 1

Sum Io = 12562, Sum Io = 12500

&lt;&lt;&lt;&lt; CHANNEL 1 &gt;&gt;&gt;&gt;

Sample % Abs (% Abs Ref)

Sample #1 = 3.6910 (-0.0140)

Sample #2 = 3.6650 (0.0200)

Sample #3 = 3.6710 (0.0060)

Sample #4 = 3.6650 (0.0280)

Avg % Abs = 3.6670 (0.0180)

STD DEV = 0.0035 (0.0111)

REL STD DEV = 0.094 (61.864)

## Optical Calibration

SN: 80-006621

Agency: Kennedy Space Center

Date: 06/01/2023

Quadratic Fit: +/- 0.002g/210L ✓

By: TDG ML

\*\*\*\*\* AUTO CAL DATA \*\*\*\*\*

&lt;&lt;&lt;&lt; CHANNEL 1 &gt;&gt;&gt;&gt;

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.123

Std Dev = 0.03 Rel Std Dev = 20.53

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 0.865

Std Dev = 0.03 Rel Std Dev = 3.37

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 1.918

Std Dev = 0.01 Rel Std Dev = 0.75

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 3.667

Std Dev = 0.00 Rel Std Dev = 0.09

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 5.341

Std Dev = 0.02 Rel Std Dev = 0.44

Zero Order Coef = -329.06

First Order Coef = 2589.19

Second Order Coef = 27.48

Standard Deviation = 19.249487

&lt;&lt;&lt;&lt; CHANNEL 2 &gt;&gt;&gt;&gt;

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.170

Std Dev = 0.02 Rel Std Dev = 12.92

Sol Val = 0.1905 mg/l or 0.040 g/210L

% Abs = 1.587

Std Dev = 0.01 Rel Std Dev = 0.67

Sol Val = 0.4762 mg/l or 0.100 g/210L

% Abs = 3.614

Std Dev = 0.01 Rel Std Dev = 0.29

Sol Val = 0.9524 mg/l or 0.200 g/210L

% Abs = 6.854

Std Dev = 0.01 Rel Std Dev = 0.17

Sol Val = 1.4286 mg/l or 0.300 g/210L

% Abs = 9.897

Std Dev = 0.03 Rel Std Dev = 0.33

Zero Order Coef = -229.18

First Order Coef = 1327.97

Second Order Coef = 13.98

Standard Deviation = 7.068134

&lt;&lt;&lt;&lt; CHANNEL 1 &gt;&gt;&gt;&gt;

Sol Val = 0.0000 mg/l or 0.000 g/210L

% Abs = 0.000

Std Dev = 0.000

Sol Val = 0.040

% Abs = 0.099

Std Dev = 0.200

Sol Val = 0.300

% Abs = 0.001

&lt;&lt;&lt;&lt; CHANNEL 2 &gt;&gt;&gt;&gt;

Sol Val = 0.0000 mg/l or 0.000 g/210L

Act Fit Residual

g/210L g/210L g/210L

0.000 -0.000 0.0001

0.040 0.040 -0.0002

0.100 0.100 0.0002

0.200 0.200 -0.0001

0.300 0.300 0.0000

Sol Value = 0.080 g/210L \*\*\*

Fit value = 0.3810 mg/l %%%

Samples Taken = 4, Discarded = 1

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*

Sample #1 = 3163.00

Sample #2 = 3143.00

Sample #3 = 3091.00

Sample #4 = 3079.00

Average Result = 3104.3333

STD DEV = 34.0196

REL STD DEV = 1.096

\*\*\*\*\* CHANNEL 2 \*\*\*\*\*

Sample #1 = 3261.00

Sample #2 = 3261.00

Sample #3 = 3242.00

Sample #4 = 3244.00

Average Result = 3249.0000

STD DEV = 10.4403

REL STD DEV = 0.321

\*\*\*\*\* CHANNEL 1 \*\*\*\*\*

Dry Gas H2O Adjust Results \*\*\*\*\*

Barometric Pressure = 1013

3 um H2O Adjust (mg/l\*10,000) = 705

9 um H2O Adjust (mg/l\*10,000) = 560

\*\*\*\*\* AUTO CAL PASS \*\*\*\*\*



Type of Test	Serial Number	Agency	Date	Performed By
Stabilities (Post-Cal)	80-00 6621	Kennedy Space Center	06/01/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✓	✓	✓
KENNEDY SPACE CENTER Intoxilyzer - Alcohol Analyzer Model 8000 06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27	KENNEDY SPACE CENTER Intoxilyzer - Alcohol Analyzer Model 8000 06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27	KENNEDY SPACE CENTER Intoxilyzer - Alcohol Analyzer Model 8000 06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27	KENNEDY SPACE CENTER Intoxilyzer - Alcohol Analyzer Model 8000 06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27  06/01/2023 Software: 8100.27
Test g/210L Time Air Blank 0.000 11:47 Control Test 0.048 11:48 Air Blank 0.000 11:48 Control Test 0.048 11:49 Air Blank 0.000 11:49 Control Test 0.048 11:50 Air Blank 0.000 11:51 Control Test Stats Average 0.0480 Std Dev 0.0000 Rel Std Dev(%) 0.0000	Test g/210L Time Air Blank 0.000 12:17 Control Test 0.078 12:18 Air Blank 0.000 12:18 Control Test 0.077 12:19 Air Blank 0.000 12:19 Control Test 0.077 12:20 Air Blank 0.000 12:21 Control Test Stats Average 0.0773 Std Dev 0.0006 Rel Std Dev(%) 0.7466	Test g/210L Time Air Blank 0.000 12:24 Control Test 0.198 12:25 Air Blank 0.000 12:26 Control Test 0.198 12:26 Air Blank 0.000 12:27 Control Test 0.199 12:28 Air Blank 0.000 12:28 Control Test Stats Average 0.1983 Std Dev 0.0006 Rel Std Dev(%) 0.2911	Test g/210L Time Air Blank 0.000 09:54 Control Test 0.079 09:54 Air Blank 0.000 09:55 Control Test 0.079 09:55 Air Blank 0.000 09:56 Control Test 0.079 09:56 Air Blank 0.000 09:56 Control Test Stats Average 0.0790 Std Dev 0.0000 Rel Std Dev(%) 0.0000
MG Operator's Signature	MG Operator's Signature	MG Operator's Signature	MG Operator's Signature

Comments:

# Florida Department of Law Enforcement Alcohol Testing Program

## DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER  
Time of Inspection: 14:36

Date of Inspection: 06/01/2023

Serial Number: 80-006621  
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG115904 Exp: 06/08/2023
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.047	0.077	0.199	0.080
0.000	0.047	0.077	0.200	0.080
0.000	0.047	0.077	0.199	0.080
0.000	0.048	0.077	0.199	0.079
0.000	0.047	0.077	0.200	0.080
0.000	0.048	0.077	0.200	0.080
0.000	0.048	0.077	0.200	0.080
0.000	0.048	0.077	0.200	0.080
0.000	0.048	0.077	0.200	0.079

Standard Deviations	0.0005	0.0000	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow TAYLOR D GUTSCHOW  
Signature and Printed Name

06/01/2023  
Date





# Calibration Certificate

Florida Department of Law Enforcement  
Alcohol Testing Program  
4700 Terminal Drive, Suite 1  
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006621, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006621</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>KENNEDY SPACE CENTER</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>06/01/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>14:36</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within  $\pm 0.005$  or 5%, whichever is greater, of the target alcohol concentration.

\*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ( $k=3$ ).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

## TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/01/2023

Date

  
TAYLOR D GUTSCHOW,

FDLE/ATP Form 69 December 2021  
Issuing Authority: Alcohol Testing Program

Department Inspector

Service • Integrity • Respect • Quality