







INSTRUMENT PROCESSING SHEET

Agency Boca Raton PDS/N 80-006622Florida Department of
Law EnforcementDate In 10/10/2023 DI Completion Date 10/10/2023☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date	10/10/2023	Flow Calibration	By	Date																																																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Missing two back feet on DGS shelf.</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>221</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.164</u> (.139 - .169) 36 mm <u>0.183</u> (.156 - .190) 53 mm <u>0.253</u> (.228 - .278) 103 mm <u>0.527</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks				<input type="checkbox"/> Flow Column # <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # 32 mm (.139 - .169) 36 mm (.156 - .190) 53 mm (.228 - .278) 103 mm (.447 - .547)																																																														
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>						Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024	Maintenance By <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other																																															
Simulator	Serial #	Lot #/Exp																																																																		
0.050	MP5094	202201C 01/11/2024																																																																		
0.080	MP5095	202201D 01/18/2024																																																																		
0.200	MP5096	202201E 01/18/2024																																																																		
0.080 DGS	N/A	AG223802 08/26/2024																																																																		
Calibration Adjustment By TDG Barometric Pressure Gauge <u>1014</u> ID # <u>28199</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>22460</td><td>12/28/2024</td></tr><tr><td>0.100</td><td>MP5099</td><td>22310</td><td>08/11/2024</td></tr><tr><td>0.200</td><td>MP5100</td><td>22050</td><td>02/07/2024</td></tr><tr><td>0.300</td><td>MP5101</td><td>22220</td><td>06/15/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203</td><td>08/10/2024</td></tr></tbody></table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C</td><td>01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D</td><td>01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E</td><td>01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802</td><td>08/26/2024</td></tr></tbody></table>						Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	22460	12/28/2024	0.100	MP5099	22310	08/11/2024	0.200	MP5100	22050	02/07/2024	0.300	MP5101	22220	06/15/2024	0.080 DGS	N/A	AG222203	08/10/2024	Simulator	Serial #	Lot #	Expiration	0.050	MP5094	202201C	01/11/2024	0.080	MP5095	202201D	01/18/2024	0.200	MP5096	202201E	01/18/2024	0.080 DGS	N/A	AG223802	08/26/2024	Department Inspection By TDG Barometric Pressure ID# <u>26932</u> Gauge <u>1013</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP4863</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment <input checked="" type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other			Simulator	Serial Number	0.000	MP4863	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																																																	
0.000	MP5097	N/A	N/A																																																																	
0.040	MP5098	22460	12/28/2024																																																																	
0.100	MP5099	22310	08/11/2024																																																																	
0.200	MP5100	22050	02/07/2024																																																																	
0.300	MP5101	22220	06/15/2024																																																																	
0.080 DGS	N/A	AG222203	08/10/2024																																																																	
Simulator	Serial #	Lot #	Expiration																																																																	
0.050	MP5094	202201C	01/11/2024																																																																	
0.080	MP5095	202201D	01/18/2024																																																																	
0.200	MP5096	202201E	01/18/2024																																																																	
0.080 DGS	N/A	AG223802	08/26/2024																																																																	
Simulator	Serial Number																																																																			
0.000	MP4863																																																																			
Interferent	MP5093																																																																			
0.050	MP5094																																																																			
0.080	MP5095																																																																			
0.200	MP5096																																																																			
Notes/Suggested Service:																																																																				
<div><div><input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use</div><div>Israel Soto <small>Digitally signed by Israel Soto Date: 2023.10.10 14:21:27 +0400</small></div><div>Phil Nicodemus <small>Digitally signed by Phil Nicodemus Date: 2023.10.10 14:32:25 -0400</small></div><div>Tech Review / Date</div><div>Admin Review / Date</div></div>																																																																				

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
✓	✗	✓	✓
<p>BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006622 10/10/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:31</p> <p>Control Test 0.048 09:32</p> <p>Air Blank 0.000 09:32</p> <p>Control Test 0.048 09:33</p> <p>Air Blank 0.000 09:33</p> <p>Control Test 0.048 09:34</p> <p>Control Test Stats</p> <p>Average 0.0480</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006622 10/10/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:37</p> <p>Control Test 0.077 09:38</p> <p>Air Blank 0.000 09:39</p> <p>Control Test 0.077 09:39</p> <p>Air Blank 0.000 09:40</p> <p>Control Test 0.076 09:41</p> <p>Air Blank 0.000 09:41</p> <p>Control Test Stats</p> <p>Average 0.0767</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7531</p> <p>Operator's Signature </p>	<p>BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006622 10/10/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:44</p> <p>Control Test 0.197 09:44</p> <p>Air Blank 0.000 09:45</p> <p>Control Test 0.197 09:46</p> <p>Air Blank 0.000 09:46</p> <p>Control Test 0.198 09:47</p> <p>Air Blank 0.000 09:47</p> <p>Control Test Stats</p> <p>Average 0.1973</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2925</p> <p>Operator's Signature </p>	<p>BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006622 10/10/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 09:50</p> <p>Control Test 0.079 09:51</p> <p>Air Blank 0.000 09:51</p> <p>Control Test 0.080 09:51</p> <p>Air Blank 0.000 09:52</p> <p>Control Test 0.078 09:52</p> <p>Air Blank 0.000 09:53</p> <p>Control Test Stats</p> <p>Average 0.0790</p> <p>Std Dev 0.0010</p> <p>Rel Std Dev(%) 1.2658</p> <p>Operator's Signature </p>

DGS

BOCA RATON PD
Intoxilyzer - Alcohol Analyzer
Model 8000
10/10/2023 09:57:44
SN 80-006622
09:57:44

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.5940 (-0.0010)
Sample #2 = 1.5810 (0.0010)
Sample #3 = 1.5810 (0.0040)
Sample #4 = 1.5540 (0.0000)
Avg % Abs = 1.5720 (0.0017)
STD DEV = 0.0156 (0.0021)
REL STD DEV = 0.992 (124.900)

Auto Calibration
Max Power Res Value = 35
Auto Range Res Value = 22

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12721, Sum Io = 12745
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1270 (-0.0170)
Sample #2 = 0.1110 (0.0140)
Sample #3 = 0.0920 (0.0570)
Sample #4 = 0.0500 (0.0750)
Avg % Abs = 0.0843 (0.0487)
STD DEV = 0.0312 (0.0313)
REL STD DEV = 37.013 (64.402)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.2770 (-0.0260)
Sample #2 = 0.2490 (-0.0040)
Sample #3 = 0.2370 (0.0030)
Sample #4 = 0.2160 (0.0160)
Avg % Abs = 0.2340 (0.0050)
STD DEV = 0.0167 (0.0101)
REL STD DEV = 7.138 (202.978)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12707, Sum Io = 12740
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.7930 (-0.0140)
Sample #2 = 0.7630 (0.0130)
Sample #3 = 0.7740 (-0.0030)
Sample #4 = 0.7270 (0.0200)
Avg % Abs = 0.7547 (0.0100)
STD DEV = 0.0246 (0.0118)
REL STD DEV = 3.257 (117.898)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 6.7060 (-0.0060)
Sample #2 = 6.6390 (0.0220)
Sample #3 = 6.6730 (-0.0010)
Sample #4 = 6.6700 (0.0100)
Avg % Abs = 6.6607 (0.0103)
STD DEV = 0.0188 (0.0115)
REL STD DEV = 0.283 (111.325)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12695, Sum Io = 12737
<<<<< CHANNEL 1 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 4.9560 (0.0000)
Sample #2 = 4.9640 (0.0060)
Sample #3 = 5.0000 (-0.0120)
Sample #4 = 5.0050 (-0.0040)
Avg % Abs = 4.9897 (-0.0033)
STD DEV = 0.0224 (0.0090)
REL STD DEV = 0.448 (270.555)

<<<<< CHANNEL 2 >>>>>
Sample % Abs (% Abs Ref)
Sample #1 = 9.6230 (-0.0060)
Sample #2 = 9.6210 (0.0040)
Sample #3 = 9.6430 (0.0000)
Sample #4 = 9.6450 (0.0040)
Avg % Abs = 9.6363 (0.0027)
STD DEV = 0.0133 (0.0023)
REL STD DEV = 0.138 (86.603)

***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.084
Std Dev = 0.03 Rel Std Dev = 37.01
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.755
Std Dev = 0.02 Rel Std Dev = 3.26
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.762
Std Dev = 0.03 Rel Std Dev = 1.48
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.387
Std Dev = 0.01 Rel Std Dev = 0.25
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 4.990
Std Dev = 0.02 Rel Std Dev = 0.45
Zero Order Coef = -236.08
First Order Coef = 2809.69
Second Order Coef = 20.31
Standard Deviation = 10.840514

<<<<< CHANNEL 2 >>>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.234
Std Dev = 0.02 Rel Std Dev = 7.14
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.572
Std Dev = 0.02 Rel Std Dev = 0.99
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.521
Std Dev = 0.01 Rel Std Dev = 0.30
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.661
Std Dev = 0.02 Rel Std Dev = 0.28
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.636
Std Dev = 0.01 Rel Std Dev = 0.14
Zero Order Coef = -325.52
First Order Coef = 1399.86
Second Order Coef = 12.04
Standard Deviation = 6.616469

Solution Stats Quadratic Fit Chan 1
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0000
0.040 0.040 0.0002
0.100 0.100 -0.0003
0.200 0.200 0.0002
0.300 0.300 -0.0001

Solution Stats Quadratic Fit Chan 2
Act Fit Residual
g/210L g/210L g/210L
0.000 0.000 -0.0001
0.040 0.040 -0.0000
0.100 0.100 0.0002
0.200 0.200 -0.0002
0.300 0.300 0.0001





Sol Value = 0.080 g/210L ***
Fit value = 0.3810 mg/l %%%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3502.00
Sample #2 = 3442.00
Sample #3 = 3412.00
Sample #4 = 3449.00
Average Result = 3434.3333
STD DEV = 19.6554
REL STD DEV = 0.572

***** CHANNEL 2 *****
Sample #1 = 3197.00
Sample #2 = 3182.00
Sample #3 = 3179.00
Sample #4 = 3182.00
Average Result = 3181.0000
STD DEV = 1.7321
REL STD DEV = 0.054

Dry Gas H2O Adjust Results *****
Barometric Pressure = 1014
3 um H2O Adjust (mg/l*10,000) = 375
9 um H2O Adjust (mg/l*10,000) = 628
***** AUTO CAL PASS *****

Optical Calibration
Adjustment
By: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
<p>BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006622 10/10/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:12</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:13</td></tr> <tr><td>Control Test</td><td>0.048</td><td>11:14</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	11:11	Control Test	0.049	11:11	Air Blank	0.000	11:12	Control Test	0.049	11:12	Air Blank	0.000	11:13	Control Test	0.048	11:14	Air Blank	0.000	11:14	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006622 10/10/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:19</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:20</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:20</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.077</td><td>11:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	11:19	Control Test	0.077	11:20	Air Blank	0.000	11:20	Control Test	0.077	11:21	Air Blank	0.000	11:22	Control Test	0.077	11:22	Air Blank	0.000	11:23	Control Test Stats			Average	0.0770		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006622 10/10/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:30</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test</td><td>0.198</td><td>11:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	11:28	Control Test	0.198	11:29	Air Blank	0.000	11:30	Control Test	0.198	11:30	Air Blank	0.000	11:31	Control Test	0.198	11:32	Air Blank	0.000	11:32	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>BOCA RATON PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-006622 10/10/2023 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>11:05</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:05</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:06</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:06</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:07</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </table> <p>Operator's Signature </p>	Air Blank	0.000	11:05	Control Test	0.079	11:05	Air Blank	0.000	11:06	Control Test	0.079	11:06	Air Blank	0.000	11:07	Control Test	0.078	11:07	Air Blank	0.000	11:08	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Air Blank	0.000	11:11																																																																																																																																					
Control Test	0.049	11:11																																																																																																																																					
Air Blank	0.000	11:12																																																																																																																																					
Control Test	0.049	11:12																																																																																																																																					
Air Blank	0.000	11:13																																																																																																																																					
Control Test	0.048	11:14																																																																																																																																					
Air Blank	0.000	11:14																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0487																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1863																																																																																																																																						
Air Blank	0.000	11:19																																																																																																																																					
Control Test	0.077	11:20																																																																																																																																					
Air Blank	0.000	11:20																																																																																																																																					
Control Test	0.077	11:21																																																																																																																																					
Air Blank	0.000	11:22																																																																																																																																					
Control Test	0.077	11:22																																																																																																																																					
Air Blank	0.000	11:23																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0770																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:28																																																																																																																																					
Control Test	0.198	11:29																																																																																																																																					
Air Blank	0.000	11:30																																																																																																																																					
Control Test	0.198	11:30																																																																																																																																					
Air Blank	0.000	11:31																																																																																																																																					
Control Test	0.198	11:32																																																																																																																																					
Air Blank	0.000	11:32																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1980																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	11:05																																																																																																																																					
Control Test	0.079	11:05																																																																																																																																					
Air Blank	0.000	11:06																																																																																																																																					
Control Test	0.079	11:06																																																																																																																																					
Air Blank	0.000	11:07																																																																																																																																					
Control Test	0.078	11:07																																																																																																																																					
Air Blank	0.000	11:08																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0787																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7339																																																																																																																																						

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BOCA RATON PD
Time of Inspection: 13:33

Date of Inspection: 10/10/2023

Serial Number: 80-006622
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test* (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.048	0.077	0.198	0.079
0.000	0.047	0.077	0.198	0.079
0.000	0.048	0.077	0.198	0.079
0.000	0.048	0.077	0.198	0.078
0.000	0.047	0.077	0.199	0.079
0.000	0.048	0.077	0.198	0.078
0.000	0.047	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.077	0.198	0.079

Standard Deviations	0.0004	0.0000	0.0003	0.0007
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

10/10/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006622, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006622</u>	UNCERTAINTY* \pm
Owning Agency:	<u>BOCA RATON PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/10/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:33</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



10/10/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality