



INSTRUMENT PROCESSING SHEET

Agency Orlando FHPS/N 80-006628Florida Department of
Law EnforcementDate In 9/13/2023 DI Completion Date 9/14/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By BS	Quality Checks	By BS	Date <u>9/14/2023</u>	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>One keyboard screw was missing, replaced it with a new screw.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>213</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.167</u> (.139 - .169) 36 mm <u>0.187</u> (.156 - .190) 53 mm <u>0.257</u> (.228 - .278) 103 mm <u>0.523</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K 3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 4/5/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5 4/5/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 3/29/2025	0.080	MP5089	202303L 3/29/2025	0.200	MP5090	202304C 4/5/2025	0.080 DGS	N/A	06723080A5 4/5/2025	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td><td></td></tr><tr><td>☐ Other _____</td><td></td><td></td></tr></tbody></table>			Maintenance		By	☐ Battery Replacement			☐ Dry Gas Regulator Replacement			☐ Breath Tube Replacement			☐ Other _____											
Simulator	Serial #	Lot #/Exp																																												
0.050	MP5088	202303K 3/29/2025																																												
0.080	MP5089	202303L 3/29/2025																																												
0.200	MP5090	202304C 4/5/2025																																												
0.080 DGS	N/A	06723080A5 4/5/2025																																												
Maintenance		By																																												
☐ Battery Replacement																																														
☐ Dry Gas Regulator Replacement																																														
☐ Breath Tube Replacement																																														
☐ Other _____																																														
Calibration Adjustment		By _____		Department Inspection			By BS _____																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28662</u> Gauge <u>1010</u> Instrument <u>1004</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP5086																																													
Interferent	MP5087																																													
0.050	MP5088																																													
0.080	MP5089																																													
0.200	MP5090																																													
☐ Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☒ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table>			☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☒ Form 40	☐ Calibration Adjustment	☐ Other _____												
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
☒ Form 41	☐ Post-Stability Checks																																													
☒ Stability Checks	☐ Flow Calibration																																													
☒ Calibration Certificate	☒ Form 40																																													
☐ Calibration Adjustment	☐ Other _____																																													
Notes/Suggested Service: _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																										
				<table border="1"><tbody><tr><td>Israel Soto Digitally signed by Israel Soto Date: 2023.09.19 07:48:19 +0400</td><td>Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.09.19 12:45:03 -04'00'</td></tr></tbody></table>				Israel Soto Digitally signed by Israel Soto Date: 2023.09.19 07:48:19 +0400	Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.09.19 12:45:03 -04'00'																																					
Israel Soto Digitally signed by Israel Soto Date: 2023.09.19 07:48:19 +0400	Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.09.19 12:45:03 -04'00'																																													
				Tech Review / Date _____ Admin Review / Date _____																																										

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 13:44

Date of Inspection: 09/14/2023

Serial Number: 80-006628

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.050	0.079	0.197	0.081
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.081
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.081
0.000	0.050	0.078	0.197	0.080
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.080
0.000	0.050	0.079	0.197	0.080

Standard Deviations	0.0000	0.0003	0.0000	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Benjamin W Siddoway

BENJAMIN W SIDDOWAY
Signature and Printed Name

09/14/2023
Date

Stability Checks

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006628
09/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:18
Control Test	0.050	10:18
Air Blank	0.000	10:19
Control Test	0.049	10:20
Air Blank	0.000	10:20
Control Test	0.050	10:21
Air Blank	0.000	10:21
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Benjamin Siddle
Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006628
09/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:28
Control Test	0.079	10:28
Air Blank	0.000	10:29
Control Test	0.079	10:30
Air Blank	0.000	10:30
Control Test	0.079	10:31
Air Blank	0.000	10:31
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Siddle
Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006628
09/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:01
Control Test	0.197	10:01
Air Blank	0.000	10:02
Control Test	0.198	10:03
Air Blank	0.000	10:03
Control Test	0.197	10:04
Air Blank	0.000	10:05
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

Benjamin Siddle
Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006628
09/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.080	10:12
Air Blank	0.000	10:13
Control Test	0.080	10:13
Air Blank	0.000	10:14
Control Test	0.080	10:14
Air Blank	0.000	10:14
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Benjamin Siddle
Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006628 , manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006628</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/14/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:44</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Benjamin
Siddoway
Digitally signed by
Benjamin Siddoway
Date: 2023.09.14 15:07:20
-04'00'

09/14/2023

Date

BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 09:23

Date of Inspection: 09/14/2023

Serial Number: 80-006628

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, BYPASS FOR DI

compliance not determined

The above instrument ~~complies (X)~~ does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Benjamin Siddoway

Signature and Printed Name

BENJAMIN W SIDDOWAY

09/14/2023
Date