



INSTRUMENT PROCESSING SHEET

Agency Manatee CSOS/N 80-006694Florida Department of
Law EnforcementDate In 11/07/2023 DI Completion Date 11/16/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>11/08/2023</u>	Flow Calibration	By	Date																																							
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>AI reports instrument will not print forms. Instrument was hand-delivered with box to ship back to the agency. Instrument arrived with regulator detached from DGS shelf and missing its plate/screws.</u>		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>206</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.511</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP5095</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5096</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5094	202303K 03/29/2025	0.080	MP5095	202303L 03/29/2025	0.200	MP5096	202304C 04/05/2025	0.080 DGS	N/A	AG223802 08/26/2024																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP5094	202303K 03/29/2025																																												
0.080	MP5095	202303L 03/29/2025																																												
0.200	MP5096	202304C 04/05/2025																																												
0.080 DGS	N/A	AG223802 08/26/2024																																												
Calibration Adjustment		By _____		Department Inspection			By TDG _____																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>26932 (x2)</u> Gauge <u>1017 / 1012</u> Instrument <u>1018 / 1013</u> Mouth Alcohol Solution Lot # <u>2023-A (x2)</u> Acetone Stock Solution Lot # <u>2022-B (x2)</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092 (x2)</td></tr><tr><td>Interferent</td><td>MP5093 (x2)</td></tr><tr><td>0.050</td><td>MP5094 (x2)</td></tr><tr><td>0.080</td><td>MP5095 (x2)</td></tr><tr><td>0.200</td><td>MP5096 (x2)</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5092 (x2)	Interferent	MP5093 (x2)	0.050	MP5094 (x2)	0.080	MP5095 (x2)	0.200	MP5096 (x2)
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP5092 (x2)																																													
Interferent	MP5093 (x2)																																													
0.050	MP5094 (x2)																																													
0.080	MP5095 (x2)																																													
0.200	MP5096 (x2)																																													
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																						
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
Notes/Suggested Service: <u>Tested forms reload by printing Form 40. (TDG)</u>				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																																										
<u>Conducted successful Department Inspection on 11/14. While attaching printer cord to print Form 41, the power cord loosened slightly and the instrument powered off without saving the Department Inspection. Conducted another Department Inspection on 11/16. (TDG)</u>				Israel Soto <small>Digitally signed by Israel Soto Date: 2023.11.16 15:16:54 +05'00'</small> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.11.17 13:27:41 -05'00'</small>																																										
				Tech Review / Date _____ Admin Review / Date _____																																										

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 13:28

Date of Inspection: 11/08/2023

Serial Number: 80-006694
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

TEST OF FORMS RELOAD. AI NOT CONDUCTED.

Not determined

mg
11/08/2023

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.





Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

11/08/2023
Date

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2023 Software: 8100.27</p> <p>SN 80-006694</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:06</p> <p>Control Test 0.050 13:06</p> <p>Air Blank 0.000 13:07</p> <p>Control Test 0.049 13:08</p> <p>Air Blank 0.000 13:08</p> <p>Control Test 0.050 13:09</p> <p>Air Blank 0.000 13:09</p> <p>Control Test Stats</p> <p>Average 0.0497</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1625</p> <p>Operator's Signature </p>	<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2023 Software: 8100.27</p> <p>SN 80-006694</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:12</p> <p>Control Test 0.079 13:12</p> <p>Air Blank 0.000 13:13</p> <p>Control Test 0.079 13:14</p> <p>Air Blank 0.000 13:14</p> <p>Control Test 0.078 13:15</p> <p>Air Blank 0.000 13:16</p> <p>Control Test Stats</p> <p>Average 0.0787</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7339</p> <p>Operator's Signature </p>	<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2023 Software: 8100.27</p> <p>SN 80-006694</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:19</p> <p>Control Test 0.196 13:19</p> <p>Air Blank 0.000 13:20</p> <p>Control Test 0.196 13:21</p> <p>Air Blank 0.000 13:21</p> <p>Control Test 0.196 13:22</p> <p>Air Blank 0.000 13:23</p> <p>Control Test Stats</p> <p>Average 0.1960</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>	<p>MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2023 Software: 8100.27</p> <p>SN 80-006694</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 13:24</p> <p>Control Test 0.079 13:24</p> <p>Air Blank 0.000 13:25</p> <p>Control Test 0.080 13:25</p> <p>Air Blank 0.000 13:26</p> <p>Control Test 0.080 13:26</p> <p>Air Blank 0.000 13:26</p> <p>Control Test Stats</p> <p>Average 0.0797</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7247</p> <p>Operator's Signature </p>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 12:12

Date of Inspection: 11/16/2023

Serial Number: 80-006694
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test* (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.078	0.197	0.079
0.000	0.050	0.078	0.197	0.079
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.079
0.000	0.051	0.079	0.198	0.079

Standard Deviations	0.0004	0.0004	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow TAYLOR D GUTSCHOW
Signature and Printed Name

11/16/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006694, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006694</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>MANATEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>11/16/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:12</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

11/16/2023

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Manatee CSOS/N 80-006694Florida Department of
Law EnforcementDate In 06/22/2023 DI Completion Date 06/23/2023☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date 06/23/2023	Flow Calibration	By	Date																																						
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>210</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.250</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																								
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5094</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5095</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5096</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5094	202201C 01/11/2024	0.080	MP5095	202201D 01/18/2024	0.200	MP5096	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024			Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____ _____																									
Simulator	Serial #	Lot #/Exp																																											
0.050	MP5094	202201C 01/11/2024																																											
0.080	MP5095	202201D 01/18/2024																																											
0.200	MP5096	202201E 01/18/2024																																											
0.080 DGS	N/A	AG223802 08/26/2024																																											
Calibration Adjustment By _____		Department Inspection By TDG _____																																											
Barometric Pressure Gauge _____ ID # _____		Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1018</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2022-B</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5092</td></tr><tr><td>Interferent</td><td>MP5093</td></tr><tr><td>0.050</td><td>MP5094</td></tr><tr><td>0.080</td><td>MP5095</td></tr><tr><td>0.200</td><td>MP5096</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP5092	Interferent	MP5093	0.050	MP5094	0.080	MP5095	0.200	MP5096
Simulator	Serial #	Lot #	Expiration																																										
0.000		N/A	N/A																																										
0.040																																													
0.100																																													
0.200																																													
0.300																																													
0.080 DGS	N/A																																												
Simulator	Serial Number																																												
0.000	MP5092																																												
Interferent	MP5093																																												
0.050	MP5094																																												
0.080	MP5095																																												
0.200	MP5096																																												
<input type="checkbox"/> Post Calibration Adjustment Stability Checks		Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																							
Simulator	Serial #	Lot #	Expiration																																										
0.050																																													
0.080																																													
0.200																																													
0.080 DGS	N/A																																												
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		Israel Soto <small>Digitally signed by Israel Soto Date: 2023.06.26 07:29:53 +04'00'</small> Phil Nicodemmo <small>Digitally signed by Phil Nicodemmo Date: 2023.06.26 15:04:48 -04'00'</small> Tech Review / Date _____ Admin Review / Date _____																																											

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-006694	Manatee CSD	06/23/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083
MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/23/2023 Software: 8100.27	MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/23/2023 Software: 8100.27	MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/23/2023 Software: 8100.27	MANATEE COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 06/23/2023 Software: 8100.27
Test g/210L Time	Test g/210L Time	Test g/210L Time	Test g/210L Time
Air Blank 0.000 08:58	Air Blank 0.000 09:08	Air Blank 0.000 09:15	Air Blank 0.000 08:51
Control Test 0.050 08:59	Control Test 0.079 09:09	Control Test 0.198 09:16	Control Test 0.080 08:52
Air Blank 0.000 09:00	Air Blank 0.000 09:09	Air Blank 0.000 09:16	Air Blank 0.000 08:52
Control Test 0.049 09:00	Control Test 0.078 09:10	Control Test 0.197 09:17	Control Test 0.079 08:53
Air Blank 0.000 09:01	Air Blank 0.000 09:10	Air Blank 0.000 09:17	Air Blank 0.000 08:53
Control Test 0.049 09:02	Control Test 0.078 09:11	Control Test 0.196 09:18	Control Test 0.079 08:53
Air Blank 0.000 09:02	Air Blank 0.000 09:12	Air Blank 0.000 09:19	Air Blank 0.000 08:54
Control Test Stats	Control Test Stats	Control Test Stats	Control Test Stats
Average 0.0493	Average 0.0783	Average 0.1970	Average 0.0793
Std Dev 0.0006	Std Dev 0.0006	Std Dev 0.0010	Std Dev 0.0006
Rel Std Dev(%) 1.1703	Rel Std Dev(%) 0.7370	Rel Std Dev(%) 0.5076	Rel Std Dev(%) 0.7277
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: MANATEE COUNTY SO
Time of Inspection: 11:51

Date of Inspection: 06/23/2023

Serial Number: 80-006694
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.079
0.000	0.048	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.079
0.000	0.049	0.078	0.198	0.078
0.000	0.048	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.078
0.000	0.049	0.078	0.198	0.079


Standard Deviations	0.0004	0.0000	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 TAYLOR D GUTSCHOW
Signature and Printed Name

06/23/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006694, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006694</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>MANATEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/23/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:51</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

06/23/2023

Date

Taylor D Gutschow

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1