

Agency Kennedy Space Center

S/N 80-006759

Date In 09-21-2023 DI Completion Date 10/2/2023

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment

By_

Barometric Pressure Gauge _____ ID # _____

❑ Post Calibration Adjustment Stability Checks

Notes/Suggested Service:

Department Inspection

By BSBarometric Pressure ID# 28662Gauge 1016 Instrument 1017Mouth Alcohol Solution Lot # 2022-A

Acetone Stock Solution Lot # 2022-B

Attachments

- | | |
|---|--|
| ☒ Form 41 | ☐ Post-Stability Checks |
| ☒ Stability Checks | ☐ Flow Calibration |
| ☒ Calibration Certificate | ☒ Form 40 |
| ☐ Calibration Adjustment | ☐ Other _____ |

- ☒ Instrument Complies with Chapter 11D-8, FAC
 - ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
 - ☒ Return to/Place into Evidentiary Use
 - ☐ Remain Out of Evidentiary Use
 - ☒ Conduct an Agency Inspection Before Evidentiary Use

Israel Soto Digitally signed by Israel Soto
Date: 2023.10.03 07:53:48 -04'00'

Digitally signed by Israel Soto
Date: 2023.10.03 07:53:48 -04'

Phil Nicodemo

Digitally signed by Phil Nicodemo
Date: 2023.10.03 09:48:49 -04'00'

Tech Review / Date

Admin Review / Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER
Time of Inspection: 11:59

Date of Inspection: 10/02/2023

Serial Number: 80-006759
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.050	0.078	0.197	0.000 / 0.080
0.000	0.050	0.078	0.197	0.000 / 0.080
0.000	0.049	0.078	0.197	0.080 / 0.079
0.000	0.050	0.078	0.196	0.079 / 0.079
0.000	0.049	0.078	0.196	0.079 / 0.079
0.000	0.050	0.078	0.197	0.079 / 0.080
0.000	0.049	0.078	0.196	0.080 / 0.079
0.000	0.050	0.078	0.197	0.079 / 0.079
0.000	0.050	0.078	0.196	0.079 / 0.079
0.000	0.050	0.078	0.197	0.080 / 0.079
Standard Deviations	0.0004	0.0000	0.0005	0.0334 / 0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, INLET UNPLUGGED. PLUGGED DRY GAS STANDARD HOSE INTO INLET AND REPEATED TEST.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Benjamin Siddoway BENJAMIN W SIDDOWAY
Signature and Printed Name
10/02/2023
Date

Stability Checks

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006759
10/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:32
Control Test	0.049	09:32
Air Blank	0.000	09:33
Control Test	0.049	09:34
Air Blank	0.000	09:34
Control Test	0.050	09:35
Air Blank	0.000	09:35
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Bevin Siddiquy
Operator's Signature

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006759
10/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:10
Control Test	0.079	09:10
Air Blank	0.000	09:11
Control Test	0.079	09:12
Air Blank	0.000	09:12
Control Test	0.079	09:13
Air Blank	0.000	09:13
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Bevin Siddiquy
Operator's Signature

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006759
10/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:15
Control Test	0.197	09:15
Air Blank	0.000	09:16
Control Test	0.196	09:17
Air Blank	0.000	09:17
Control Test	0.196	09:18
Air Blank	0.000	09:18
Control Test Stats		
Average	0.1963	
Std Dev	0.0006	
Rel Std Dev(%)	0.2941	

Bevin Siddiquy
Operator's Signature

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006759
10/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:21
Control Test	0.079	09:21
Air Blank	0.000	09:22
Control Test	0.079	09:22
Air Blank	0.000	09:23
Control Test	0.079	09:23
Air Blank	0.000	09:24
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Bevin Siddiquy
Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-006759, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-006759</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>KENNEDY SPACE CENTER</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>10/02/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:59</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Digitally signed by
Benjamin Siddoway
Date: 2023.10.02 15:13:16
-04'00'

Benjamin
Siddoway

10/02/2023

Date

BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: KENNEDY SPACE CENTER
Time of Inspection: 08:50

Date of Inspection: 10/02/2023

Serial Number: 80-006759
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, BYPASS TO OPERATE

N/A compliance not determined

The above instrument ~~complies (X)~~ does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Benjamin Siddoway

BENJAMIN W SIDDOWAY

Signature and Printed Name

10/02/2023
Date

Agency Kennedy Space Center

S/N 80-006759

Date In 02-13-2023 DI Completion Date _____

☐ Ship ☐ P/U ☐ H/D ☐ CMI ☒ EE

Calibration Adjustment

By IS

Simulator	Serial #	Lot #	Expiration
0.000	MP6294	N/A	N/A
0.040	MP6295	21410	09-30-2023
0.100			
0.200			
0.300			
0.080 DGS	N/A		

☐ Post Calibration Adjustment Stability Checks

Simulator	Serial #	Lot #	Expiration
0.050			
0.080			
0.200			
0.080 DGS	N/A		

Notes/Suggested Service: Values in Stability Checks
outside nominal range. Repeated 0.050 Stability Check
after tightening simulator and ensuring no leaks were
present. Stability Checks performed 02-17-2023.
Instrument failed 0.040 g/210L adjustment during
Optical Bench Calibration Adjustment. Sending
instrument to repair. Compliance with 11D-8 not
determined. IS

Department Inspection

By _____

Gauge _____ Instrument _____

Mouth Alcohol Solution Lot # _____

Acetone Stock Solution Lot #

Simulator	Serial Number
0.000	
Interferent	
0.050	
0.080	
0.200	

Attachments

- | | |
|--|---|
| ☐ Form 41 | ☐ Post-Stability Checks |
| ☒ Stability Checks | ☐ Flow Calibration |
| ☐ Calibration Certificate | ☐ Form 40 |
| ☒ Calibration Adjustment | ☒ Other Form 51 |

- ☐ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC
-
- ☐ Return to/Place into Evidentiary Use
- ☒ Remain Out of Evidentiary Use
-
- ☐ Conduct an Agency Inspection Before Evidentiary Use

Tech Review / Date

Admin Review / Date

stability checks

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006759
02/17/2023
Software: 8100.27

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006759
02/17/2023
Software: 8100.27

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006759
02/17/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:12
Control Test	0.051	08:13
Air Blank	0.000	08:14
Control Test	0.046	08:14
Air Blank	0.000	08:15
Control Test	0.045	08:16
Air Blank	0.000	08:16
Control Test Stats		
Average	0.0473	
Std Dev	0.0032	
Rel Std Dev(%)	6.7913	

Test	g/210L	Time
Air Blank	0.000	08:21
Control Test	0.049	08:22
Air Blank	0.000	08:22
Control Test	0.048	08:23
Air Blank	0.000	08:24
Control Test	0.048	08:24
Air Blank	0.000	08:25
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	

Test	g/210L	Time
Air Blank	0.000	08:26
Control Test	0.079	08:26
Air Blank	0.000	08:27
Control Test	0.073	08:28
Air Blank	0.000	08:28
Control Test	0.078	08:29
Air Blank	0.000	08:29
Control Test Stats		
Average	0.0767	
Std Dev	0.0032	
Rel Std Dev(%)	4.1929	

0.050
#1

0.050 #2

wet

Operator's Signature

Operator's Signature

Operator's Signature

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006759
02/17/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:30
Control Test	0.197	08:31
Air Blank	0.000	08:32
Control Test	0.197	08:32
Air Blank	0.000	08:33
Control Test	0.198	08:33
Air Blank	0.000	08:34
Control Test Stats		
Average	0.1973	
Std Dev	0.0006	
Rel Std Dev(%)	0.2926	

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006759
02/17/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:36
Control Test	0.081	08:37
Air Blank	0.000	08:37
Control Test	0.080	08:37
Air Blank	0.000	08:38
Control Test	0.082	08:38
Air Blank	0.000	08:39
Control Test Stats		
Average	0.0810	
Std Dev	0.0010	
Rel Std Dev(%)	1.2346	

Dry

Operator's Signature

Operator's Signature

KENNEDY SPACE CENTER
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-006759
02/23/2023 11:57:22

Auto Calibration
Max Power Res Value = 56
Auto Range Res Value = 34

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12660, 9um Io = 13055
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.0620 (0.0010)
Sample #2 = 0.0810 (0.0020)
Sample #3 = 0.0630 (0.0470)
Sample #4 = 0.1000 (0.0500)
Avg % Abs = 0.0813 (0.0330)
STD DEV = 0.0185 (0.0269)
REL STD DEV = 22.749 (81.481)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.1150 (-0.0680)
Sample #2 = 0.1480 (-0.0700)
Sample #3 = 0.1980 (-0.0530)
Sample #4 = 0.3230 (-0.0480)
Avg % Abs = 0.2230 (-0.0570)
STD DEV = 0.0901 (0.0115)
REL STD DEV = 40.421 (20.233)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12644, 9um Io = 13056
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8020 (-0.0040)
Sample #2 = 0.7980 (0.0160)
Sample #3 = 0.7600 (0.0510)
Sample #4 = 0.7580 (0.0720)
Avg % Abs = 0.7720 (0.0463)
STD DEV = 0.0225 (0.0283)
REL STD DEV = 2.920 (61.058)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.4300 (0.1490)
Sample #2 = 1.5770 (-0.1230)
Sample #3 = 1.3130 (0.2050)
Sample #4 = 1.0560 (0.0660)
Avg % Abs = 1.3153 (0.0493)
STD DEV = 0.2605 (0.1646)
REL STD DEV = 19.805 (333.717)

Optical Bench Calibration Adjustment

failed
Instrument 0.040 g/210L test,
ensured simulator did not have
a leak, repeated test, repeat
not successful.

✓

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
3um Io = 12632, 9um Io = 13027
<<<< CHANNEL 1 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 0.8100 (-0.0110)
Sample #2 = 0.7750 (0.0240)
Sample #3 = 0.8140 (0.0170)
Sample #4 = 0.7870 (0.0620)
Avg % Abs = 0.7920 (0.0343)
STD DEV = 0.0200 (0.0242)
REL STD DEV = 2.522 (70.527)

<<<< CHANNEL 2 >>>>
Sample % Abs (% Abs Ref)
Sample #1 = 1.6670 (-0.2370)
Sample #2 = 1.5800 (-0.1940)
Sample #3 = 1.7310 (-0.3500)
Sample #4 = 1.4740 (-0.1060)
Avg % Abs = 1.5950 (-0.2167)
STD DEV = 0.1292 (0.1236)
REL STD DEV = 8.097 (57.032)

**** AUTO CAL FAIL

Return Material Authorization

Ship to: ☐ CMI, Inc.
☒ Enforcement Electronics

Shipment to repair facility authorized by: Karen Joyner on 02-23-2023

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-006759

Bill To Address:

Kennedy Space Center

Florida

Ship to Address:

FDLE Off-Site Mail Facility

Florida Department of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Instrument is not measuring alcohol reference solutions consistently. When attempting to
perform an Optical Bench Calibration Adjustment the instrument could not pass the
adjustment and failed the adjustment.

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Karen Joyner

Phone #: 321-861-6313

Email: karen.joyner@nasa.gov

ATP Contact Name: Israel Soto

ATP Email: israelsoto@fdle.state.fl.us