



INSTRUMENT PROCESSING SHEET

Agency Broward CSOS/N 80-007006Florida Department of
Law EnforcementDate In 02/03/2023 DI Completion Date 02/16/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>02/08/2023</u>	Flow Calibration	By TDG	Date <u>02/08/2023</u>																												
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>128</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.128*</u> (.139 - .169) 36 mm <u>0.148*</u> (.156 - .190) 53 mm <u>0.222*</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>68639</u> <input checked="" type="checkbox"/> Stability Checks			Flow Column # <u>ATP106</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>128</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.144</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547)																														
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5092</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5093</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5094</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C 01/11/2024	0.080	MP5093	202201D 01/18/2024	0.200	MP5094	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024			Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____ _____															
Simulator	Serial #	Lot #/Exp																																	
0.050	MP5092	202201C 01/11/2024																																	
0.080	MP5093	202201D 01/18/2024																																	
0.200	MP5094	202201E 01/18/2024																																	
0.080 DGS	N/A	AG223802 08/26/2024																																	
Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																															
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28663</u>																															
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1020</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>			
Simulator	Serial #	Lot #	Expiration																																
0.000		N/A	N/A																																
0.040																																			
0.100																																			
0.200																																			
0.300																																			
0.080 DGS	N/A																																		
<input type="checkbox"/> Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5095</td></tr><tr><td>Interferent</td><td>MP5097</td></tr><tr><td>0.050</td><td>MP5092</td></tr><tr><td>0.080</td><td>MP5093</td></tr><tr><td>0.200</td><td>MP5094</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP5095	Interferent	MP5097	0.050	MP5092	0.080	MP5093	0.200	MP5094																
Simulator	Serial Number																																		
0.000	MP5095																																		
Interferent	MP5097																																		
0.050	MP5092																																		
0.080	MP5093																																		
0.200	MP5094																																		
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____											
Simulator	Serial #	Lot #	Expiration																																
0.050																																			
0.080																																			
0.200																																			
0.080 DGS	N/A																																		
Notes/Suggested Service: <u>*Flow values were outside the nominal range. Conducted a flow calibration. (TDG)</u> _____ _____ _____ _____ _____ _____ _____ _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use																															
Israel Soto <small>Digitally signed by Israel Soto Date: 2023.02.17 07:48:44 +05'00'</small>				Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.02.17 10:44:45 -05'00'</small>																															
Tech Review / Date _____				Admin Review / Date _____																															

80-007006

Flow Cal

02/08/2023

MG

BROWARD COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007006
02/08/2023
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(DIFF) = 6.633

2: Rate (Liters/min) = 15

SQRT(DIFF) = 11.355

3: Rate (Liters/min) = 30

SQRT(DIFF) = 20.516

Dependent Data Scale Factor = 10000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 695

Rounded Intercept = -616086

Correlation = 0.99772

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-007006	Broward CSD	02/08/2023	TDG MG

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ≤0.003 of Wet																																																																																																																																																
<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007006 02/08/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th><th>g/210L</th><th>Time</th></tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:17</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.049</td><td>14:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test</td><td>0.048</td><td>14:19</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:20</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0487</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1863</td><td></td></tr> </tbody> </table> <p>MG Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	14:16	Control Test	0.049	14:17	Air Blank	0.000	14:17	Control Test	0.049	14:18	Air Blank	0.000	14:18	Control Test	0.048	14:19	Air Blank	0.000	14:20	Control Test Stats			Average	0.0487		Std Dev	0.0006		Rel Std Dev(%)	1.1863		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007006 02/08/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th><th>g/210L</th><th>Time</th></tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:10</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:12</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:12</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:13</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:14</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0780</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>MG Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	14:10	Control Test	0.078	14:11	Air Blank	0.000	14:11	Control Test	0.078	14:12	Air Blank	0.000	14:12	Control Test	0.078	14:13	Air Blank	0.000	14:14	Control Test Stats			Average	0.0780		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007006 02/08/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th><th>g/210L</th><th>Time</th></tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:26</td></tr> <tr><td>Control Test</td><td>0.201</td><td>14:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:27</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:28</td></tr> <tr><td>Control Test</td><td>0.201</td><td>14:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:30</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2007</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2877</td><td></td></tr> </tbody> </table> <p>MG Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	14:26	Control Test	0.201	14:27	Air Blank	0.000	14:27	Control Test	0.200	14:28	Air Blank	0.000	14:28	Control Test	0.201	14:29	Air Blank	0.000	14:30	Control Test Stats			Average	0.2007		Std Dev	0.0006		Rel Std Dev(%)	0.2877		<p>DGS</p> <p>BROWARD COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007006 02/08/2023 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th><th>g/210L</th><th>Time</th></tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:21</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:21</td></tr> <tr><td>Control Test</td><td>0.078</td><td>14:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:23</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0790</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2658</td><td></td></tr> </tbody> </table> <p>MG Operator's Signature</p>	Test	g/210L	Time	Air Blank	0.000	14:21	Control Test	0.080	14:21	Air Blank	0.000	14:21	Control Test	0.078	14:22	Air Blank	0.000	14:22	Control Test	0.079	14:23	Air Blank	0.000	14:23	Control Test Stats			Average	0.0790		Std Dev	0.0010		Rel Std Dev(%)	1.2658	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:16																																																																																																																																																	
Control Test	0.049	14:17																																																																																																																																																	
Air Blank	0.000	14:17																																																																																																																																																	
Control Test	0.049	14:18																																																																																																																																																	
Air Blank	0.000	14:18																																																																																																																																																	
Control Test	0.048	14:19																																																																																																																																																	
Air Blank	0.000	14:20																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0487																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1863																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:10																																																																																																																																																	
Control Test	0.078	14:11																																																																																																																																																	
Air Blank	0.000	14:11																																																																																																																																																	
Control Test	0.078	14:12																																																																																																																																																	
Air Blank	0.000	14:12																																																																																																																																																	
Control Test	0.078	14:13																																																																																																																																																	
Air Blank	0.000	14:14																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0780																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:26																																																																																																																																																	
Control Test	0.201	14:27																																																																																																																																																	
Air Blank	0.000	14:27																																																																																																																																																	
Control Test	0.200	14:28																																																																																																																																																	
Air Blank	0.000	14:28																																																																																																																																																	
Control Test	0.201	14:29																																																																																																																																																	
Air Blank	0.000	14:30																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.2007																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2877																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:21																																																																																																																																																	
Control Test	0.080	14:21																																																																																																																																																	
Air Blank	0.000	14:21																																																																																																																																																	
Control Test	0.078	14:22																																																																																																																																																	
Air Blank	0.000	14:22																																																																																																																																																	
Control Test	0.079	14:23																																																																																																																																																	
Air Blank	0.000	14:23																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0790																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	1.2658																																																																																																																																																		

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BROWARD COUNTY SO
Time of Inspection: 14:23

Date of Inspection: 02/16/2023

Serial Number: 80-007006
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.048	0.078	0.201	0.078
0.000	0.048	0.078	0.201	0.078
0.000	0.048	0.078	0.200	0.077
0.000	0.048	0.078	0.200	0.077
0.000	0.049	0.078	0.200	0.078
0.000	0.049	0.078	0.200	0.078
0.000	0.048	0.079	0.200	0.077
0.000	0.048	0.078	0.201	0.078
0.000	0.049	0.079	0.201	0.078
0.000	0.049	0.078	0.201	0.078
Standard Deviations	0.0005	0.0004	0.0005	0.0004

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow TAYLOR D GUTSCHOW
Signature and Printed Name

02/16/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007006, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007006</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>BROWARD COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/16/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:23</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/16/2023

Date


TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1