



INSTRUMENT PROCESSING SHEET

Agency Suwannee County Sheriff's Office

S/N 80-007382

Florida Department of
Law Enforcement

Date In 03-27-2023

DI Completion Date _____

☐ Ship

☒ P/U

☐ H/D

☐ CMI

☐ EE

Intake	By _____	Quality Checks	By _____	Date _____	Flow Calibration	By _____	Date _____															
<input type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input type="checkbox"/> Case <input type="checkbox"/> Handle <input type="checkbox"/> Keyboard <input type="checkbox"/> Dry Gas Shelf <input type="checkbox"/> Feet <input type="checkbox"/> Breath Tube <input type="checkbox"/> Ports <input type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: Agency Inspector <u>indicated the instrument was</u> <u>receiving DSP Fail during</u> <u>Diagnostic Check.</u> 		<input type="checkbox"/> Breath Tube Screen <input type="checkbox"/> Replace External O-Rings <input type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # _____ <input type="checkbox"/> Stability Checks			Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																	
		<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A				Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ 		
Simulator	Serial #	Lot #/Exp																				
0.050																						
0.080																						
0.200																						
0.080 DGS	N/A																					

Calibration Adjustment	By _____	Department Inspection	By _____																																																												
Barometric Pressure Gauge _____ ID # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A				Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____ <table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> </tr> <tr> <td>Interferent</td> <td></td> </tr> <tr> <td>0.050</td> <td></td> </tr> <tr> <td>0.080</td> <td></td> </tr> <tr> <td>0.200</td> <td></td> </tr> </tbody> </table>	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200		
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000																																																															
Interferent																																																															
0.050																																																															
0.080																																																															
0.200																																																															
Notes/Suggested Service: <u>Instrument has DSP Fail</u> <u>during startup Diagnostic Check, instrument unable to</u> <u>enter Ready Mode. Instrument set to Factory Disabled</u> <u>Mode and returned to agency 07-26-2023. Compliance</u> <u>with 11D-8 not determined. IS</u> 		Attachments <input type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use 																																																													
Tech Review / Date _____		Admin Review / Date _____																																																													



INSTRUMENT PROCESSING SHEET

Agency Suwannee County Sheriff's OfficeS/N 80-007382Florida Department of
Law EnforcementDate In 12-19-2022 DI Completion Date 1/19/2023 PN 1/20/2023
☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By IS _____ <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____	Quality Checks By IS _____ Date <u>01-11-2023</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>257</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>30793</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202201C 01-11-2024</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202201D 01-18-2024</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202201E 01-18-2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403 05-14-2023</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG113403 05-14-2023	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____
Simulator	Serial #	Lot #/Exp															
0.050	MP6291	202201C 01-11-2024															
0.080	MP6292	202201D 01-18-2024															
0.200	MP6293	202201E 01-18-2024															
0.080 DGS	N/A	AG113403 05-14-2023															

Calibration Adjustment By PN _____ Barometric Pressure Gauge <u>1015/1013/1014</u> ID # <u>28421</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5091</td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td>MP5082</td> <td>21410</td> <td>09/30/2023</td> </tr> <tr> <td>0.100</td> <td>MP5083</td> <td>22310</td> <td>08/11/2024</td> </tr> <tr> <td>0.200</td> <td>MP6297</td> <td>22050</td> <td>02/07/2024</td> </tr> <tr> <td>0.300</td> <td>MP5085</td> <td>22220</td> <td>06/15/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>08121080A1</td> <td>05/05/2023</td> </tr> </tbody> </table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP5088</td> <td>202201C</td> <td>01/11/2024</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> <td>202201D</td> <td>01/18/2024</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> <td>202201E</td> <td>01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG113403</td> <td>05/14/2023</td> </tr> </tbody> </table> Notes/Suggested Service: _____ Tech Review: See attachment 'Notes'. PN <u>1/19/23</u> _____ Tech Review: DI Completion Date corrected PN <u>1/20/2023</u> _____ _____ _____	Simulator	Serial #	Lot #	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	21410	09/30/2023	0.100	MP5083	22310	08/11/2024	0.200	MP6297	22050	02/07/2024	0.300	MP5085	22220	06/15/2024	0.080 DGS	N/A	08121080A1	05/05/2023	Simulator	Serial #	Lot #	Expiration	0.050	MP5088	202201C	01/11/2024	0.080	MP5089	202201D	01/18/2024	0.200	MP5084	202201E	01/18/2024	0.080 DGS	N/A	AG113403	05/14/2023	Department Inspection By PN _____ Barometric Pressure ID# <u>28421/30793</u> Gauge <u>1017/1012</u> Instrument <u>1017/1013</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP8087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 X2 <input checked="" type="checkbox"/> Post-Stability Checks X2 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate X2 <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment X3 <input type="checkbox"/> Other <u>Notes</u> <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between;"> <div> Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2023.01.20 13:33:23 -0500</small> </div> <div style="text-align: right;"> 23 09:32:47 05/00 </div> </div> Tech Review / Date _____ Admin Review / Date _____	Simulator	Serial Number	0.000	MP5086	Interferent	MP8087	0.050	MP5088	0.080	MP5089	0.200	MP5084
Simulator	Serial #	Lot #	Expiration																																																										
0.000	MP5091	N/A	N/A																																																										
0.040	MP5082	21410	09/30/2023																																																										
0.100	MP5083	22310	08/11/2024																																																										
0.200	MP6297	22050	02/07/2024																																																										
0.300	MP5085	22220	06/15/2024																																																										
0.080 DGS	N/A	08121080A1	05/05/2023																																																										
Simulator	Serial #	Lot #	Expiration																																																										
0.050	MP5088	202201C	01/11/2024																																																										
0.080	MP5089	202201D	01/18/2024																																																										
0.200	MP5084	202201E	01/18/2024																																																										
0.080 DGS	N/A	AG113403	05/14/2023																																																										
Simulator	Serial Number																																																												
0.000	MP5086																																																												
Interferent	MP8087																																																												
0.050	MP5088																																																												
0.080	MP5089																																																												
0.200	MP5084																																																												

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT – INTOXILYZER 8000

Agency: SUWANNEE COUNTY SO
Time of Inspection: 12:01

Date of Inspection: 01/12/2023

Serial Number: 80-007382
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.078

Standard Deviations	0.0000	0.0000	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

01/12/2023
Date

Stability checks

SUWANNEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:59
Control Test	0.049	13:59
Air Blank	0.000	14:00
Control Test	0.049	14:01
Air Blank	0.000	14:01
Control Test	0.049	14:02
Air Blank	0.000	14:02
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

SUWANNEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:04
Control Test	0.079	14:05
Air Blank	0.000	14:05
Control Test	0.079	14:06
Air Blank	0.000	14:06
Control Test	0.078	14:07
Air Blank	0.000	14:08
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

SUWANNEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:09
Control Test	0.201	14:09
Air Blank	0.000	14:10
Control Test	0.201	14:11
Air Blank	0.000	14:11
Control Test	0.200	14:12
Air Blank	0.000	14:12
Control Test Stats		
Average	0.2007	
Std Dev	0.0006	
Rel Std Dev(%)	0.2877	

Wet

Operator's Signature

Operator's Signature

Operator's Signature

SUWANNEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/11/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:14
Control Test	0.075	14:15
Air Blank	0.000	14:15
Control Test	0.080	14:16
Air Blank	0.000	14:16
Control Test	0.080	14:16
Air Blank	0.000	14:17
Control Test Stats		
Average	0.0783	
Std Dev	0.0029	
Rel Std Dev(%)	3.6852	

Dry

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007382, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007382</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>SUWANNEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/12/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:01</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/12/2023

Date

PHIL NICODEMO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: SUWANNEE COUNTY SO

Time of Inspection: 13:34

Date of Inspection: 01/11/2023

Serial Number: 80-007382

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

N/A Compliance not determined

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

01/11/2023
Date

Florida Department of Law Enforcement

Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SUWANNEE COUNTY SO
Time of Inspection: 15:22

Date of Inspection: 01/19/2023

Serial Number: 80-007382
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080
0.000	0.050	0.078	0.200	0.080
0.000	0.049	0.078	0.200	0.080


Standard Deviations	0.0003	0.0000	0.0000	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0000 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PHIL NICODEMO

Signature and Printed Name

01/19/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007382, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007382</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>SUWANNEE COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/19/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:22</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

01/19/2023

Date


PHIL NICODEMO,

Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

SUWANNEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/18/2023 12:35:02

#1

Auto Calibration

pg 1 of 2

<<<<< 3um >>>>>			<<<<< 9um >>>>>		

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.0880	(-0.0030)	0.0940	(0.0040)	
Sample #2	0.0780	(0.0700)	0.0590	(0.0410)	
Sample #3	0.1140	(0.1110)	0.0780	(0.0580)	
Sample #4	0.0780	(0.1630)	0.0580	(0.0800)	
Avg % Abs	0.0900	(0.1147)	0.0650	(0.0597)	
STD DEV	0.0208	(0.0466)	0.0113	(0.0196)	
REL STD DEV	23.094	(40.647)	17.338	(32.771)	

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	0.7920	(-0.0030)	1.4240	(-0.0150)	
Sample #2	0.7900	(0.0270)	1.4180	(0.0070)	
Sample #3	0.8130	(0.0460)	1.4200	(-0.0010)	
Sample #4	0.7970	(0.0780)	1.4360	(0.0120)	
Avg % Abs	0.8000	(0.0503)	1.4247	(0.0060)	
STD DEV	0.0118	(0.0258)	0.0099	(0.0066)	
REL STD DEV	1.474	(51.208)	0.692	(109.291)	

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	1.9130	(-0.0040)	3.4240	(0.0110)	
Sample #2	1.9020	(0.0250)	3.4020	(0.0230)	
Sample #3	1.8830	(0.0570)	3.4000	(0.0300)	
Sample #4	1.8720	(0.0760)	3.4110	(0.0420)	
Avg % Abs	1.8857	(0.0527)	3.4043	(0.0317)	
STD DEV	0.0152	(0.0258)	0.0059	(0.0096)	
REL STD DEV	0.805	(48.939)	0.172	(30.344)	

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	5.3260	(-0.0070)	9.5110	(0.0000)	
Sample #2	5.3000	(0.0340)	9.4890	(0.0300)	
Sample #3	5.2690	(0.0650)	9.4920	(0.0500)	
Sample #4	5.3040	(0.0670)	9.4890	(0.0470)	
Avg % Abs	5.2910	(0.0553)	9.4900	(0.0423)	
STD DEV	0.0192	(0.0185)	0.0017	(0.0108)	
REL STD DEV	0.362	(33.438)	0.018	(25.478)	

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1					
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)	
Sample #1	3.6260	(-0.0110)	6.5360	(0.0040)	
Sample #2	3.6460	(0.0000)	6.5560	(0.0070)	
Sample #3	3.6460	(0.0160)	6.5820	(0.0040)	
Sample #4	3.6330	(0.0390)	6.5610	(0.0250)	
Avg % Abs	3.6417	(0.0183)	6.5663	(0.0120)	
STD DEV	0.0075	(0.0196)	0.0138	(0.0114)	
REL STD DEV	0.206	(106.933)	0.210	(94.648)	

/

SUWANNEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/18/2023 12:35:02

Auto Calibration

pg 2 of 2

<<<<< 3um >>>>>

Zero Order Coef -1912.34
First Order Coef 6303.95
Second Order Coef -742.86

<<<<< 9um >>>>>

-1586.06
3349.24
-216.35

Act Fit Residual
(g/210L) (g/210L) (g/210L)
0.000 -0.028 0.0284
0.040 0.056 -0.0158
0.100 0.154 -0.0540
0.200 0.224 -0.0236
0.300 0.235 0.0649

Act Fit Residual
(g/210L) (g/210L) (g/210L)
0.000 -0.029 0.0288
0.040 0.058 -0.0177
0.100 0.153 -0.0535
0.200 0.225 -0.0250
0.300 0.233 0.0674

<<<<< 3um >>>>> <<<<< 9um >>>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 4916.00 5426.00
Sample #2 4929.00 5385.00
Sample #3 4818.00 5407.00
Sample #4 4885.00 5338.00
Avg 4877.3335 5376.6665
STD DEV 55.8957 35.2467
REL STD DEV 1.146 0.656
H2O adjust (mg/l*10k) -1067 -1566

Barometric Pressure = 1014

*****CALIBRATION SUCCESSFUL*****

OPTICAL CALIBRATION ADJUSTMENT #2

INTOXILYZER 8000
Instrument Initialization
14:24 01/18/2023

SUWANNEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/19/2023 08:27:13

Auto Calibration
Max Power Res Value = 61
Auto Range Res Value = 61
Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum lo = 0, Sum hi = 0

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (0.0000)
Sample #2 = 0.0000 (0.0000)
Sample #3 = 0.0000 (0.0000)
Sample #4 = 0.0000 (0.0000)
Avg % Abs = 0.0000 (0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.000 (0.000)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (0.0000)
Sample #2 = 0.0000 (0.0000)
Sample #3 = 0.0000 (0.0000)
Sample #4 = 0.0000 (0.0000)
Avg % Abs = 0.0000 (0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.000 (0.000)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum lo = 0, Sum hi = 0

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (0.0000)
Sample #2 = 0.0000 (0.0000)
Sample #3 = 0.0000 (0.0000)
Sample #4 = 0.0000 (0.0000)
Avg % Abs = 0.0000 (0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.000 (0.000)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (0.0000)
Sample #2 = 0.0000 (0.0000)
Sample #3 = 0.0000 (0.0000)
Sample #4 = 0.0000 (0.0000)
Avg % Abs = 0.0000 (0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.000 (0.000)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum lo = 0, Sum hi = 0

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (0.0000)
Sample #2 = 0.0000 (0.0000)
Sample #3 = 0.0000 (0.0000)
Sample #4 = 0.0000 (0.0000)
Avg % Abs = 0.0000 (0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.000 (0.000)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (0.0000)
Sample #2 = 0.0000 (0.0000)
Sample #3 = 0.0000 (0.0000)
Sample #4 = 0.0000 (0.0000)
Avg % Abs = 0.0000 (0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.000 (0.000)

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum lo = 0, Sum hi = 0

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (0.0000)
Sample #2 = 0.0000 (0.0000)
Sample #3 = 0.0000 (0.0000)
Sample #4 = 0.0000 (0.0000)
Avg % Abs = 0.0000 (0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.000 (0.000)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (0.0000)
Sample #2 = 0.0000 (0.0000)
Sample #3 = 0.0000 (0.0000)
Sample #4 = 0.0000 (0.0000)
Avg % Abs = 0.0000 (0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.000 (0.000)

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum lo = 0, Sum hi = 0

Channel 1 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (0.0000)
Sample #2 = 0.0000 (0.0000)
Sample #3 = 0.0000 (0.0000)
Sample #4 = 0.0000 (0.0000)
Avg % Abs = 0.0000 (0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.000 (0.000)

Channel 2 Data:
Sample % Abs (% Abs Ref)
Sample #1 = 0.0000 (0.0000)
Sample #2 = 0.0000 (0.0000)
Sample #3 = 0.0000 (0.0000)
Sample #4 = 0.0000 (0.0000)
Avg % Abs = 0.0000 (0.0000)
STD DEV = 0.0000 (0.0000)
REL STD DEV = 0.000 (0.000)

***** AUTO CAL DATA *****
Channel 1 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.000
Std Dev = 0.00 Rel Std Dev = 0.00
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.000
Std Dev = 0.00 Rel Std Dev = 0.00
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 0.000
Std Dev = 0.00 Rel Std Dev = 0.00
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 0.000
Std Dev = 0.00 Rel Std Dev = 0.00
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 0.000
Std Dev = 0.00 Rel Std Dev = 0.00
Zero Order Coef = 12190.47
First Order Coef = 0.00
Second Order Coef = -6095.23
Standard Deviation = 8959.469727

Channel 2 Data:
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.000
Std Dev = 0.00 Rel Std Dev = 0.00
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.000
Std Dev = 0.00 Rel Std Dev = 0.00
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 0.000
Std Dev = 0.00 Rel Std Dev = 0.00
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 0.000
Std Dev = 0.00 Rel Std Dev = 0.00
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 0.000
Std Dev = 0.00 Rel Std Dev = 0.00
Zero Order Coef = 12190.47
First Order Coef = 0.00
Second Order Coef = -6095.23
Standard Deviation = 8959.469727

OPTICAL CALIBRATION ADJUSTMENT #2

PA

```

: Solution Stats Quadratic Fit Chan 1 :
: Act      Fit      Residual      :
: g/210L   g/210L   g/210L      :
: 0.000    0.256    -0.2560      :
: 0.040    0.256    -0.2160      :
: 0.100    0.256    -0.1560      :
: 0.200    0.256    -0.0560      :
: 0.300    0.256    0.0440       :

```

```

: Solution Stats Quadratic Fit Chan 2 :
: Act      Fit      Residual      :
: g/210L   g/210L   g/210L      :
: 0.000    0.256    -0.2560      :
: 0.040    0.256    -0.2160      :
: 0.100    0.256    -0.1560      :
: 0.200    0.256    -0.0560      :
: 0.300    0.256    0.0440       :

```

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1

***** CHANNEL 1

Sample #1 = 0.00
 Sample #2 = 0.00
 Sample #3 = 0.00
 Sample #4 = 0.00
 Average Result = 0.0000
 STD DEV = 0.0000
 REL STD DEV = 0.000

***** CHANNEL 2

Sample #1 = 0.00
 Sample #2 = 0.00
 Sample #3 = 0.00
 Sample #4 = 0.00
 Average Result = 0.0000
 STD DEV = 0.0000
 REL STD DEV = 0.000

Dry Gas H2O Adjust Results *****

Barometric Pressure = 1013
 3 um H2O Adjust (mg/l*10,000) = 3809
 9 um H2O Adjust (mg/l*10,000) = 3809

**** AUTO CAL PASS

OPTICAL CALIBRATION ADJUSTMENT #3

[Handwritten signature]

INTOXILYZER 8000
Instrument Initialization
11:35 01/19/2023

SUWANNEE COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/19/2023 11:42:39

Auto Calibration
Max Power Res Value = 85
Auto Range Res Value = 62

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %XX%
Samples Taken = 4, Discarded = 1
Sum lo = 12831, Sum lo = 13145

=====

Sample	% Abs	(% Abs Ref)
Sample #1	0.1040	(-0.0090)
Sample #2	0.1010	(0.0310)
Sample #3	0.0750	(0.0700)
Sample #4	0.0950	(0.0840)
Avg % Abs	0.0903	(0.0617)
STD DEV	0.0136	(0.0275)
REL STD DEV	15.071	(44.538)

=====

=====

Sample	% Abs	(% Abs Ref)
Sample #1	0.0720	(-0.0040)
Sample #2	0.0830	(0.0000)
Sample #3	0.0810	(0.0170)
Sample #4	0.0940	(0.0190)
Avg % Abs	0.0850	(0.0120)
STD DEV	0.0070	(0.0104)
REL STD DEV	8.140	(87.003)

=====

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %XX%
Samples Taken = 4, Discarded = 1
Sum lo = 12815, Sum lo = 13139

=====

Sample	% Abs	(% Abs Ref)
Sample #1	0.8220	(-0.0060)
Sample #2	0.7980	(0.0230)
Sample #3	0.7800	(0.0360)
Sample #4	0.8160	(0.0260)
Avg % Abs	0.7980	(0.0283)
STD DEV	0.0180	(0.0068)
REL STD DEV	2.256	(24.024)

=====

=====

Sample	% Abs	(% Abs Ref)
Sample #1	1.4500	(-0.0120)
Sample #2	1.4200	(0.0050)
Sample #3	1.4160	(0.0040)
Sample #4	1.4480	(0.0100)
Avg % Abs	1.4280	(0.0063)
STD DEV	0.0174	(0.0032)
REL STD DEV	1.221	(50.756)

=====

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %XX%
Samples Taken = 4, Discarded = 1
Sum lo = 12807, Sum lo = 13136

=====

Sample	% Abs	(% Abs Ref)
Sample #1	1.8700	(-0.0180)
Sample #2	1.8690	(0.0040)
Sample #3	1.8650	(0.0080)
Sample #4	1.8870	(0.0020)
Avg % Abs	1.8737	(0.0047)
STD DEV	0.0117	(0.0031)
REL STD DEV	0.625	(65.465)

=====

=====

Sample	% Abs	(% Abs Ref)
Sample #1	3.3920	(0.0000)
Sample #2	3.3840	(0.0150)
Sample #3	3.3850	(0.0190)
Sample #4	3.4110	(0.0070)
Avg % Abs	3.3933	(0.0137)
STD DEV	0.0153	(0.0061)
REL STD DEV	0.451	(44.708)

=====

Sol Value = 0.200 g/210L ***
Fit value = 0.9524 mg/l %XX%
Samples Taken = 4, Discarded = 1
Sum lo = 12804, Sum lo = 13134

=====

Sample	% Abs	(% Abs Ref)
Sample #1	3.6370	(-0.0040)
Sample #2	3.6080	(0.0150)
Sample #3	3.5930	(0.0300)
Sample #4	3.6030	(0.0330)
Avg % Abs	3.6013	(0.0260)
STD DEV	0.0076	(0.0096)
REL STD DEV	0.212	(37.091)

=====

=====

Sample	% Abs	(% Abs Ref)
Sample #1	6.5690	(-0.0140)
Sample #2	6.5070	(0.0270)
Sample #3	6.5020	(0.0380)
Sample #4	6.5180	(0.0290)
Avg % Abs	6.5090	(0.0313)
STD DEV	0.0082	(0.0059)
REL STD DEV	0.126	(18.700)

=====

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %XX%
Samples Taken = 4, Discarded = 1
Sum lo = 12801, Sum lo = 13133

=====

Sample	% Abs	(% Abs Ref)
Sample #1	5.3100	(-0.0140)
Sample #2	5.2950	(0.0060)
Sample #3	5.2540	(0.0260)
Sample #4	5.2980	(0.0140)
Avg % Abs	5.2823	(0.0153)
STD DEV	0.0246	(0.0101)
REL STD DEV	0.465	(65.651)

=====

=====

Sample	% Abs	(% Abs Ref)
Sample #1	9.5240	(-0.0140)
Sample #2	9.5010	(0.0090)
Sample #3	9.4900	(0.0250)
Sample #4	9.4890	(0.0290)
Avg % Abs	9.4933	(0.0210)
STD DEV	0.0067	(0.0106)
REL STD DEV	0.070	(50.395)

=====

=====

***** AUTO CAL DATA *****

Channel	Sol Val	% Abs	Std Dev	Rel Std Dev
1	0.0000 mg/l or 0.000 g/210L	0.090	0.01	15.07
2	0.1905 mg/l or 0.040 g/210L	0.798	0.02	2.26
3	0.4762 mg/l or 0.100 g/210L	1.874	0.01	0.63
4	0.9524 mg/l or 0.200 g/210L	3.601	0.01	0.21
5	1.4286 mg/l or 0.300 g/210L	5.282	0.02	0.47
6	1.9048 mg/l or 0.400 g/210L	6.969	0.02	0.29
7	2.3810 mg/l or 0.500 g/210L	8.756	0.02	0.23
8	2.8572 mg/l or 0.600 g/210L	10.543	0.02	0.19
9	3.3334 mg/l or 0.700 g/210L	12.330	0.02	0.16
10	3.8096 mg/l or 0.800 g/210L	14.117	0.02	0.14
11	4.2858 mg/l or 0.900 g/210L	15.904	0.02	0.13
12	4.7620 mg/l or 1.000 g/210L	17.691	0.02	0.12
13	5.2382 mg/l or 1.100 g/210L	19.478	0.02	0.11
14	5.7144 mg/l or 1.200 g/210L	21.265	0.02	0.10
15	6.1906 mg/l or 1.300 g/210L	23.052	0.02	0.09
16	6.6668 mg/l or 1.400 g/210L	24.839	0.02	0.08
17	7.1430 mg/l or 1.500 g/210L	26.626	0.02	0.07
18	7.6192 mg/l or 1.600 g/210L	28.413	0.02	0.06
19	8.0954 mg/l or 1.700 g/210L	30.200	0.02	0.05
20	8.5716 mg/l or 1.800 g/210L	31.987	0.02	0.04

=====

=====

Channel	Sol Val	% Abs	Std Dev	Rel Std Dev
1	0.0000 mg/l or 0.000 g/210L	0.086	0.01	8.14
2	0.1905 mg/l or 0.040 g/210L	1.428	0.02	1.22
3	0.4762 mg/l or 0.100 g/210L	3.393	0.02	0.45
4	0.9524 mg/l or 0.200 g/210L	6.509	0.01	0.13
5	1.4286 mg/l or 0.300 g/210L	9.493	0.01	0.07
6	1.9048 mg/l or 0.400 g/210L	12.330	0.01	0.04
7	2.3810 mg/l or 0.500 g/210L	15.167	0.01	0.03
8	2.8572 mg/l or 0.600 g/210L	17.691	0.01	0.02
9	3.3334 mg/l or 0.700 g/210L	20.215	0.01	0.01
10	3.8096 mg/l or 0.800 g/210L	22.739	0.01	0.01
11	4.2858 mg/l or 0.900 g/210L	25.263	0.01	0.01
12	4.7620 mg/l or 1.000 g/210L	27.787	0.01	0.01
13	5.2382 mg/l or 1.100 g/210L	30.311	0.01	0.01
14	5.7144 mg/l or 1.200 g/210L	32.835	0.01	0.01
15	6.1906 mg/l or 1.300 g/210L	35.359	0.01	0.01
16	6.6668 mg/l or 1.400 g/210L	37.883	0.01	0.01
17	7.1430 mg/l or 1.500 g/210L	40.407	0.01	0.01
18	7.6192 mg/l or 1.600 g/210L	42.931	0.01	0.01
19	8.0954 mg/l or 1.700 g/210L	45.455	0.01	0.01
20	8.5716 mg/l or 1.800 g/210L	47.979	0.01	0.01

=====

OPTICAL CALIBRATION ADJUSTMENT #3

P

Solution Stats Quadratic Fit Chan 1		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	0.000	-0.0002
0.040	0.040	0.0005
0.100	0.100	-0.0002
0.200	0.200	-0.0000
0.300	0.300	0.0000

Solution Stats Quadratic Fit Chan 2		
Act	Fit	Residual
g/210L	g/210L	g/210L
0.000	-0.000	0.0000
0.040	0.040	0.0001
0.100	0.100	-0.0002
0.200	0.200	0.0002
0.300	0.300	-0.0000

Sol Value = 0.080 g/210L ***
 Fit value = 0.3810 mg/l %%%
 Samples Taken = 4, Discarded = 1
 ***** CHANNEL 1
 Sample #1 = 3135.00
 Sample #2 = 3130.00
 Sample #3 = 3118.00
 Sample #4 = 3112.00
 Average Result = 3120.0000
 STD DEV = 9.1652
 REL STD DEV = 0.294

 ***** CHANNEL 2
 Sample #1 = 3423.00
 Sample #2 = 3428.00
 Sample #3 = 3431.00
 Sample #4 = 3407.00
 Average Result = 3422.0000
 STD DEV = 13.0767
 REL STD DEV = 0.382

 Dry Gas H2O Adjust Results *****
 Barometric Pressure = 1013
 3 um H2O Adjust (mg/l x 10,000) = 689
 9 um H2O Adjust (mg/l x 10,000) = 387
 ***** AUTO CAL PASS


Post stability Checks #1 (CALIBRATION ADJUSTMENT #2)



SUMNER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:23
Control Test	INT*	09:24
Air Blank	PUR**	09:25
Air Blank	PUR**	09:25

*Interferent Detect
**Purge Fail


Operator's Signature

SUMNER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:26
Control Test	INT*	09:27
Air Blank	AMB**	09:27
Air Blank	PUR***	09:28


*Interferent Detect
**Ambient Fail
***Purge Fail


Operator's Signature

SUMNER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:29
Control Test	INT*	09:30
Air Blank	AMB**	09:30
Air Blank	PUR***	09:31


*Interferent Detect
**Ambient Fail
***Purge Fail


Operator's Signature

SUMNER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:33
Control Test	INT*	09:34
Air Blank	PUR**	09:34
Air Blank	PUR**	09:35

*Interferent Detect
**Purge Fail


Operator's Signature

Post Stability Checks #2 (CALIBRATION ADJUSTMENT #3)

[Handwritten signature]

SUMNER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:36
Control Test	0.049	12:37
Air Blank	0.000	12:37
Control Test	0.049	12:38
Air Blank	0.000	12:39
Control Test	0.049	12:39
Air Blank	0.000	12:40
Control Test Stats		
Average	0.0490	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Handwritten signature]

Operator's Signature

SUMNER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:41
Control Test	0.079	12:41
Air Blank	0.000	12:42
Control Test	0.079	12:43
Air Blank	0.000	12:43
Control Test	0.078	12:44
Air Blank	0.000	12:45
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

[Handwritten signature]

Operator's Signature

SUMNER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:45
Control Test	0.200	12:46
Air Blank	0.000	12:47
Control Test	0.200	12:47
Air Blank	0.000	12:48
Control Test	0.200	12:49
Air Blank	0.000	12:49
Control Test Stats		
Average	0.2000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

[Handwritten signature]

Operator's Signature

SUMNER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007382
01/19/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	12:54
Control Test	0.080	12:54
Air Blank	0.000	12:55
Control Test	0.080	12:55
Air Blank	0.000	12:56
Control Test	0.080	12:56
Air Blank	0.000	12:56
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

[Handwritten signature]

Operator's Signature

NOTES

Tech Review: Instrument had a DGS measurement that was outside of acceptable range (see Stability Checks). Calibration adjustment (i.e., #1) was performed on 1/18. Instrument failed a diagnostic check after the calibration adjustment was performed (i.e., analytical). It is suspected that adjustment solution order may have been erroneously arranged. A repeated calibration adjustment (i.e., #2) was performed on 1/19. A failed diagnostic check occurred prior to the adjustment (i.e., analytical & DSP), and post stability check resulted exception messages (see Post Stability Checks #1). A final calibration adjustment (i.e., #3) was performed on 1/19. Post Stability Checks (see Post Stability Checks #2) resulted in nominal values. A Department Inspection was completed on 1/19. PN 1/19/2023