



INSTRUMENT PROCESSING SHEET

Agency Baker County SOS/N 80-007439Florida Department of
Law EnforcementDate In 12/12/2023 DI Completion Date 12/12/2023 Ship P/U H/D CMI EE

Intake By BS _____ <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____	Quality Checks By BS _____ Date <u>12/12/2023</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>183</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.242</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28427</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td rowspan="2">0.050</td> <td rowspan="2">MP6291</td> <td>202303K</td> </tr> <tr> <td>3/29/2025</td> </tr> <tr> <td rowspan="2">0.080</td> <td rowspan="2">MP6292</td> <td>202303L</td> </tr> <tr> <td>3/29/2025</td> </tr> <tr> <td rowspan="2">0.200</td> <td rowspan="2">MP6293</td> <td>202304C</td> </tr> <tr> <td>4/5/2025</td> </tr> <tr> <td rowspan="2">0.080 DGS</td> <td rowspan="2">N/A</td> <td>AG325603</td> </tr> <tr> <td>9/13/2025</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202303K	3/29/2025	0.080	MP6292	202303L	3/29/2025	0.200	MP6293	202304C	4/5/2025	0.080 DGS	N/A	AG325603	9/13/2025	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____																																								
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP6291	202303K																																																											
		3/29/2025																																																											
0.080	MP6292	202303L																																																											
		3/29/2025																																																											
0.200	MP6293	202304C																																																											
		4/5/2025																																																											
0.080 DGS	N/A	AG325603																																																											
		9/13/2025																																																											
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> Notes/Suggested Service: _____ _____ _____ _____ _____ _____	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By BS _____ Barometric Pressure ID# <u>28427</u> Gauge <u>1026</u> Instrument <u>1028</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP6289</td> </tr> <tr> <td>Interferent</td> <td>MP6290</td> </tr> <tr> <td>0.050</td> <td>MP6291</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> </tr> </tbody> </table> Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Israel Soto <small>Digitally signed by Israel Soto Date: 2023.12.12 13:31:33 -05'00'</small> Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2023.12.12 13:42:30 -05'00'</small> Tech Review / Date _____ Admin Review / Date _____	Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293
Simulator	Serial #	Lot #	Expiration																																																										
0.000		N/A	N/A																																																										
0.040																																																													
0.100																																																													
0.200																																																													
0.300																																																													
0.080 DGS	N/A																																																												
Simulator	Serial #	Lot #	Expiration																																																										
0.050																																																													
0.080																																																													
0.200																																																													
0.080 DGS	N/A																																																												
Simulator	Serial Number																																																												
0.000	MP6289																																																												
Interferent	MP6290																																																												
0.050	MP6291																																																												
0.080	MP6292																																																												
0.200	MP6293																																																												

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 12:39

Date of Inspection: 12/12/2023

Serial Number: 80-007439
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG325603 Exp: 09/13/2025
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.051	0.080	0.199	0.080
0.000	0.051	0.080	0.199	0.080

Standard Deviations	0.0004	0.0004	0.0004	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Benjamin W Siddoway BENJAMIN W SIDDOWNAY
Signature and Printed Name

12/12/2023
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 10:00

Date of Inspection: 12/12/2023

Serial Number: 80-007439
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____

Remarks:
COMPLIANCE NOT DETERMINED, BYPASS TO OPERATE

N/A compliance not determined

The above instrument ~~complies (X)~~ does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Benjamin W. Siddoway BENJAMIN W SIDDOWAY
Signature and Printed Name

12/12/2023
Date

Stability Checks

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
12/12/2023
Software: 8100.27

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
12/12/2023
Software: 8100.27


BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
12/12/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:20
Control Test	0.050	10:21
Air Blank	0.000	10:21
Control Test	0.050	10:22
Air Blank	0.000	10:22
Control Test	0.050	10:23
Air Blank	0.000	10:24
Control Test Stats		
Average	0.0500	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	10:25
Control Test	0.080	10:26
Air Blank	0.000	10:26
Control Test	0.080	10:27
Air Blank	0.000	10:28
Control Test	0.080	10:28
Air Blank	0.000	10:29
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Test	g/210L	Time
Air Blank	0.000	10:31
Control Test	0.198	10:32
Air Blank	0.000	10:32
Control Test	0.198	10:33
Air Blank	0.000	10:33
Control Test	0.198	10:34
Air Blank	0.000	10:35
Control Test Stats		
Average	0.1980	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature



Operator's Signature


Operator's Signature

BAKER COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007439
12/12/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:38
Control Test	0.079	10:39
Air Blank	0.000	10:39
Control Test	0.079	10:40
Air Blank	0.000	10:40
Control Test	0.079	10:40
Air Blank	0.000	10:41
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS


Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007439, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007439</u>	UNCERTAINTY* ±	
Owning Agency:	<u>BAKER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>12/12/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:39</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Benjamin Siddoway
Digitally signed by Benjamin Siddoway
Date: 2023.12.12 13:26:20 -05'00'

Date 12/12/2023
BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Baker County Sheriff's OfficeS/N 80-007439Florida Department of
Law EnforcementDate In 1/10/2023DI Completion Date 1/10/2023 Ship P/U H/D CMI EE

Intake By PN _____ <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____	Quality Checks By PN _____ Date <u>1/10/2023</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>191</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP 105</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.164</u> (.156 - .190) 53 mm <u>.234</u> (.228 - .278) 103 mm <u>.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td rowspan="2">MP5088</td> <td>202201C</td> </tr> <tr> <td></td> <td>01/11/2024</td> </tr> <tr> <td>0.080</td> <td rowspan="2">MP5089</td> <td>202201D</td> </tr> <tr> <td></td> <td>01/18/2024</td> </tr> <tr> <td>0.200</td> <td rowspan="2">MP5084</td> <td>202201E</td> </tr> <tr> <td></td> <td>01/18/2024</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG229803</td> </tr> <tr> <td></td> <td></td> <td>10/25/2024</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202201C		01/11/2024	0.080	MP5089	202201D		01/18/2024	0.200	MP5084	202201E		01/18/2024	0.080 DGS	N/A	AG229803			10/25/2024	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____																																					
Simulator	Serial #	Lot #/Exp																																																													
0.050	MP5088	202201C																																																													
		01/11/2024																																																													
0.080	MP5089	202201D																																																													
		01/18/2024																																																													
0.200	MP5084	202201E																																																													
		01/18/2024																																																													
0.080 DGS	N/A	AG229803																																																													
		10/25/2024																																																													
Calibration Adjustment By _____ Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td></td> <td>N/A</td> <td>N/A</td> </tr> <tr> <td>0.040</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.100</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.300</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #</th> <th>Expiration</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.200</td> <td></td> <td></td> <td></td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td></td> <td></td> </tr> </tbody> </table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Department Inspection By PN _____ Barometric Pressure ID# <u>28421</u> Gauge <u>1019</u> Instrument <u>1019</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>Simulator</th> <th>Serial Number</th> </tr> </thead> <tbody> <tr> <td>0.000</td> <td>MP5086</td> </tr> <tr> <td>Interferent</td> <td>MP5087</td> </tr> <tr> <td>0.050</td> <td>MP5088</td> </tr> <tr> <td>0.080</td> <td>MP5089</td> </tr> <tr> <td>0.200</td> <td>MP5084</td> </tr> </tbody> </table>		Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5084
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000	MP5086																																																														
Interferent	MP5087																																																														
0.050	MP5088																																																														
0.080	MP5089																																																														
0.200	MP5084																																																														
Notes/Suggested Service: _____ _____ _____ _____ _____ _____		Attachments <input checked="" type="checkbox"/> Form 41 <input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____																																																													
		<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="text-align: right;"> <small>2023.01.11 0 16:07:22 -05'00'</small> </div>																																																													
		Israel Soto <small>Digitally signed by Israel Soto Date: 2023.01.10 15:00:06 -05'00'</small> Tech Review / Date _____	Admin Review / Date _____																																																												

Type of Test	Serial Number	Agency	Date	Performed By
Stability Checks	80-007439	Baker County Sheriff's Office	1/10/2023	PN

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083

<p>BAKER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/10/2023 Software: 8100.27</p>	<p>BAKER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/10/2023 Software: 8100.27</p>	<p>BAKER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/10/2023 Software: 8100.27</p>	<p>BAKER COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 01/10/2023 Software: 8100.27</p>
---	---	---	---

<p>Test</p> <p>Air Blank 0.000 09:24</p> <p>Control Test 0.049 09:25</p> <p>Air Blank 0.000 09:25</p> <p>Control Test 0.049 09:26</p> <p>Air Blank 0.000 09:27</p> <p>Control Test 0.049 09:27</p> <p>Air Blank 0.000 09:28</p> <p>Control Test Stats</p> <p>Average 0.0490</p> <p>Std Dev 0.0000</p> <p>Rel Std Dev(%) 0.0000</p>	<p>Test</p> <p>Air Blank 0.000 09:31</p> <p>Control Test 0.078 09:32</p> <p>Air Blank 0.000 09:33</p> <p>Control Test 0.078 09:33</p> <p>Air Blank 0.000 09:34</p> <p>Control Test 0.079 09:35</p> <p>Air Blank 0.000 09:35</p> <p>Control Test Stats</p> <p>Average 0.0783</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7370</p>	<p>Test</p> <p>Air Blank 0.000 09:37</p> <p>Control Test 0.199 09:38</p> <p>Air Blank 0.000 09:39</p> <p>Control Test 0.198 09:39</p> <p>Air Blank 0.000 09:40</p> <p>Control Test 0.199 09:41</p> <p>Air Blank 0.000 09:41</p> <p>Control Test Stats</p> <p>Average 0.1987</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2906</p>	<p>Test</p> <p>Air Blank 0.000 09:44</p> <p>Control Test 0.079 09:44</p> <p>Air Blank 0.000 09:45</p> <p>Control Test 0.079 09:45</p> <p>Air Blank 0.000 09:45</p> <p>Control Test 0.080 09:46</p> <p>Air Blank 0.000 09:47</p> <p>Control Test Stats</p> <p>Average 0.0793</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7277</p>
--	--	--	--

Operator's Signature _____

Operator's Signature _____

Operator's Signature _____

Operator's Signature _____

Comments: N/A

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: BAKER COUNTY SO
Time of Inspection: 11:51

Date of Inspection: 01/10/2023

Serial Number: 80-007439
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.048	0.078	0.199	0.079
0.000	0.049	0.078	0.200	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.078	0.199	0.079
0.000	0.049	0.079	0.200	0.079
0.000	0.049	0.079	0.199	0.079
0.000	0.049	0.079	0.199	0.079
Standard Deviations	0.0004	0.0005	0.0004	0.0000

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


PHIL NICODEMO

 Signature and Printed Name

01/10/2023
 Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007439, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007439</u>	UNCERTAINTY* ±	
Owning Agency:	<u>BAKER COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>01/10/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>11:51</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.
Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Phil Nicodemo
Nicodemo
Date: 2023.01.10 14:49:03
-05'00'

Digitally signed by Phil

01/10/2023

Date

PHIL NICODEMO,

Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality