



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-007459Florida Department of
Law EnforcementDate In 9/13/2023DI Completion Date 9/26/2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By BS	Quality Checks	By BS	Date <u>9/26/2023</u>	Flow Calibration	By	Date																																							
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☐ Handle ☒ Keyboard ☐ Dry Gas Shelf ☒ Feet ☐ Breath Tube ☒ Ports ☐ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>211</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.492</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																									
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K 3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 4/5/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5 4/5/2025</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 3/29/2025	0.080	MP5089	202303L 3/29/2025	0.200	MP5090	202304C 4/5/2025	0.080 DGS	N/A	06723080A5 4/5/2025																													
Simulator	Serial #	Lot #/Exp																																												
0.050	MP5088	202303K 3/29/2025																																												
0.080	MP5089	202303L 3/29/2025																																												
0.200	MP5090	202304C 4/5/2025																																												
0.080 DGS	N/A	06723080A5 4/5/2025																																												
Calibration Adjustment		By _____		Department Inspection			By BS _____																																							
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28662</u> Gauge <u>1015</u> Instrument <u>1013</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A					<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090
Simulator	Serial #	Lot #	Expiration																																											
0.000		N/A	N/A																																											
0.040																																														
0.100																																														
0.200																																														
0.300																																														
0.080 DGS	N/A																																													
Simulator	Serial Number																																													
0.000	MP5086																																													
Interferent	MP5087																																													
0.050	MP5088																																													
0.080	MP5089																																													
0.200	MP5090																																													
☐ Post Calibration Adjustment Stability Checks				Attachments																																										
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A					<table border="1"><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☒ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☐ Other _____</td></tr></tbody></table>			☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☒ Form 40	☐ Calibration Adjustment	☐ Other _____												
Simulator	Serial #	Lot #	Expiration																																											
0.050																																														
0.080																																														
0.200																																														
0.080 DGS	N/A																																													
☒ Form 41	☐ Post-Stability Checks																																													
☒ Stability Checks	☐ Flow Calibration																																													
☒ Calibration Certificate	☒ Form 40																																													
☐ Calibration Adjustment	☐ Other _____																																													
Notes/Suggested Service: _____ _____ _____ _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																										
		Israel Soto Digitally signed by Israel Soto Date: 2023.09.27 09:09:40 +04'00'		Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.09.27 10:17:56 -04'00'																																										
		Tech Review / Date		Admin Review / Date																																										

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 15:37

Date of Inspection: 09/26/2023

Serial Number: 80-007459

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.050	0.079	0.197	0.081
0.000	0.050	0.079	0.197	0.081
0.000	0.049	0.079	0.196	0.080
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.197	0.079
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.080	0.197	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.197	0.077
0.000	0.050	0.080	0.197	0.075


Standard Deviations	0.0005	0.0004	0.0003	0.0018
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0007 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BENJAMIN W SIDDOWAY
Signature and Printed Name

09/26/2023
Date

Stability Checks

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
09/26/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:13
Control Test	0.050	11:13
Air Blank	0.000	11:14
Control Test	0.049	11:15
Air Blank	0.000	11:15
Control Test	0.049	11:16
Air Blank	0.000	11:16
Control Test Stats		
Average	0.0493	
Std Dev	0.0006	
Rel Std Dev(%)	1.1703	

Benjamin Siddle
Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
09/26/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:22
Control Test	0.079	11:23
Air Blank	0.000	11:23
Control Test	0.078	11:24
Air Blank	0.000	11:24
Control Test	0.079	11:25
Air Blank	0.000	11:26
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Benjamin Siddle
Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
09/26/2023 23:59 04/25/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:47
Control Test	0.197	10:47
Air Blank	0.000	10:48
Control Test	0.197	10:49
Air Blank	0.000	10:49
Control Test	0.197	10:50
Air Blank	0.000	10:50
Control Test Stats		
Average	0.1970	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Benjamin Siddle
Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
09/26/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:58
Control Test	0.081	10:58
Air Blank	0.000	10:59
Control Test	0.081	10:59
Air Blank	0.000	10:59
Control Test	0.081	11:00
Air Blank	0.000	11:00
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS

Benjamin Siddle
Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007459, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007459</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/26/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>15:37</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Benjamin
Siddoway
Digitally signed by Benjamin
Siddoway
Date: 2023.09.26 16:06:03
+04'00'

09/26/2023

Date

BENJAMIN W SIDDOWAY,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 10:28

Date of Inspection: 09/26/2023

Serial Number: 80-007459

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

COMPLIANCE NOT DETERMINED, BYPASS TO OPERATE

compliance not determined

The above instrument ~~complies (x) does not comply ()~~ with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Benjamin W Siddoway

Signature and Printed Name

BENJAMIN W SIDDOWAY

09/26/2023
Date



INSTRUMENT PROCESSING SHEET

Agency Florida Highway PatrolS/N 80-007459Florida Department of
Law EnforcementDate In 02-28-2023 DI Completion Date 03-03-2023☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By IS	Quality Checks	By IS	Date	03-02-2023	Flow Calibration	By	Date																																								
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: Note included with instrument indicating instrument received Int Printer Test Fail during startup Diagnostic Check.		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>209</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP-105</u> 32 mm <u>0.160</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.515</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>30793</u> ☒ Stability Checks				☐ Flow Column # _____ ☐ 5L/min – 17mm ☐ 15L/min – 53mm ☐ 30L/min – 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																										
		<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C 01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D 01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E 01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803 10-25-2024</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202201C 01-11-2024	0.080	MP6292	202201D 01-18-2024	0.200	MP6293	202201E 01-18-2024	0.080 DGS	N/A	AG229803 10-25-2024	<table border="1"><thead><tr><th colspan="2">Maintenance</th><th>By IS</th></tr></thead><tbody><tr><td colspan="2">☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Cleared internal printer jam.</u> <u>Instrument passed startup Diagnostic Check and entered Ready Mode.</u> <u>03-02-2023</u></td><td></td></tr></tbody></table>			Maintenance		By IS	☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Cleared internal printer jam.</u> <u>Instrument passed startup Diagnostic Check and entered Ready Mode.</u> <u>03-02-2023</u>																						
Simulator	Serial #	Lot #/Exp																																														
0.050	MP6291	202201C 01-11-2024																																														
0.080	MP6292	202201D 01-18-2024																																														
0.200	MP6293	202201E 01-18-2024																																														
0.080 DGS	N/A	AG229803 10-25-2024																																														
Maintenance		By IS																																														
☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☒ Other <u>Cleared internal printer jam.</u> <u>Instrument passed startup Diagnostic Check and entered Ready Mode.</u> <u>03-02-2023</u>																																																
Calibration Adjustment		By IS			Department Inspection			By IS																																								
Barometric Pressure Gauge <u>1002</u> ID # <u>28427</u>					Barometric Pressure ID# <u>30793</u> Gauge <u>1004</u> Instrument <u>1003</u> Mouth Alcohol Solution Lot # <u>2022-A</u> Acetone Stock Solution Lot # <u>2022-B</u>																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP6294</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP6295</td><td>21410</td><td>09-30-2023</td></tr><tr><td>0.100</td><td>MP6296</td><td>22310</td><td>08-11-2024</td></tr><tr><td>0.200</td><td>MP6297</td><td>22050</td><td>02-07-2024</td></tr><tr><td>0.300</td><td>MP6298</td><td>22220</td><td>06-15-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>08121080A1</td><td>05-05-2023</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.000	MP6294	N/A	N/A	0.040	MP6295	21410	09-30-2023	0.100	MP6296	22310	08-11-2024	0.200	MP6297	22050	02-07-2024	0.300	MP6298	22220	06-15-2024	0.080 DGS	N/A	08121080A1	05-05-2023				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6289</td></tr><tr><td>Interferent</td><td>MP6290</td></tr><tr><td>0.050</td><td>MP6291</td></tr><tr><td>0.080</td><td>MP6292</td></tr><tr><td>0.200</td><td>MP6293</td></tr></tbody></table>			Simulator	Serial Number	0.000	MP6289	Interferent	MP6290	0.050	MP6291	0.080	MP6292	0.200	MP6293	
Simulator	Serial #	Lot #	Expiration																																													
0.000	MP6294	N/A	N/A																																													
0.040	MP6295	21410	09-30-2023																																													
0.100	MP6296	22310	08-11-2024																																													
0.200	MP6297	22050	02-07-2024																																													
0.300	MP6298	22220	06-15-2024																																													
0.080 DGS	N/A	08121080A1	05-05-2023																																													
Simulator	Serial Number																																															
0.000	MP6289																																															
Interferent	MP6290																																															
0.050	MP6291																																															
0.080	MP6292																																															
0.200	MP6293																																															
☒ Post Calibration Adjustment Stability Checks					Attachments																																											
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202201C</td><td>01-11-2024</td></tr><tr><td>0.080</td><td>MP6292</td><td>202201D</td><td>01-18-2024</td></tr><tr><td>0.200</td><td>MP6293</td><td>202201E</td><td>01-18-2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG229803</td><td>10-25-2024</td></tr></tbody></table>		Simulator	Serial #	Lot #	Expiration	0.050	MP6291	202201C	01-11-2024	0.080	MP6292	202201D	01-18-2024	0.200	MP6293	202201E	01-18-2024	0.080 DGS	N/A	AG229803	10-25-2024				☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☒ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☒ Form 40 ☐ Other _____																							
Simulator	Serial #	Lot #	Expiration																																													
0.050	MP6291	202201C	01-11-2024																																													
0.080	MP6292	202201D	01-18-2024																																													
0.200	MP6293	202201E	01-18-2024																																													
0.080 DGS	N/A	AG229803	10-25-2024																																													
Notes/Suggested Service: <u>Wet 0.080 g/210L Stability</u> <u>Check values outside nominal range. Optical Bench</u> <u>Calibration Adjustment performed. IS</u>					☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																											
					Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2023.03.03 12:42:21 -05'00'			Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.03.07 10:16:13 -05'00'																																								
					Tech Review / Date			Admin Review / Date																																								

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 11:07

Date of Inspection: 03/03/2023

Serial Number: 80-007459

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG229803 Exp: 10/25/2024
0.000	0.048	0.077	0.199	0.080
0.000	0.048	0.077	0.200	0.080
0.000	0.048	0.078	0.199	0.079
0.000	0.048	0.077	0.199	0.079
0.000	0.048	0.077	0.199	0.080
0.000	0.048	0.077	0.199	0.080
0.000	0.048	0.077	0.199	0.080
0.000	0.049	0.077	0.199	0.080
0.000	0.048	0.078	0.199	0.080
0.000	0.048	0.078	0.200	0.080

Standard Deviations	0.0003	0.0004	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

03/03/2023
Date

stability checks

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
03/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:24
Control Test	0.047	13:24
Air Blank	0.000	13:25
Control Test	0.047	13:26
Air Blank	0.000	13:26
Control Test	0.047	13:27
Air Blank	0.000	13:27
Control Test Stats		
Average	0.0470	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
03/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:29
Control Test	0.076	13:29
Air Blank	0.000	13:30
Control Test	0.076	13:31
Air Blank	0.000	13:31
Control Test	0.077	13:32
Air Blank	0.000	13:32
Control Test Stats		
Average	0.0763	
Std Dev	0.0006	
Rel Std Dev(%)	0.7564	

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
03/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:33
Control Test	0.196	13:34
Air Blank	0.000	13:35
Control Test	0.195	13:35
Air Blank	0.000	13:36
Control Test	0.195	13:37
Air Blank	0.000	13:37
Control Test Stats		
Average	0.1953	
Std Dev	0.0006	
Rel Std Dev(%)	0.2956	

wet

Operator's Signature

Operator's Signature

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
03/02/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:39
Control Test	0.078	13:39
Air Blank	0.000	13:39
Control Test	0.079	13:40
Air Blank	0.000	13:40
Control Test	0.079	13:41
Air Blank	0.000	13:41
Control Test Stats		
Average	0.0787	
Std Dev	0.0006	
Rel Std Dev(%)	0.7339	

Dry

Operator's Signature



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007459, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007459</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FHP</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/03/2023</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:07</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

03/03/2023

Date

Israel Soto

ISRAEL SOTO,
Department Inspector

Digitally signed by Israel

Soto

Date: 2023.03.03

11:15:49 -05'00'

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
03/03/2023 07:46:38

Optical Bench
Calibration Adjustment


Auto Calibration

pg 1 of 2

	<<<<<	3um	>>>>>	<<<<<	9um	>>>>>

Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	0.0880	(-0.0090)		0.0910	(0.0040)	
Sample #2	0.1010	(-0.0060)		0.1010	(-0.0140)	
Sample #3	0.0850	(0.0260)		0.1040	(-0.0040)	
Sample #4	0.1110	(0.0160)		0.1070	(0.0040)	
Avg % Abs	0.0990	(0.0120)		0.1040	(-0.0047)	
STD DEV	0.0131	(0.0164)		0.0030	(0.0090)	
REL STD DEV	13.247	(136.423)		2.885	(193.254)	

Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	0.8090	(-0.0150)		1.4560	(-0.0090)	
Sample #2	0.8040	(0.0060)		1.4310	(0.0000)	
Sample #3	0.7930	(0.0220)		1.4550	(0.0000)	
Sample #4	0.8480	(0.0100)		1.4650	(-0.0020)	
Avg % Abs	0.8150	(0.0127)		1.4503	(-0.0007)	
STD DEV	0.0291	(0.0083)		0.0175	(0.0012)	
REL STD DEV	3.571	(65.737)		1.205	(173.205)	

Solution = 0.100 g/210L or 0.4762 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	1.9530	(-0.0130)		3.4500	(0.0020)	
Sample #2	1.8850	(0.0390)		3.4210	(0.0310)	
Sample #3	1.8980	(0.0500)		3.4310	(0.0230)	
Sample #4	1.9100	(0.0620)		3.4300	(0.0300)	
Avg % Abs	1.8977	(0.0503)		3.4273	(0.0280)	
STD DEV	0.0125	(0.0115)		0.0055	(0.0044)	
REL STD DEV	0.659	(22.855)		0.161	(15.567)	

Solution = 0.200 g/210L or 0.9524 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	3.7020	(-0.0180)		6.6040	(0.0010)	
Sample #2	3.6760	(0.0240)		6.5750	(0.0320)	
Sample #3	3.7020	(0.0240)		6.5930	(0.0350)	
Sample #4	3.6600	(0.0560)		6.5610	(0.0440)	
Avg % Abs	3.6793	(0.0347)		6.5763	(0.0370)	
STD DEV	0.0212	(0.0185)		0.0160	(0.0062)	
REL STD DEV	0.576	(53.294)		0.244	(16.878)	

Solution = 0.300 g/210L or 1.4286 mg/l, Samples = 4, Discarded = 1						
Sample	% Abs	(% Abs Ref)		% Abs	(% Abs Ref)	
Sample #1	5.4170	(-0.0110)		9.5660	(-0.0030)	
Sample #2	5.3660	(0.0320)		9.5470	(0.0300)	
Sample #3	5.3740	(0.0480)		9.5530	(0.0320)	
Sample #4	5.3900	(0.0320)		9.5660	(0.0290)	
Avg % Abs	5.3767	(0.0373)		9.5553	(0.0303)	
STD DEV	0.0122	(0.0092)		0.0097	(0.0015)	
REL STD DEV	0.227	(24.744)		0.102	(5.036)	

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
03/03/2023 07:46:38

Optical Bench
Calibration Adjustment

OK

Auto Calibration

pg 2 of 2

<<<< 3um >>>>

Zero Order Coef -234.09
First Order Coef 2580.19
Second Order Coef 21.95

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0005
0.040	0.040	0.0004
0.100	0.100	0.0004
0.200	0.201	-0.0007
0.300	0.300	0.0003

<<<< 9um >>>>

-138.07
1384.16
13.08

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0001
0.040	0.040	0.0002
0.100	0.100	0.0000
0.200	0.200	-0.0001
0.300	0.300	0.0001

<<<< 3um >>>> <<<< 9um >>>>

Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1 3134.00 3399.00
Sample #2 3195.00 3403.00
Sample #3 3188.00 3411.00
Sample #4 3267.00 3402.00
Avg 3216.6667 3405.3333
STD DEV 43.7302 4.9329
REL STD DEV 1.359 0.145
H2O adjust (mg/l*10k) 593 404

Barometric Pressure = 1002

*****CALIBRATION SUCCESSFUL*****

Post Stability Checks

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
03/03/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:48
Control Test	0.048	08:49
Air Blank	0.000	08:49
Control Test	0.048	08:50
Air Blank	0.000	08:51
Control Test	0.048	08:51
Air Blank	0.000	08:52
Control Test Stats		
Average	0.0480	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
03/03/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:53
Control Test	0.077	08:54
Air Blank	0.000	08:54
Control Test	0.077	08:55
Air Blank	0.000	08:56
Control Test	0.078	08:56
Air Blank	0.000	08:57
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
03/03/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:59
Control Test	0.199	09:00
Air Blank	0.000	09:00
Control Test	0.198	09:01
Air Blank	0.000	09:02
Control Test	0.199	09:02
Air Blank	0.000	09:03
Control Test Stats		
Average	0.1987	
Std Dev	0.0006	
Rel Std Dev(%)	0.2906	

Operator's Signature

FHP
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007459
03/03/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:04
Control Test	0.081	09:04
Air Blank	0.000	09:05
Control Test	0.080	09:05
Air Blank	0.000	09:06
Control Test	0.080	09:06
Air Blank	0.000	09:07
Control Test Stats		
Average	0.0803	
Std Dev	0.0006	
Rel Std Dev(%)	0.7187	

Operator's Signature

wet

Dry

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FHP

Time of Inspection: 11:38

Date of Inspection: 03/02/2023

Serial Number: 80-007459

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

BYPASSED AI TO OPERATE INSTRUMENT

N/A compliance not determined

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

03/02/2023

Date