



Agency Polk County Sheriff's Office

S/N 80-007503

Date In 07/11/2023 DI Completion Date 07-14-2023

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

FDLE/ATP Form 48 January 2022
Issuing Authority: Alcohol Testing Program

Page 1 of 1

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK CSO

Time of Inspection: 10:18

Date of Inspection: 07/14/2023

Serial Number: 80-007503

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:06723080A5 Exp: 04/05/2025
0.000	0.049	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.080
0.000	0.049	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.079
0.000	0.049	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.080
0.000	0.050	0.080	0.201	0.079
0.000	0.050	0.080	0.201	0.079

Standard Deviations	0.0005	0.0000	0.0000	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0002 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Israel Soto

ISRAEL SOTO

Signature and Printed Name

07/14/2023
Date

stability checks

POLK CSO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007503
07/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:23
Control Test	0.050	08:23
Air Blank	0.000	08:24
Control Test	0.049	08:25
Air Blank	0.000	08:25
Control Test	0.050	08:26
Air Blank	0.000	08:26
Control Test Stats		
Average	0.0497	
Std Dev	0.0006	
Rel Std Dev(%)	1.1625	

Operator's Signature

POLK CSO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007503
07/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:27
Control Test	0.080	08:28
Air Blank	0.000	08:29
Control Test	0.080	08:29
Air Blank	0.000	08:30
Control Test	0.080	08:31
Air Blank	0.000	08:31
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

POLK CSO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007503
07/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:32
Control Test	0.201	08:33
Air Blank	0.000	08:33
Control Test	0.201	08:34
Air Blank	0.000	08:34
Control Test	0.201	08:35
Air Blank	0.000	08:36
Control Test Stats		
Average	0.2010	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

POLK CSO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-007503
07/14/2023
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	08:37
Control Test	0.079	08:37
Air Blank	0.000	08:38
Control Test	0.079	08:38
Air Blank	0.000	08:39
Control Test	0.079	08:39
Air Blank	0.000	08:40
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

wet

Dry



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007503, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007503</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>POLK CSO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/14/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:18</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Israel Soto
Soto
Digitally signed by Israel Soto
Date: 2023.07.14 10:50:59 -04'00'

07/14/2023

Date

ISRAEL SOTO,
Department Inspector

FDLE/ATP Form 69 March 2022

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Polk CSOS/N 80-007503Florida Department of
Law EnforcementDate In 12/12/2022 DI Completion Date 02/21/2023☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By TDG	Quality Checks	By TDG	Date <u>02/21/2023</u>	Flow Calibration	By	Date																														
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: <u>Instrument retained after</u> <u>Dec 2022 Department Inspection</u> <u>with permission from the Agency</u> <u>Inspector.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>190</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.230</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>68639</u> ☒ Stability Checks			Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																
				<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5092</td><td>202201C 01/11/2024</td></tr><tr><td>0.080</td><td>MP5093</td><td>202201D 01/18/2024</td></tr><tr><td>0.200</td><td>MP5094</td><td>202201E 01/18/2024</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG223802 08/26/2024</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP5092	202201C 01/11/2024	0.080	MP5093	202201D 01/18/2024	0.200	MP5094	202201E 01/18/2024	0.080 DGS	N/A	AG223802 08/26/2024	<table border="1"><thead><tr><th>Maintenance</th><th>By</th></tr></thead><tbody><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr></tbody></table>			Maintenance	By	☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____						
Simulator	Serial #	Lot #/Exp																																			
0.050	MP5092	202201C 01/11/2024																																			
0.080	MP5093	202201D 01/18/2024																																			
0.200	MP5094	202201E 01/18/2024																																			
0.080 DGS	N/A	AG223802 08/26/2024																																			
Maintenance	By																																				
☐ Battery Replacement																																					
☐ Dry Gas Regulator Replacement																																					
☐ Breath Tube Replacement																																					
☐ Other _____																																					
Calibration Adjustment				Department Inspection																																	
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>28199</u>																																	
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1019</u> Instrument <u>1017</u> Mouth Alcohol Solution Lot # <u>2021-D</u> Acetone Stock Solution Lot # <u>2021-C</u>					
Simulator	Serial #	Lot #	Expiration																																		
0.000		N/A	N/A																																		
0.040																																					
0.100																																					
0.200																																					
0.300																																					
0.080 DGS	N/A																																				
☐ Post Calibration Adjustment Stability Checks				<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5095</td></tr><tr><td>Interferent</td><td>MP5097</td></tr><tr><td>0.050</td><td>MP5092</td></tr><tr><td>0.080</td><td>MP5093</td></tr><tr><td>0.200</td><td>MP5094</td></tr></tbody></table>				Simulator	Serial Number	0.000	MP5095	Interferent	MP5097	0.050	MP5092	0.080	MP5093	0.200	MP5094																		
Simulator	Serial Number																																				
0.000	MP5095																																				
Interferent	MP5097																																				
0.050	MP5092																																				
0.080	MP5093																																				
0.200	MP5094																																				
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th colspan="2">Attachments</th></tr></thead><tbody><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☒ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☒ Other <u>Form 51</u></td></tr></tbody></table>				Attachments		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☒ Form 40	☐ Calibration Adjustment	☒ Other <u>Form 51</u>
Simulator	Serial #	Lot #	Expiration																																		
0.050																																					
0.080																																					
0.200																																					
0.080 DGS	N/A																																				
Attachments																																					
☒ Form 41	☐ Post-Stability Checks																																				
☒ Stability Checks	☐ Flow Calibration																																				
☒ Calibration Certificate	☒ Form 40																																				
☐ Calibration Adjustment	☒ Other <u>Form 51</u>																																				
Notes/Suggested Service: <u>Checked breath tube screen</u> <u>and replaced o-rings on 1/9. Battery clip broke during</u> <u>attempted battery replacement on 2/6, and the battery</u> <u>was subsequently unable to be replaced. Conducted</u> <u>Department Inspection on 2/21. Will send to CMI for</u> <u>repair. (TDG)</u>				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☐ Return to/Place into Evidentiary Use ☒ Remain Out of Evidentiary Use ☐ Conduct an Agency Inspection Before Evidentiary Use																																	
Admin Review: <u>Updated billing address, added remark to</u> <u>Reason for Return, and added X to estimate option box and</u> <u>items returned box on Form 51. Indicated instrument should remain out of evidential use on Form 48. (TDG 02/27/2023)</u>				Israel Soto Digitally signed by Israel Soto Date: 2023.02.23 06:56:39 +05'00' Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2023.02.27 13:25:55 -05'00'																																	
Tech Review / Date				Admin Review / Date																																	

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Nathan Lucas on 2/6/2023

TDG 02/27/2023
Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____
Instrument Model: Intoxilyzer 8000 **Serial Number:** 80-007503

Bill To Address:

Polk CSO Florida Dept. of Law Enforcement

Attn: Nathan Lucas Attn: Shayla Platt

TDG 02/27/2023

Ship to Address:

Florida Department of Law Enforcement

Alcohol Testing Program (Tallahassee)

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

The battery clip is broken and needs to be replaced. The battery also needs to be replaced.

FDLE will cover the cost of replacing the battery clip and battery. For anything else, the agency should be contacted for approval/payment. (TDG 02/27/2023)

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$ _____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Nathan Lucas

Phone #: 863-398-3748

Email: NLucas@polksheiff.org

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us

TDG 02/27/2023

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK CSO
Time of Inspection: 09:24

Date of Inspection: 02/21/2023

Serial Number: 80-007503
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____

Number of Simulators Used: _____

Remarks:

AI NOT CONDUCTED. BYPASSED TO OPERATE INSTRUMENT

Not determined

NG
02/21/2023

The above instrument complies (☒ X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

02/21/2023
Date

Type of Test	Serial Number	Agency	Date	Performed By
Stabilities	80-00 7503	POLK CSO	02/21/2023	TDG MG

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053 ✓	0.077 to 0.083 ✓	0.194 to 0.206 ✓	0.077 to 0.083 ✓ ≤0.003 of Wet ✓																																																																																																																																																
POLK CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007503 02/21/2023 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:07</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:08</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:09</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:09</td></tr><tr><td>Control Test</td><td>0.049</td><td>10:10</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:10</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0490</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> Operator's Signature <i>MG</i>	Test	g/210L	Time	Air Blank	0.000	10:07	Control Test	0.049	10:07	Air Blank	0.000	10:08	Control Test	0.049	10:09	Air Blank	0.000	10:09	Control Test	0.049	10:10	Air Blank	0.000	10:10	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		POLK CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007503 02/21/2023 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:13</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:14</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:15</td></tr><tr><td>Control Test</td><td>0.078</td><td>10:15</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:16</td></tr><tr><td>Control Test</td><td>0.079</td><td>10:17</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:17</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></tbody></table> Operator's Signature <i>MG</i>	Test	g/210L	Time	Air Blank	0.000	10:13	Control Test	0.078	10:14	Air Blank	0.000	10:15	Control Test	0.078	10:15	Air Blank	0.000	10:16	Control Test	0.079	10:17	Air Blank	0.000	10:17	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		POLK CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007503 02/21/2023 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr><tr><td>Control Test</td><td>0.197</td><td>10:21</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:22</td></tr><tr><td>Control Test</td><td>0.197</td><td>10:23</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr><tr><td>Control Test</td><td>0.196</td><td>10:24</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1967</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2936</td><td></td></tr></tbody></table> Operator's Signature <i>MG</i>	Test	g/210L	Time	Air Blank	0.000	10:21	Control Test	0.197	10:21	Air Blank	0.000	10:22	Control Test	0.197	10:23	Air Blank	0.000	10:23	Control Test	0.196	10:24	Air Blank	0.000	10:24	Control Test Stats			Average	0.1967		Std Dev	0.0006		Rel Std Dev(%)	0.2936		<i>DGS</i> POLK CSO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-007503 02/21/2023 Software: 8100.27 <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>10:02</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:02</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:02</td></tr><tr><td>Control Test</td><td>0.080</td><td>10:03</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:03</td></tr><tr><td>Control Test</td><td>0.081</td><td>10:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>10:04</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></tbody></table> Operator's Signature <i>MG</i>	Test	g/210L	Time	Air Blank	0.000	10:02	Control Test	0.080	10:02	Air Blank	0.000	10:02	Control Test	0.080	10:03	Air Blank	0.000	10:03	Control Test	0.081	10:04	Air Blank	0.000	10:04	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:07																																																																																																																																																	
Control Test	0.049	10:07																																																																																																																																																	
Air Blank	0.000	10:08																																																																																																																																																	
Control Test	0.049	10:09																																																																																																																																																	
Air Blank	0.000	10:09																																																																																																																																																	
Control Test	0.049	10:10																																																																																																																																																	
Air Blank	0.000	10:10																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0490																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:13																																																																																																																																																	
Control Test	0.078	10:14																																																																																																																																																	
Air Blank	0.000	10:15																																																																																																																																																	
Control Test	0.078	10:15																																																																																																																																																	
Air Blank	0.000	10:16																																																																																																																																																	
Control Test	0.079	10:17																																																																																																																																																	
Air Blank	0.000	10:17																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:21																																																																																																																																																	
Control Test	0.197	10:21																																																																																																																																																	
Air Blank	0.000	10:22																																																																																																																																																	
Control Test	0.197	10:23																																																																																																																																																	
Air Blank	0.000	10:23																																																																																																																																																	
Control Test	0.196	10:24																																																																																																																																																	
Air Blank	0.000	10:24																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1967																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2936																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	10:02																																																																																																																																																	
Control Test	0.080	10:02																																																																																																																																																	
Air Blank	0.000	10:02																																																																																																																																																	
Control Test	0.080	10:03																																																																																																																																																	
Air Blank	0.000	10:03																																																																																																																																																	
Control Test	0.081	10:04																																																																																																																																																	
Air Blank	0.000	10:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

Comments:

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: POLK CSO
Time of Inspection: 13:41

Date of Inspection: 02/21/2023

Serial Number: 80-007503
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202201C Exp: 01/11/2024	0.08g/210L Test (g/210L) Lot#:202201D Exp: 01/18/2024	0.20g/210L Test (g/210L) Lot#:202201E Exp: 01/18/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG223802 Exp: 08/26/2024
0.000	0.049	0.077	0.197	0.080
0.000	0.048	0.077	0.197	0.080
0.000	0.048	0.077	0.196	0.080
0.000	0.049	0.077	0.197	0.080
0.000	0.049	0.077	0.197	0.080
0.000	0.049	0.077	0.197	0.080
0.000	0.048	0.077	0.196	0.080
0.000	0.048	0.077	0.197	0.080
0.000	0.049	0.077	0.197	0.080
0.000	0.049	0.078	0.197	0.080

Standard Deviations	0.0005	0.0003	0.0004	0.0000
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/21/2023
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-007503, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-007503</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>POLK CSO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>02/21/2023</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>13:41</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

02/21/2023

Date


TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1