

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF COUNTY SO
Time of Inspection: 22:53

Date of Inspection: 01/25/2024

Serial Number: 80-000795
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402607186 Exp: 12/02/2025
0.000	0.049	0.078	0.199	0.080
0.000	0.049	0.079	0.200	0.080
0.000	0.049	0.079	0.199	0.080

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded.

Too much mouth alcohol used on first attempt.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

OPC. Dalton J. Brammer #107

DALTON J BRAMMER
Signature and Printed Name

01/25/2024
Date