

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: DIXIE COUNTY SO
Time of Inspection: 12:02

Date of Inspection: 01/31/2024

Serial Number: 80-000810
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14323080A3 Exp: 09/05/2025
0.000	0.048 / 0.048	0.079	0.187 / 0.199	0.082
0.000	/ 0.048	0.078	0.198 / 0.199	0.081
0.000	/ 0.048	0.079	0.200 / 0.200	0.082

Number of Simulators Used: 4


Remarks:

05: Purge Fail. 20: Control Outside TEMP LOW.

Purge fail
Low Temp
Corrected on 01-31-2024
At 11:58 AM
M.D.F.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

MATTHEW MANDOLA

01/31/2024
Date