

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 13:16

Date of Inspection: 12/28/2024

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202402E Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#:202410E Exp: 10/22/2026	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23324080A2 Exp: 10/05/2026
0.000	0.050	0.078	0.196	0.082
0.000	0.050	0.078	0.196	0.082
0.000	0.049	0.078	0.196	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

12/28/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Serial Number: 80-000812

Time of Inspection: 11:42

Date of Inspection: 09/22/2024

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000				
0.000				
0.000				

Number of Simulators Used: 5

Remarks:

Int Det: LOOSE CONNECTION, . Non-compliance: FAILED INT DETECT DUE TO LOOSE CONNECTIONS .

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER


Signature and Printed Name

09/22/2024
Date

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000812
09/22/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:43
Control Test	INT*	11:43
Air Blank	0.000	11:44
Control Test	INT*	11:45
Air Blank	0.000	11:45
Control Test	INT*	11:46
Air Blank	0.000	11:46
Control Test Stats		
Average	0.0023	
Std Dev	0.0040	
Rel Std Dev(%)	173.2051	

*Interferent Detect


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 12:51

Date of Inspection: 09/22/2024

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202402F Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08223080A3 Exp: 06/05/2025
0.000	0.049	0.078	0.195	0.081
0.000	0.050	0.079	0.196	0.083
0.000	0.052	0.079	0.196	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

09/22/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO
Time of Inspection: 15:11

Date of Inspection: 09/22/2024

Serial Number: 80-000814
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202402F Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08223080A3 Exp: 06/05/2025
0.000	0.051	0.079	0.197	0.078
0.000	0.051	0.080	0.196	0.078
0.000	0.050	0.080	0.197	0.078

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

09/22/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 20:55

Date of Inspection: 06/28/2024

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.012 / 0.000	0.049	0.080	0.198	0.081
0.000 / 0.000	0.050	0.080	0.198	0.081
0.000 / 0.000	0.049	0.080	0.198	0.080

Number of Simulators Used: 5

Remarks:

00: Control Outside Tolerance.

Utilization of new simulators resulted in trial and error. 191443

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

PATRICK M BAKER

06/28/2024
Date

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 11:23

Date of Inspection: 08/15/2024

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202402E Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000	0.050	0.080	0.196	0.082
0.000	0.049	0.080	0.196	0.082
0.000	0.049	0.080	0.196	0.082

Number of Simulators Used: 5

Remarks:

A F / M A: No Sample Provided.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

08/15/2024

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 11:51

Date of Inspection: 07/23/2024

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000	0.049	0.078	0.195	0.081
0.000	0.049	0.078	0.196	0.081
0.000	0.049	0.078	0.196	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature

BRAD R COLLITO

Signature and Printed Name

07/23/2024

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 19:34

Date of Inspection: 02/28/2024

Serial Number: 80-000812

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000	0.048	0.077	0.194	0.083
0.000	0.048	0.077	0.195	0.082
0.000	0.048	0.077	0.195	0.082

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

SP ALIN62

PATRICK M BAKER

Signature and Printed Name

02/28/2024

Date