

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO
Time of Inspection: 11:29

Date of Inspection: 12/28/2024

Serial Number: 80-000814
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202402E Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#:202410E Exp: 10/22/2026	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:23324080A2 Exp: 10/05/2026
0.000	0.050	0.079	0.196	0.080
0.000	0.050	0.078	0.197	0.079
0.000	0.050	0.079	0.197	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

12/28/2024
Date

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
12/28/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:43
Control Test	0.000	09:43
Air Blank	0.000	09:44
Control Test	0.000	09:45
Air Blank	0.000	09:45
Control Test	0.000	09:46
Air Blank	0.000	09:46
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	


Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
12/28/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:47
Control Test	INT*	09:48
Air Blank	0.000	09:48
Control Test	INT*	09:49
Air Blank	0.000	09:50
Control Test	INT*	09:50
Air Blank	0.000	09:51
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*Interferent Detect


Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
12/28/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:52
Control Test	0.051	09:52
Air Blank	0.000	09:53
Control Test	0.050	09:53
Air Blank	0.000	09:54
Control Test	0.049	09:55
Air Blank	0.000	09:55
Control Test Stats		
Average	0.0500	
Std Dev	0.0010	
Rel Std Dev(%)	2.0000	



Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
12/28/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:52
Control Test	0.078	10:53
Air Blank	0.000	10:54
Control Test	0.077	10:54
Air Blank	0.000	10:55
Control Test	0.077	10:56
Air Blank	0.000	10:56
Control Test Stats		
Average	0.0773	
Std Dev	0.0006	
Rel Std Dev(%)	0.7466	



Operator's Signature

HERNANDO COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000814
12/28/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	10:12
Control Test	0.196	10:13
Air Blank	0.000	10:13
Control Test	0.196	10:14
Air Blank	0.000	10:15
Control Test	0.195	10:15
Air Blank	0.000	10:16
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	


Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 21:58

Date of Inspection: 06/25/2024

Serial Number: 80-000814

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.011 / 0.000	0.044 / 0.048	0.074 / 0.078	0.172 / 0.191	0.080
0.000 / 0.000	0.046 / 0.049	0.077 / 0.079	0.186 / 0.195	0.080
0.000 / 0.000	0.047 / 0.048	0.079 / 0.079	0.191 / 0.195	0.080

Number of Simulators Used: 5

Remarks:

00: Control Outside Tolerance. 05: Control Outside Tolerance. 08: Control Outside Tolerance. 20: Control Outside Tolerance.

Utilization of new simulators resulted in trial and error. P81163

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

06/25/2024

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 08:33

Date of Inspection: 07/23/2024

Serial Number: 80-000814

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000	0.050	0.079	0.196	0.080
0.000	0.049	0.079	0.197	0.080
0.000	0.049	0.079	0.197	0.080

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

07/23/2024

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Serial Number: 80-000814

Time of Inspection: 20:57

Date of Inspection: 05/28/2024

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: 202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot #: 08223080A3 Exp: 06/05/2025
0.000	0.048	0.078	0.193	0.079
0.000	0.048	0.078	0.194	0.079
0.000	0.049	0.079	0.194	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M BAKER

Signature and Printed Name

05/28/2024

Date

FDLE/ATP Form 40 -- March 2004

Reprinted From Database

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 17:16

Date of Inspection: 03/19/2024

Serial Number: 80-000814

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08223080A3 Exp: 06/05/2025
0.000	0.049	0.078	0.196	0.079
0.000	0.049	0.078	0.196	0.079
0.000	0.048	0.078	0.196	0.079

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



PATRICK M. BAKER

Signature and Printed Name

03/19/2024

Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY S.O.

Time of Inspection: 19:55

Date of Inspection: 02/28/2024

Serial Number: 80-000814

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:08223080A3 Exp: 06/05/2025
0.000	0.049	0.078	0.195	0.000 / 0.080
0.000	0.049	0.078	0.196	0.080 / 0.080
0.000	0.049	0.079	0.197	0.080 / 0.080

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

PATRICK M BAKER

02/28/2024
Date