

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HILLSBOROUGH CO SO

Time of Inspection: 10:14

Date of Inspection: 04/10/2024

Serial Number: 80-000830

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

ACCIDENTALLY SELECTED NO FOR DIAGNOSTIC. Non-compliance: ACCIDENTALLY SELECTED NO FOR DIAGNOSTIC.

AI B. KROME #263389 PRESENT FOR INSPECTION. BK 263389

SECOND INSPECTION PERFORMED AFTER CONTACTING
FOLE / ATP - SHAYLA PLATT.


The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 261252

STEPHANIE R KNOX
Signature and Printed Name

04/10/2024
Date

 4-10-24