

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH PD

Time of Inspection: 11:01

Date of Inspection: 07/11/2024

Serial Number: 80-000876

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402370890 Exp: 05/09/2025
0.000	0.048	0.078	0.201	0.072 / 0.076
0.000	0.049	0.080	0.202	0.072 / 0.075
0.000	0.050	0.079	0.201	0.072 / 0.075

Number of Simulators Used: 5

Remarks:

A F / M A: . CHANGED DRY GAS CYLINDER AND RECONNECTED TUBE. FDLE INSPECT.

AF/MA: First test too little mouth Alcohol.

Dry gas. changed Dry gas cylinder and reconnected tube. Retried the test and it passed. FDLE Inspector Gutschow advised.

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jose Montero Jr #1763

JOSE MONTERO

Signature and Printed Name

07/11/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIALEAH PD
Time of Inspection: 09:54

Date of Inspection: 07/11/2024

Serial Number: 80-000876
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08g/210L Test (g/210L) Lot#:_____ Exp:_____	0.20g/210L Test (g/210L) Lot#:_____ Exp:_____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:_____ Exp:_____
0.000				
0.000				
0.000				

Number of Simulators Used: 5

Remarks:

ABORTED INSPECTION AFTER TALKING TO FDLE INSPECTOR GUTSCHOW ~~Non-compliance~~ *Will redo monthly inspection.*
due to two (2) reading on the AF/MA.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Jose Montero Jr #1763

JOSE MONTERO

Signature and Printed Name

07/11/2024
Date