

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TALLAHASSEE PD

Time of Inspection: 09:49

Date of Inspection: 10/22/2024

Serial Number: 80-000949

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:33922080A1 Exp: 01/05/2025
0.000	0.047	0.076	0.195	0.000 / 0.000
0.000	0.047	0.076	0.196	0.080 / 0.000
0.000	0.047	0.076	0.196	0.080 / 0.080


Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance, Control Outside Tolerance. Non-compliance: .

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 **TP0321**

JAMES J HARRIS

Signature and Printed Name

10/22/2024
Date