AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO Time of Inspection: 13:22

Date of Inspection: 05/19/2024

Serial Number: 80-001071

Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | |
| | | No |
| Diagnostic Check (Pre-Inspection): OK | | |
| | Yes | |
| Alcohol Free Subject Test: 0.000 | | |
| | Yes | |
| Mouth Alcohol Test: Slope Not Met | | |
| | Yes | |
| Interferent Detect Test: Interferent Detect | | |
| | Yes | |
| Diagnostic Check (Post-Inspection): OK | | |
| | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:1524080A1 Exp: 03/05/2026 |
|----------------------------------|--|--|--|--|
| 0.000 | 0.048 | 0.077 | 0.197 | 0.079 |
| 0.000 | 0.048 | 0.078 | 0.198 | 0.079 |
| 0.000 | 0.048 | 0.079 | 0.198 | 0.079 |

Number of Simulators Used: 5

Remarks:

MONTHLY INSPECTION RETURN FOR ANNUAL INSP

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

RICHARD L PATRONE

Signature and Printed Name

05/19/2024 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO Time of Inspection: 12:22

Date of Inspection: 04/21/2024

Serial Number: 80-001071 Software: 8100.27

| Check or Test | YES | NO |
|---|-----|-----|
| Date and/or Time Adjusted | | 140 |
| Diagnostic Check (Pre-Inspection): OK | | No |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |
| | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:07622080A2 |
|----------------------------------|--|--|--|--|
| 0.000 | 0.048 | 0.078 | 0.197 | 0.079 |
| 0.000 | 0.048 | 0.078 | | |
| 0.000 | 0.048 | | 0.198 | 0.079 |
| | 0.048 | 0.079 | 0.198 | 0.078 |

Number of Simulators Used: 5

Remarks:

MONTHLY INSPECTION

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW P JOHNSON

Signature and Printed Name

04/21/2024 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO

Time of Inspection: 09:50

Date of Inspection: 03/24/2024

Serial Number: 80-001071 Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | 125 | NO |
| Diagnostic Check (Pre-Inspection): OK | | No |
| | Yes | |
| Alcohol Free Subject Test: 0.000 | | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| | Yes | |
| Interferent Detect Test: Interferent Detect | | |
| Diagnostic Check (Post-Inspection): OK | Yes | |
| | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:07622080A2 Exp: 05/05/2024 |
|----------------------------------|--|--|--|---|
| 0.000 | 0.048 | 0.077 | 0.198 | 0.078 |
| 0.000 | 0.048 | 0.078 | 0.199 | 0.079 |
| 0.000 | 0.048 | 0.077 | 0.199 | 0.079 |

Number of Simulators Used: 5

Remarks:

MONTHLY AI

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed that inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

MATTHEW P JOHNSON

03/24/2024 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO

Time of Inspection: 07:15

Date of Inspection: 02/25/2024

Serial Number: 80-001071

Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | |
| | | No |
| Diagnostic Check (Pre-Inspection): OK | | |
| | Yes | |
| Alcohol Free Subject Test: 0.000 | | |
| | Yes | |
| Mouth Alcohol Test: Slope Not Met | | |
| | Yes | |
| Interferent Detect Test: Interferent Detect | | |
| | Yes | |
| Diagnostic Check (Post-Inspection): OK | | |
| • | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:07622080A2 Exp: 05/05/2024 |
|----------------------------------|--|--|--|---|
| 0.000 | 0.048 | 0.078 | 0.199 | 0.078 |
| 0.000 | 0.049 | 0.078 | 0.198 | 0.079 |
| 0.000 | 0.049 | 0.079 | 0.198 | 0.079 |

Number of Simulators Used: 5

Remarks:

MONTHY INSPECTION

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Signature and Printed Name

02/25/2024

RICHARD L PATRONE

Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLAY COUNTY SO

Time of Inspection: 12:02

Date of Inspection: 01/27/2024

Serial Number: 80-001071

Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | |
| | | No |
| Diagnostic Check (Pre-Inspection): OK | | |
| | Yes | |
| Alcohol Free Subject Test: 0.000 | | |
| | Yes | |
| Mouth Alcohol Test: Slope Not Met | | |
| | Yes | |
| Interferent Detect Test: Interferent Detect | | |
| | Yes | |
| Diagnostic Check (Post-Inspection): OK | | |
| | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025 | 0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024 | 0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:07622080A2 Exp: 05/05/2024 |
|----------------------------------|--|--|--|---|
| 0.000 | 0.047 | 0.077 | 0.197 | 0.078 |
| 0.000 | 0.048 | 0.077 | 0.198 | 0.078 |
| 0.000 | 0.048 | 0.078 | 0.198 | 0.078 |

Number of Simulators Used: 5

Remarks:

MONTHLY INSPECTION

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

MATTHEW P JOHNSON

Signature and Printed Name

01/27/2024 Date