

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ZEPHYRHILLS PD

Serial Number: 80-001076

Time of Inspection: 08:51

Date of Inspection: 05/30/2024

Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 202303J<br>Exp: 03/28/2023 <sup>45</sup> | 0.08g/210L Test (g/210L)<br>Lot#: 202307A<br>Exp: 07/11/2025 | 0.20g/210L Test (g/210L)<br>Lot#: 202307C<br>Exp: 07/12/2025 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot #: 06023080A2<br>Exp: 04/05/2025 |
|----------------------------|----------------------------------------------------------------------------|--------------------------------------------------------------|--------------------------------------------------------------|-------------------------------------------------------------------------------|
| 0.000                      | 0.049                                                                      | 0.079                                                        | 0.197                                                        | 0.077                                                                         |
| 0.000                      | 0.049                                                                      | 0.079                                                        | 0.199                                                        | 0.077                                                                         |
| 0.000                      | 0.050                                                                      | 0.079                                                        | 0.200                                                        | 0.077                                                                         |

Number of Simulators Used: 4

**Remarks:**

A F / M A: GOT READING ON MA.

*Amended .05 Exp date to 3/28/2025*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

AILISA J STARNES

Signature and Printed Name

05/30/2024

Date