

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: ORMOND BEACH PD
Time of Inspection: 15:38

Date of Inspection: 08/31/2024

Serial Number: 80-001156
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 5


Remarks:

A F / M A: Range Exceeded.BLEW TO HARD. NOT ENOUGH MAS. Non-compliance:WILL CONDUCT ANOTHER TEST.

Appears the first bottle of MAS was no longer good.
Used a new bottle of MAS and second inspection was fine.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 40371
Signature and Printed Name

MICHAEL D GARNER

08/31/2024
Date