

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: CLERMONT P.D.

Time of Inspection: 10:33

Date of Inspection: 07/25/2024

Serial Number: 80-001174

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402357111 Exp: 02/10/2025
0.000	0.050	0.078	0.196	0.080 / 0.081
0.000	0.050	0.078	0.197	0.086 / 0.081
0.000	0.050	0.078	0.198	0.087 / 0.081

Number of Simulators Used: 4

Remarks:

08: Control Outside Tolerance. → Spoke with Department Inspector and took corrective actions which included:

- checked Dry Gas PSI
- aired out with fan and moved simulators farther away
- disconnected and reconnected Dry Gas cylinder hose

- In compliance

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Chad Wheeler A210

CHAD T WHEELER

Signature and Printed Name

07/25/2024
Date