

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Jacksonville Sheriff's Office

Serial Number: 80-001279

Time of Inspection: 1000

Date of Inspection: 02/16/2024

Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		NO
Diagnostic Check (Pre-Inspection): OK		NO
Alcohol Free Subject Test: 0.000		NO
Mouth Alcohol Test: Slope Not Met		NO
Interferent Detect Test: Interferent Detect		NO
Diagnostic Check (Post-Inspection): OK		NO

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: 0

Instrument would not come to (Ready Mode). Unable to perform Agency Inspection due to Diagnostic Failure of the DSP System in Instrument #80-001279. Notified my Departmental Inspector of Issue. Sent back to manufacturer (CMI) for inspection and maintenance. Then PDLE for Departmental Inspection.

The above instrument complies () does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Robert D Lightsey

Signature and Printed Name

02-16-2024

Date

