

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: INDIAN RIVER CSO
Time of Inspection: 10:14

Date of Inspection: 05/23/2024

Serial Number: 80-001328
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402806340 Exp: 07/26/2026
0.000	0.044 / 0.045	0.074 / 0.075	0.188 / 0.190	0.080
0.000	0.045 / 0.045	0.074 / 0.075	0.190 / 0.191	0.079
0.000	0.045 / 0.046	0.075 / 0.075	0.191 / 0.192	0.079

Number of Simulators Used: 4

Remarks:

05: Control Outside Tolerance. 08: Control Outside Tolerance..

Simulator tubes need to be replaced.
Contacting Tech Source
Instrument in Compliance

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



BRITTANY D RUDOLPH

Signature and Printed Name

05/23/2024
Date