

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 10:54

Date of Inspection: 11/25/2024

Serial Number: 80-001345

Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#: 202303J<br>Exp: 03/28/2025 | 0.08g/210L Test (g/210L)<br>Lot#: 202307A<br>Exp: 07/11/2025 | 0.20g/210L Test (g/210L)<br>Lot#: 202307C<br>Exp: 07/12/2025 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#: 26824080A1<br>Exp: 11/05/2026 |
|----------------------------|--------------------------------------------------------------|--------------------------------------------------------------|--------------------------------------------------------------|------------------------------------------------------------------------------|
| 0.000                      | 0.049                                                        | 0.079                                                        | 0.198                                                        | 0.000 / 0.079                                                                |
| 0.000                      | 0.049                                                        | 0.080                                                        | 0.199                                                        | 0.000 / 0.079                                                                |
| 0.000                      | 0.049                                                        | 0.081                                                        | 0.199                                                        | 0.000 / 0.079                                                                |


Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. DRY GAS CANISTER NOT FULLY INSERTED; TIGHTEN FOR A SEALED FIT  
HE RETESTED, MET STANDARD. 11-25-24

The above instrument complies ( X ) does not comply ( ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

  
Signature and Printed Name

CAMILO J ALMEIDA

11/25/2024  
Date