

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LEE COUNTY SO
Time of Inspection: 09:21

Date of Inspection: 01/25/2024

Serial Number: 80-001722
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:22521080A1 Exp: 10/05/2024
0.000	0.047	0.072 / 0.077	0.175 / 0.193	0.080
0.000	0.047	0.078 / 0.078	0.189 / 0.195	0.080
0.000	0.048	0.078 / 0.078	0.191 / 0.194	0.081

Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance- WEAK O-RING SEAL. 20: Control Outside Tolerance- SIM NOT HOLDING TEMP

08 correction: O-ring seal was tightened.

20 correction: sim was brought back to temp and held at temp before continuing.

MA 0542

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Nicholas M Duncan

NICHOLAS M DUNCAN

Signature and Printed Name

01/25/2024
Date