

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: South Miami Police Department
Time of Inspection: N/A

Date of Inspection: (Unable)

Serial Number: 80-001737
Software: 8100.27

CHECK OR TEST	YES	NO
Date and/or Time Adjusted		X
Diagnostic Check (Pre-Inspection): OK		X
Alcohol Free Subject Test: 0.000		X
Mouth Alcohol Test: Slope Not Met		X
Interferent Detect Test: Interferent Detect		X
Diagnostic Check (Post-Inspection): OK		X

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

Instrument will not stay on; therefore inspection unable to be completed.

The above instrument complies (____) does not comply (X) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

N. Waddell

Nathaniel Waddell

Signature and Printed Name

7/17/2024

Date

#1818