AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: Lakeland Time of Inspec		of Inspection: 07/1		Number: 1865 ce: 8100.27	
CHECK OR TES	r			YES	NO
Date and/or T:	ime Adjusted				
Diagnostic Che	eck (Pre-Inspectio	n): OK			
Alcohol Free	Subject Test: 0.00	0			
Mouth Alcohol	Test: Slope Not M	et			
Interferent De	etect Test: Interf	erent Detect			
Diagnostic Che	eck (Post-Inspecti	on): OK			
Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210 Dry Gas St (g/210L) Lot#: Exp:	
		unresponsive keyboard. A	ttempted to swap out with	other keyboard, a	nd there
I certify that I	hold a valid Florida	Department of Law Enwith the provisions of	forcement Agency Inspe Chapter 11D-8, FAC.		nd that
Copie		Camilo Jo Signature and Printed	Name		

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Serial Number: 80-001865

Time of Inspection: 11:36

Date of Inspection: 05/03/2024

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14323080A3 Exp: 09/05/2025
0.000	0.000 / 0.048	0.079	0.195	0.080
0.000	0.046 / 0.048	0.079	0.196	0.080
0.000	0.047 / 0.048	0.079	0.197	0.079

Number	of	Simulators	Used:	5
--------	----	------------	-------	---

Remarks:

05: Control Outside Tolerance. SIM WAS NOT ATTACHED FOR FIRST TEST, RE ATTAC RE TESTED AFTER PATTACHING, MET STANDARD. 6/3/24 CA

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CAMILO J ALMEIDA

Signature and Printed Name

05/03/2024 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Serial Number: 80-001865

Time of Inspection: 11:36

Date of Inspection: 05/03/2024

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14323080A3 Exp: 09/05/2025
0.000	0.000 / 0.048	0.079	0.195	0.080
0.000	0.046 / 0.048	0.079	0.196	0.080
0.000	0.047 / 0.048	0.079	0.197	0.079

Number	of	Simulators	Used:	5
--------	----	------------	-------	---

Remarks:

05: Control Outside Tolerance. SIM WAS NOT ATTACHED FOR FIRST TEST, RE ATTAC RE TESTED AFTER PATTACHING, MET STANDARD. 6/3/24 CA

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CAMILO J ALMEIDA

Signature and Printed Name

05/03/2024 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Serial Number: 80-001865

Time of Inspection: 10:37

Date of Inspection: 06/18/2024

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307¢ Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14323080A3 Exp: 09/05/2025
0.000	0.048	0.078	0.196	0.078
0.000	0.048	0.078	0.198	0.078
0.000	0.049	0.079	0.198	0.078

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded. - TOO MUCH MA., WAITED RETEST BOTH AF/MA, MET STANDARD. 4/18/24

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CAMILO J ALMEIDA

Signature and Printed Name

06/18/2024 Date

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Serial Number: 80-001865

Time of Inspection: 10:35

Date of Inspection: 01/29/2024

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		
		No
Diagnostic Check (Pre-Inspection): OK		
	Yes	
Alcohol Free Subject Test: 0.000		
	Yes	
Mouth Alcohol Test: Slope Not Met		
	Yes	
Interferent Detect Test: Interferent Detect		
	Yes	
Diagnostic Check (Post-Inspection): OK		
	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:14323080A3 Exp: 09/05/2025
0.000	0.048	0.079	0.197	0.081
0.000	0.048	0.080	0.198	0.081
0.000	0.049	0.080	0.199	0.081

Number o	ρf	Simulators	Used:	5
----------	----	------------	--------------	---

Remarks:

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CAMILO J ALMEIDA

Signature and Printed Name

01/29/2024 Date